Robotic Retinal Surgery¹

Emmanuel Vander Poorten^{a,*}, Cameron N. Riviere^{**}, Jake J. Abbott[†],
Christos Bergeles[‡], M. Ali Nasseri[§], Jin U. Kang[¶], Raphael Sznitman^{||},
Koorosh Faridpooya^{††} and Iulian Iordachita^{‡‡}

* KU Leuven, Department of Mechanical Engineering, Celestijnenlaan 300B, 3001 Heverlee, Belgium

** Carnegie Mellon University, Robotics Institute, 5000 Forbes Avenue, Pittsburgh, PA, 15213, USA

† University of Utah, Department of Mechanical Engineering, 1495 E. 100 S., MEK 1550, Salt Lake City, UT, 84112, USA

‡ King's College London, School of Biomedical Engineering and Imaging Sciences, Becket House, 1 Lambeth Palace Road, SE1 7EU, London, UK

§ Technische Universität München, Ophthalmology Department, Ismaninger Str. 22, 81675 München, Germany

¶ Johns Hopkins University, Department of Electrical and Computer Engineering, Baltimore, Maryland USA

|| University of Bern, ARTORG Center for Biomedical Engineering Research, Murtenstrasse 50, CH-3008, Bern, Switzerland

†† Eye Hospital Rotterdam, Schiedamse vest 180, 3011BH, Rotterdam, Netherlands

‡‡ Johns Hopkins University, Department of Mechanical Engineering, Baltimore, Maryland USA

^a Corresponding: emmanuel.vanderpoorten@kuleuven.be

Abstract

Retinal surgery has long drawn the attention of engineers and clinicians who identified a clear use case for robotics and assistive technology. In retinal surgery, precision is paramount. Skilled practitioners operate on the boundaries of human capability, dealing with minuscule anatomic structures that are both fragile and hard to discern. Surgical operations on the retina, a hair-thick multilayered structure that is an integral part of the central nervous system responsible for vision, spurred the development of robotic system that enhance perception, precision, and dexterity. This chapter provides an encompassing overview of the progress that has been made during the last two decades in terms of sensing, modeling, visualization, stabilization, and control. The chapter reports on recent breakthroughs with first-in-human experiences, as well as on new venues that hold the potential to expand retinal surgery to techniques that would be infeasible or challenging without robotics.

Keywords: Robotics, Retinal Surgery, Vitreoretinal Surgery, Eye Surgery, Microsurgery, Sensing, Augmented Reality, Modeling and Control

Chapter points

- The current state-of-the-art in robotic retinal surgery is surveyed as well as
- newest trends in instrumentation, robotics, image-processing and modelling
- and an outlook upon future work.

¹For submission to Elsevier's Handbook of Robotic and Image-Guided Surgery

1. The Clinical Need

The retina is a ‘layer of nervous tissue that covers the inside of the back two-thirds of the eyeball, in which stimulation by light occurs, initiating the sensation of vision’ and ‘is actually an extension of the brain, formed embryonically from neural tissue and connected to the brain proper by the optic nerve’ [190]. Any damage to the retina may cause irreversible and permanent visual field defect or even blindness. Key structures that are the subject of different surgical interventions are depicted in Fig. 0.1 and include the sclera, retinal vessels, scar tissue or epiretinal membranes and, recently, the retinal layers. A list of parameters and dimensions that characterize these structures is provided in Table 1. In comparison, we note that the diameter of the average human hair is $50\mu\text{m}$, which highlights the micromanipulation challenges in retinal surgery.

Open-sky surgery is a less-than-desirable option when treating critically fragile structures within the eye, such as the retina. Surgeons approach the retina through a ‘key-hole’ set-up, inserting slender instruments through small incisions in the sclera to operate at a micrometer scale on structures whose complexity rivals or exceeds that of the brain. Visualization occurs through a stereo operating microscope. The

Figure 0.1 A cross-section of a human eye. A cannula is placed at 4 mm from the cornea limbus (the border of the cornea and the sclera), providing access to the intraocular space.

Table 0.1 Governing dimensions in retinal surgery

Structure	Dimension	Comment/Sources
human eye	24.6 mm avg	axial length [174]
human retina	100–300 μm	thickness [115]
internal limiting membrane	0.5–2.5 μm , 1–3 μm	maximal at macula [3], [38]
epiretinal membrane	60 μm	cellular preretinal layer [203]
retinal vessel	40–350 μm , 40–120 μm	branch to central [14, 115]
vessel puncture force	20 mN avg, 181 mN max	cadaver pig eye [96]
	63% < 5 mN	cadaver pig eye [73]
	0.6–17.5 mN; 80% < 7 mN	cadaver pig eye [64]
	2 mN avg, 1 mN std	fertilized chicken egg [183]
	80% < 5 mN	fertilized chicken egg [55]
vessel dissection force	67 mN avg, 82 mN max	cadaver pig eye [96]
peeling force	8–12 mN, 15–45 mN	inner shell membrane (ISM) of chicken egg [183, 95]
damage during peeling	from 5.1 mN	fertilized chicken egg [184]
	from 6.4mN	rabbit [184]
retina damage	1788 Pa	17.2mN on 3.5mm diameter [19]
breathing frequency	3 Hz; 0.2 Hz	rat [173]; pig [146]
breathing amplitude	50 μm ; 300 μm	rat [173]; pig [146]
heartbeat frequency	0.84 Hz; 2 Hz	rat [173]; pig [146]
heartbeat amplitude	15 μm ; 100 μm	rat [173]; pig [146]
required positioning accuracy	10 μm	general [4, 39].
required positioning accuracy	25 μm	subretinal injection [221].

incision forms a fulcrum point. This fulcrum complicates hand-eye coordination due to the inverted relationship between hand and instrument motion (Fig. 0.2). If the instrument is not pivoted exactly about the fulcrum point a net force will be applied to the sclera which could damage the sclera or could potentially cause the eye to rotate in its socket. When the eye rotates it becomes more difficult to reach a location the retina precisely as the target location changes dynamically. The surgeon uses the support of an armrest (elbow) and the patient’s head (wrists) to stabilize the hands. Lightweight instruments are maneuvered within the confined space between the patient’s head and the microscope. A wide-angle lens is often placed between the eye and microscope offering a larger view of the retina. It does limits the work volume that is available.

1.1. Human Factors and Technical Challenges

Retinal microsurgery demands advanced surgical skills. The requirements for vision, depth perception, and fine motor control are high (Table 0.2), exceeding the fundamental physiological capability of many individuals [36, 75, 148]. A primary cause of tool positioning error is physiological tremor [201]. Even when microsurgical procedures

Figure 0.2 Overall layout and view during retinal surgery. (Left) Retinal surgical scene using surgical microscope surgeon holding vitrectome in right hand and light pipe in the left; (right) typical view during an internal limiting layer (ILM) peeling

Table 0.2 Human factors and technical limitations in retinal surgery

Parameter	Value	Comment/Sources
physiologic tremor	182 μm , 100 μm rms 156 μm rms 8–12 Hz	epiretinal membrane removal [162, 175] artificial eye model [150] neurogenic tremor component [4]
fulcrum motion	area up to 12.6mm ²	during e.g manual vitrectomy [88]
maximum velocity	0.7 m/s	epiretinal membrane removal [162]
typical velocity	0.1–0.5 mm/s	epiretinal membrane peeling [7]
maximum acceleration	30.1 m/s ²	epiretinal membrane removal [162]
manipulation forces	< 7.5 mN in 75%	ex-vivo pig eye membrane [73]

are successfully performed, in the presence of tremor they require greater concentration and effort and are attended by greater risk. Patient movement is another important confounding factor. Among patients who snore under monitored anesthesia ($\approx 16\%$), half have sudden head movements during surgery, leading to a higher risk of complications [130]. The challenges of retinal microsurgery are further exacerbated by the fact that in the majority of contact events, the forces encountered are below the tactile

perception threshold of the surgeon [73]. Inability to detect surgically relevant forces leads to a lack of control over potentially injurious factors that result in complications.

Aside from the poor ergonomics of operating through a surgical microscope, leading to an elevated risk for back and neck injuries, with incidence of 30% to 70% for neck pain and 40% to 80% for back pain [134], this approach is associated with difficult hand-eye coordination. Without haptic feedback, the surgeon can only rely on visual feedback. However, the quality of the visual feedback is still not good enough. Surgeons expend considerable effort adjusting the optics and illumination to obtain the appropriate level of clarity, detail, and overview of the target scene. Depth perception is suboptimal. Even with modern stereo microscopes, surgeons are still sometimes unsure exactly when contact with the retina is established. Poor visualization due to factors such as corneal scars or intense vitreous hemorrhage can affect the outcome of retinal surgeries and increase the chance of complications.

1.2. Motivation for Robotic Technology

Given the size and fragile nature of the structures involved, complication rates are not negligible [56, 104, 214]. Surgical steps that are considered too risky or even impossible may be facilitated through robotics. There is also an interest in automating repetitive tasks to reduce cognitive load and allow experts to focus on critical steps of a procedure. Ergonomy represents another area of potential innovation. One can reconsider the operating layout and optimize usability to reduce physical burdens.

Some appealing characteristics of robotic technology for treating the retina include improved positioning accuracy through some combination of motion scaling and tremor reduction, the ability to keep an instrument steady and immobilized for a prolonged period of time, and the ability to save coordinates for future use. The retina is neural tissue; even a small mistake can cause irreversible damage, including blindness. Through robotics, procedures that cannot be performed safely using conventional manual techniques due to limitations in precision, such as microcannulation and subretinal injection, may be considered. In current manual practice, surgeons can only use two instruments simultaneously, although three or more instruments would be helpful in complicated cases, such as delaminations. Robotics further facilitates integration with advanced tooling. Dedicated interfaces could help manage instruments with articulating end-effectors. User interfaces can be tailored to provide feedback from a broad range of sensors embedded in a new line of ‘intelligent’ instruments. Robotic surgery may enable operation with narrower instruments, which would decrease the size of scleral incisions and reduce damage to the sclera.

Taken in combination, the above characteristics could create a highly effective therapeutic system for performing advanced microsurgical procedures. Not only could the added functionality decrease complication rates, it could also speed up healing and shorten the duration of admission in clinic. For robotics to be successful, the above

arguments would need to outweigh the disadvantages of elevated operational cost and increased operation time that seem inevitable, based on today’s technology.

1.3. Main Targeted Interventions

The following retinal procedures have received considerable attention from researchers who identified opportunities for improvement by use of robotic technology.

1.3.1. Epiretinal Membrane Peeling

An epiretinal membrane (ERM) is an avascular, fibrocellular membrane, such as a scar tissue, that may form on the inner surface of the retina and cause blurred and distorted central vision. Risk for ERM increases with age, primarily affecting people over age 50. ERM is mostly idiopathic and related to an abnormality of the vitreoretinal interface in conjunction with a posterior vitreous detachment. ERM can also be triggered by certain eye diseases such as a retinal tear, retinal detachment, and inflammation of the eye (uveitis). The prevalence of ERM is 2% in individuals under age 60 and 12% in those over age 70 [135]. Although asymptomatic, ERM often leads to reduced visual acuity and metamorphopsia, where straight lines can appear wavy due to contraction forces acting over the macular region [5]. Treatment is surgical and only when the patient suffers from binocular metamorphopsia and progressive visual decrease less than 50%. The procedure involves pars-plana vitrectomy, followed by removal (peeling) of the ERM, with or without peeling of the native internal limiting membrane (ILM) in order to decrease the recurrence of ERM afterwards [37].

1.3.2. Retinal Vein Cannulation

Retinal vein occlusion is the second-most-prevalent vasculature-related eye disease [164]. A blood clot clogs the vein, which leads to a sudden halt in retinal perfusion. Since arterial inflow continues, hemorrhages develop and the retina may become ischemic, leading to retinal neural cell apoptosis. Depending on the thrombus location, one distinguishes between central retinal vein occlusion (CRVO) and branch retinal vein occlusion (BRVO), *i.e.*, when the thrombus resides in a smaller branch vein. BRVO can be asymptomatic but may lead up to sudden painless legal blindness. Secondary macular edema can develop and cause metamorphopsia. Later on, neovascularization can occur because of ischemic retina and cause secondary glaucoma, retinal detachment, and vitreous hemorrhage [131]. There is no etiologic curative treatment at present. One of the few symptomatic treatments that are offered are injections to prevent neovascularization, delivered directly into the eye. The injected medicine can help reduce the swelling of the macula. Steroids may also be injected to help treat the swelling and limit the damage to the occluded tissue. If CRVO is severe, ophthalmologists may apply panretinal photocoagulation wherein a laser is used to make tiny burns in areas of the retina. This lowers the chance of intraocular bleedings and can

prevent eye pressure from rising to sight-threatening levels. Retinal vein cannulation,.

1.3.3. Subretinal Injection

In procedures such as antivascularization treatment, drugs are commonly administered in the vitreous humor to slow down neovascularization. Although intravitreal injections are fairly simple, when targeting cells in subretinal spaces the dose that actually reaches those cells could be very small. Subretinal injection is an alternative where drugs are directly injected in the closed subretinal space. Subretinal injection is regarded as the most effective delivery method for cell and gene therapy—including stem-cell therapy for degenerative vitreoretinal diseases such as retinitis pigmentosa, age-related macular degeneration, and Leber’s congenital amaurosis [182]—despite it potentially leading more often to adverse events and possible complications [149].

1.4. Models Used for Replicating the Anatomy

To support technology development for the above mentioned procedures, a variety of synthetic, in-vitro, and in-vivo models have been proposed over the past decade. Table 0.3 provides an overview of the most commonly used models and some indicative references to works where they are described or deployed. Due to the complexity of the human eye, different models are suited for each surgical intervention, with no single model satisfying all requirements. Despite the abundance of available models, research is still ongoing to further improve the existing models. For example, for membrane peeling, Gupta *et al.* have been searching for representative in-silico models [72]. For vein cannulation, the Rotterdam Eye hospital has been developing an ex-vivo perfused pig eye model that can be used to evaluate retinal motion or vessel coagulation [195]. A modified Rose bengal method has been developed to create clotted vessels in live pigs for validating cannulation performance [45, 204].

Table 0.3 Models used for simulating and testing retinal surgeries, including membrane peeling, vein cannulation, and injections

Model	Peeling	Cannulation	Inj.	Comment
synthetic membranes	[7, 66, 72, 140]			peeling of membrane
gelatin phantom			[99]	10% mimics tissue
soft cheese			[221]	similar OCT response
rubber band	[79]			simulates scleral contact
agar	[107]	[199]	[17]	vitreous humour
raw chicken egg	[66, 95, 178]			peeling ISM
fertilized chicken egg	[95, 115, 183]	[59, 115, 183]		peeling ISM
cadaver bovine eye	[41]		[99]	w/o cornea, lens, vitreous
cadaver pig eye		[96, 14, 192]	[17]	open-sky; 40–60 μm
perfused pig eye	[195]			closure of vessels
in-vivo pig eye	[45, 204]			w/ lasering to form clots
in-vivo rabbit eye	[96, 209]	[2, 96]		preretinal 60 μm vessels
in-vivo cat eye		[2]		intraretinal vessels

2. Visualization in Retinal Surgery

As force levels remain predominantly below human perceptual thresholds, haptic feedback is of no avail in current surgical practice. This section explains the basic technology that is available for visualization. Over the years, a broad range of medical imaging technologies have played crucial roles in imaging the retina pre-operatively and during interventions. In the following, we describe some of the most important modalities related to robotic microsurgery, with an emphasis on the stereo microscope (§2.1), as it plays a central role in the link between the patient and the operating physician. The second part of this section (§2.2) introduces optical coherence tomography (OCT) as an imaging modality with rapidly increasing importance in retinal surgery.

2.1. Basic Visualization through Operative Stereo-microscope

Operative microscopes are the primary tool to image the surgical site during retinal microsurgery and are fully integrated into the standard of care worldwide. With a number of commercial vendors offering stereo microscopes (Zeiss, Leica Microsystems, Haag-streit Surgical, TopCon Medical Systems), most provide high-quality magnified and illuminated viewing of the surgical area. The obtained image quality is a result of a plurality of components briefly summarized in the following.

2.1.1. Stereo Microscope

At its core, a stereo microscope is composed of a binocular head mount that allows the operating clinician to view the surgical site via an optical system. Typically, the optical system consists of a set of lenses and prisms that connect to an objective lens that dictates the working distance to the viewing site. Critically, the stereo microscope relies on two optical feeds that allow the operating clinician to view the retina with depth perception. To modulate imaging magnification, different inbuilt lenses can be selected during the procedure by means of a control knob or pedal that comes with the system. Most recent systems feature focal lengths of 150–200 mm, allowing crisp visualization of the eye posterior. Fig. 0.3 provides a view upon the retina for different zoom factors. In addition, a secondary set of binoculars is often available by means of a beam splitter so that additional personnel can view the surgical procedure simultaneously.

Physically, stereo microscopes are mounted on the ceiling or suspended via a floor stand arm. They come with a dedicated foot pedal to control specific functionalities—including precise placement of the stereo-microscope, and changing of focus or zoom—with the benefit of providing the operating clinician maximal freedom with their hands.

Figure 0.3 Field of view from microscope. Retina visualization with stereo microscope and two different zoom factors. Surgical tweezers are used to delicately interact with the retina.

2.1.2. Additional Lenses

In addition to the optical system in the stereo microscope, it is common to use an additional lens during procedures in order to provide a wider field of view or improve visualization at dedicated locations of the retina. In practice, the choice of this additional lens is based on the surgical task in question. We briefly discuss some of the choices common to retinal microsurgery.

In practice, there are two types of additional lenses used: non-contact and contact lenses. As the name indicates, the difference in these lies in whether or not the lens is touching the cornea. In the case of non-contact lenses, these are typically attached to the microscope itself by means of an adapter that can be moved in and out of the viewing path manually. In contrast, contact lenses are placed in physical contact with the eye during dedicated portions of the procedures. These are typically hand held by an assistant while in use or directly sutured to the eye. Both types have their advantages: non-contact lenses are convenient as they do not require an additional personnel or cause trauma to the eye, but they are not always properly aligned with viewing region under consideration; conversely, hand-held or sutured lenses provide improved viewing comfort but require an additional hand.

In terms of visualization, additional lenses serve two important purposes. The first is to provide a wider field of view that can range up to 130° of view (e.g., BIOM or Eibos). Such wide-angle lenses are common during vitrectomy procedures. In contrast, for procedures related to the macula such as ILM or ERM peeling, lenses that provide smaller fields of view with greater resolution are often preferred. Perhaps the most popular of this kind is the Machemer lens that provides a highly magnified 30° field of view.

2.1.3. Light Sources

In order to see the surgical site, light from the exterior must be directed onto the retina. A variety of options now exist to do so, and the use of multiple illumination types during a single procedure is common. However, an important risk factor and consequence of used illumination systems is induced retinal phototoxicity. First reported in 1966 in patients having undergone cataract surgery, phototoxicity can be either thermal or photochemical in nature from excessive ultraviolet (UV) or blue light toxicity. Reports indicate that roughly 7% of macular hole repair patients have experienced significant phototoxicity. As such, the operating clinician is always forced to compromise illumination with patient safety.

As a primary illumination system, an integrated light source is already available with the surgical system itself. This light source is coaxial with the microscope’s optical system, allowing the light source to travel the same path as the viewing path, which reduces shadowing affects.

Alternatively, endoilluminators are fiber-optic light pipes inserted through one of the trocars in the eye sclera. Most common in surgical practice are two types of light sources for such light pipes: xenon and halogen. Although both have the potential to induce phototoxicity, both are considered safe. Light pipes of this nature come in 20, 23, and 25 gauge sizes, providing a spectrum of pipe stiffness useful for eye manipulations during procedures. Today, such illumination pipes provide cone-like illuminations of up to 40° to 80° angles depending on the system.

Naturally, a consequence of the light pipe endoilluminator is that the operator physician is forced to use one hand to manipulate this light source during the procedure. While this can be effective to augment depth perception (via an instruments project shadow on the retina) or improve illumination of specific retinal regions, chandelier illuminations offer an alternative and give the clinician freedom in both hands. Chandelier endoilluminators provide excellent wide-angle illumination in situations where bimanual surgical manoeuvres are necessary.

2.1.4. Additional Imaging

Prior to the surgical intervention, an important aspect is to visualize what areas of the retina should be manipulated during an envisioned procedure. To do this, a variety of imaging devices and modalities are typically used in routine clinical care. These include but are not limited to:

- **Color Fundus Imaging** relies on a digital camera, with an electronic control of focus and aperture to image a 30° to 50° field of view of the retina. The technology dates back to the 1880s and can used to capture over 140° for peripheral imaging using additional lenses. Nowadays, acquiring color fundus images is easy and relatively inexpensive method to diagnose, document, and monitor dis-

eases affecting the retina. Variants to color fundus photography such as red-free imaging, which enhance the visibility of retinal vessels by removing red wavelengths, are also common.

- **Fluorescein Angiography** is similar to color fundus photography except that it takes advantage of different filters and fluorescein intravenous injections to produce high-contrast images at the early stages of an angiogram. By using the camera light flashes, which is excited using a filter and then absorbed by the fluorescein, blood flow regions of the vasculature are strongly highlighted. This can then be recorded via the camera and help depict the complete vasculature of the retina. Such imaging is extremely effective in identifying regions of the retina that have venous occlusions and other related pathologies.
- **Optical Coherence Tomography** is a fast and non-invasive imaging modality that can acquire micrometer-resolution three-dimensional scans of the anterior and posterior segments of the eye. Since its introduction in 1991, it has become one of the most widely used diagnostic techniques in ophthalmology. Today, OCT is used to diagnose and manage a variety of chronic eye conditions, as it provides high-resolution imaging and visualization of relevant biomarkers such as inter- or subretinal fluid buildup, retinal detachments, or pigment epithelium detachments. In addition, it enables careful measurement of retinal thickness, which can be important during retinal detachment or macular hole repair procedures. Optical Coherence Tomography Angiography (OCT-A) can also be used to yield 2D volumes of the vasculature, bypassing fluorescein injections. Similarly, Doppler OCT can be used to quantify blood perfusion. Given its strong clinical relevance and its pertinent role in the future of robotic retinal surgery, the following sections will describe OCT in detail.

2.2. Real-Time OCT for Retinal Surgery

Retinal surgery requires both visualization and physical access to limited space in order to perform surgical tasks on delicate tissue with micrometer scale. When it comes to viewing critical parts of the surgical region and to working with micrometer accuracy, excellent visibility and precise instrument manipulation are essential. Conventionally, visualization during microsurgery is realized by surgical microscopes, as shown in Fig. 0.2, which limits the surgeon’s field of view and prevents perception of micro-structures and tissue planes beneath the retinal surface. The left image in Fig. 0.2 and both sides of Fig. 0.3 show a typical microscope view of the retina surface during ILM peeling. The entire thickness of human retina, which consists of 12 layers, is only about $350\mu\text{m}$ and the ILM is as thin as 1 to $3\mu\text{m}$ [38]. Therefore, even with the

Figure 0.4 Diagnostic imaging modalities. Fundus color photography (upper left), fluorescein angiography (upper right), and optical coherence tomography (lower) are pre-operative imaging modalities commonly used before retinal interventions.

aid of advanced surgical microscope systems, such operation is extremely challenging and requires rigorous and long-term training for the retinal surgeons.

So far, several well developed imaging modalities such as magnetic resonance imaging (MRI), X-ray computed tomography (CT), and ultrasound sonogram (US) have been utilized in image-guided intervention for various kinds of surgeries [151]. However, these conventional imaging modalities are not suitable for retinal surgery because their resolution is too low, which prevents resolving the retinal microstructures. The slow imaging speed is problematic here as well. In recent years, OCT emerged as a popular intraoperative imaging modality for retinal surgery. OCT systems are now capable of achieving high speed imaging in excess of 100 cross-sectional images per second, large imaging depths of a few millimeters, and micrometer-level transverse and longitudinal resolution [22, 101].

OCT systems evolved rapidly over the past thirty years, and currently there are many different types of commercial systems in the market. Below follows a short description of each type.

- **TD OCT:** Time-domain OCT is the first variant of OCT that achieves depth scanning (i.e., A-scan imaging) by physically translating the position of a reference plane in function of the depth of the imaging layer that one wants to visualize. To detect the signal, a simple photodetector directly captures the intensity of the interference signal. Because the reference plane can be translated over a long distance using mechanical stages, a very long imaging depth typically on the order of several centimeters to tens of centimeters can be achieved. However, the typical A-scan speed is less than 1 kHz. Therefore the major drawbacks of TD OCT systems are slower scanning speed and low signal-to-noise ratio (SNR).
- **FD OCT:** Unlike TD OCT, frequency-domain OCT (FD OCT) systems perform spectral measurements and the depth information is deduced from Fourier transforming the OCT spectral data. Since FD OCT does not need the physical movement of the reference plane, it can be made high-speed. Furthermore, the use of spectral measurements significantly improve the SNR compared to TD OCT [42, 144]. FD OCT system characteristics are described in detail in the next section.
- **SD OCT:** Spectral-domain OCT is the original variant of FD OCT that uses a spectrometer and a broadband light source to measure the OCT spectral interference signal. Most commercial OCT systems are SD OCT type and generally operate with A-scan speeds in the range of 70 Hz to 20 kHz. SD OCT systems exhibit significant improvements in SNR compared to TD OCT and allow high-speed OCT imaging where the imaging speed depends on the speed of a line-scan camera used in the spectrometer.
- **SS OCT:** The latest development in OCT technology is swept-source OCT. It uses a wavelength-swept laser and a high-speed single photodetector to measure the OCT spectral interference signal. Typical commercial versions exhibit A-scan speeds in the range of 50 to 200 kHz. Typically SS OCT systems are faster, exhibit larger imaging depth, and offer higher SNR compared to SD OCT. However, they are more expensive than SD OCT. For example, a typical swept source OCT engine operating at 100 kHz would cost approximately 30,000 dollars whereas a 70 kHz OCT spectrometer engine would be in the 10,000 dollar range.
- **iOCT:** Intraoperative OCT generally refers to an FD OCT system integrated into a surgical microscope that allows OCT visualization during surgical proce-

dures. Typical commercial iOCT systems provides realtime B-mode (i.e., cross-sectional) images. A post-processed C-mode (i.e., volumetric) image can be typically generated in a few seconds. Several companies provide iOCT as an option for their high-end surgical microscope systems.

- **CP OCT:** Common-path OCT, unlike the standard OCT systems that use Michelson interferometer setup, does not have a separate reference arm [172, 100]. Instead it uses the signal arm as the reference arm and the reference signal is produced from the distal end of the signal arm. Therefore, the signal and the reference beam mostly share the same beam path. This allows a much simpler system design, lower associated costs, and the ability to use interchangeable probes, as well as the freedom to use any arbitrary probe arm length. CP OCT is also immune to polarization, dispersion effects, and fiber bending. This makes CP OCT systems ideal for endoscopic applications [100].
- **FD CP OCT:** The Fourier domain common-path OCT is the Fourier Domain variant of CP OCT.

2.3. Principle of Fourier Domain OCT

FD OCT was first described by Fercher *et al.* at 1995 [57]. Over the past two decades [42, 144, 206, 217, 218] it has been developed rapidly and most of the commercial OCT systems are of this type. Compared to TD OCT, FD OCT has more than two orders of magnitude higher sensitivity and significantly faster imaging speed [42] with the typical A-scan imaging speed in the order of a few hundred kHz. There are two different types of FD OCT as mentioned above: spectral-domain OCT (SD OCT) which uses a broadband light source and a dispersive spectrometer with a line-scan array detector, and swept-source OCT (SS OCT) which uses a narrow-single-wavelength swept laser with a high-speed PIN detector.

Figure 0.5 shows the schematic layout and signal processing steps of a typical spectrometer based FD OCT (i.e., SD OCT). The spectrometer in SD OCT uses a diffraction grating that disperses the broad-band light, several collimating lenses, and a high-speed line-scan CCD or CMOS camera to detect the spectrum of the OCT signal. The signal arriving at the line-scan camera is the combined interferogram of the light waves from different depths within the sample. The resultant signal spectrum

Figure 0.5 A schematic of SD OCT. A typical layout of Fourier Domain OCT system based on a spectrometer (i.e., SD OCT) is shown schematically with simplified signal processing steps

$I_D(k)$ can be written as [48]:

$$\begin{aligned}
 I_D(k) = & \frac{\rho}{4} \left[S(k) \left(R_R + \sum_{n=1}^N R_{S_n} \right) \right] \\
 & + \frac{\rho}{8} \left[S(k) \sum_{n=1}^N \sqrt{R_R R_{S_n}} \cos(2k(z_R - z_{S_n})) \right] \\
 & + \frac{\rho}{8} \left[S(k) \sum_{m \neq n=1}^N \sqrt{R_{S_m} R_{S_n}} \cos(2k(z_{S_m} - z_{S_n})) \right] \quad (0.1)
 \end{aligned}$$

where k is the wavenumber, $S(k)$ is the power spectrum of the light source, R_R is the power reflectivity of the reference mirror, R_{S_i} is the power reflectivity of the i^{th} layer

of the sample. The depth profile or A-scan image of the sample can be obtained by taking the Fourier transform of the spectrum in (0.1). This results in a spatial domain A-scan image which can be expressed as:

$$\begin{aligned}
 i_D(z) = & \frac{\rho}{8} \left[\gamma(z)(R_R + \sum_{n=1}^N R_{S_n}) \right] && \text{DC terms} \\
 & + \frac{\rho}{8} \left[\gamma(z) \otimes \sum_{n=1}^N \sqrt{R_R R_{S_n}} \delta(z \pm 2(z_R - z_{S_n})) \right] && \text{cross-correlation terms} \\
 & + \frac{\rho}{8} \left[\gamma(z) \otimes \sum_{m \neq n=1}^N \sqrt{R_{S_m} R_{S_n}} \delta(z \pm 2(z_{S_m} - z_{S_n})) \right] && \text{auto-correlation terms}
 \end{aligned}$$

where $\gamma(z)$ is the Fourier transform of $S(k)$. The ‘DC terms’ correspond to the spectrum of the light source. Usually, this is the largest component of the detector signal, which needs to be subtracted before A-scan images can be displayed. The ‘cross-correlation terms’ is the terms that forms desired OCT A-scan image. It contains several peaks whose locations are determined by the distance offset from the reference mirror position z_R and the target positions z_S . The amplitude of these peaks changes according to the light source power and the reflectivity of the reference and the target positions within the sample. The last component, the ‘auto-correlation terms’, come from the interference of the light between different reflectors within the target. This results in ghost image artifact. However, this component is usually located away from the desired signal, since the distances between the different reflectors within the sample are small.

The OCT signal can be visualized as a depth-resolved 1-D image (A-Mode), a cross-sectional 2-D image (B-Mode), or a volumetric 3-D image (C-Mode); schematically shown in Fig. 0.6. In most SD OCT systems, the signal is detected as a spectral modulation using a spectrometer which samples them uniformly in wavelength, and this can be described as in (0.1). This implies that they are nonlinear in wavenumber domain. Thus, applying the discrete Fourier transform or fast Fourier transform to such signal will seriously de-grade the imaging quality. A specific procedure, both in hardware and software, has been developed to reconstruct the image from the non-linear wavenumber domain spectrum. Compared to the hardware solutions that usually complicate the design of the spectrometer and increases the cost, the software solutions are usually much more flexible and cost-efficient. There are two widely used software methods: the first one is based on numerical interpolation that includes various linear interpolations and cubic interpolation; the other one uses the non-uniform discrete Fourier transform or the non-uniform fast Fourier transformation.

Figure 0.6 OCT imaging modes. Three different scanning/imaging modes of OCT are schematically described: A-scan (1-D), B-scan (2-D), and C-scan (3D)

2.3.1. Axial Resolution of SD OCT

The OCT light source having a Gaussian spectral shape with a bandwidth $\Delta\lambda$ for wavelength and Δk for wavenumber, can be described mathematically as:

$$S(k) = \frac{1}{\Delta k \sqrt{\pi}} e^{-\left[\frac{(k-k_0)}{\Delta k}\right]^2} \quad (0.2)$$

Here, k_0 is the center wavenumber. It can be shown that its Fourier transform $\gamma(z)$ is:

$$\gamma(z) = e^{-z^2 \Delta k^2} \quad (0.3)$$

From (0.2) the A-scan signal is the convolution of $\gamma(z)$ and the samples structure function $\delta(z \pm 2(z_R - z_S))$. Thus, the resolution l_{axial} of the SD OCT in axial direction can be defined as the full width at half maximum (FWHM) of $\gamma(z)$:

$$l_{axial} = \frac{2\sqrt{\ln(2)}}{\Delta k} = \frac{2\ln(2)}{\pi} \frac{\lambda_0^2}{\Delta\lambda} \quad (0.4)$$

Here, λ_0 is the central wavelength of the light source. As you can see the axial resolution of the OCT is determined by the bandwidth of the light source. Thus a broad band light source is usually used in SD OCT system to achieve high resolution imaging.

2.3.2. Lateral Resolution of SD OCT

In SD OCT, the lateral resolution is defined as the Full-Width-Half-Max (FWHM) of the point spread function of the probe beam at the beam waist. Assume the numerical aperture of the objective lens before the sample is denoted as NA. Then the lateral resolution of SD OCT can be expressed as:

$$l_{lateral} = \frac{2\sqrt{\ln(2)}}{\pi} \frac{\lambda_0}{NA} \quad (0.5)$$

2.3.3. Imaging Depth of SD OCT

In SD OCT, the imaging depth is influenced by two factors. The first is the sample's scattering and absorption. This causes the light intensity to decrease exponentially with depth. Another factor is the spectrometer's spectral resolution. It is determined by the light bandwidth, Δk and the number of the pixels in the line-scan camera, which is denoted as N . Based on the Shannon/Nyquist theory, the maximum imaging depth of SD OCT system limited by the resolution of the spectrometer is given by:

$$z_{max} = \frac{N\pi}{2\Delta k} \quad (0.6)$$

Equation (0.4) shows that the axial resolution of SD OCT is inverse proportional to the bandwidth of the light source. Thus, both high-resolution and large bandwidth spectral measurements are needed for SD OCT imaging that requires both large imaging depth and high axial resolution. This requires a large linear array camera which can be quite expensive. In addition, a slow sampling rate will increase the imaging time, which makes the imaging susceptible to motion artifacts. It also produces a large amount of data that becomes a heavy burden on the image storage and transferring.

2.3.4. Sensitivity of SD OCT

The sensitivity of an SD OCT system can be expressed as [114]:

$$\sum_{SDOCT} = \frac{\frac{1}{N}(\frac{\rho\eta T}{h\nu_0}P_0)^2\gamma_r\gamma_sR_r}{\frac{\rho\eta T}{h\nu_0}\frac{P_0}{N}\gamma_rR_r[1 + \frac{1+\Pi^2}{2}\frac{\rho\eta}{h\nu_0}\frac{P_0}{N}\gamma_rR_r\frac{N}{\Delta\nu_{eff}}] + \sigma_{rec}^2} \quad (0.7)$$

Here, N is the number of pixels obtained at the detector, ρ is the efficiency of the spectrometer, η denotes the quantum efficiency of the detector, T is the CCD/CMOS detector integration time, h is Planck's constant, ν_0 is the center frequency, P_0 is the output of the source power and γ_R and γ_S are the parts of the input power that enter

the spectrometer from the reference and sample arms, respectively. R_r is the power reflectivity of the reference mirror, Π is the polarization state of the source, $\Delta\nu_{eff}$ is the effective spectral line width of the light source and σ_{rec} is the RMS of the receiver noise. The 3 terms in the denominator of (0.7) have different meanings: the first is the shot noise; the second the excess noise and the third the receiver noise.

2.4. High-Speed OCT using GPU Processing

Due to their fast working speed OCT systems are suitable for use as clinical interventional imaging systems. To provide accurate and timely visualization, real-time image acquisition, reconstruction and visualization are essential. However, in current ultra-high speed OCT technology, the reconstruction and visualization speeds (especially 3D volume rendering) are generally far behind the data acquisition speed. Therefore, most high-speed 3D OCT systems usually work in either low-resolution modes or in a post-processing mode, which limits their intraoperative surgical applications. To overcome this issue, several parallel processing methods have been implemented to improve A-scan data of FD OCT images. The technique that was adopted by most commercial systems is based on multi-core CPU parallel processing. Such systems have shown to achieve 80,000 line/s processing rate on nonlinear-k polarization-sensitive OCT systems and 207,000 line/s on linear-k systems, both with 1024-point/A-scan [117, 153]. Nevertheless the CPU based processing is inadequate for realtime 3D video imaging even a single 3D image display can take multiple seconds. To achieve ultrahigh speed processing, GPGPU (acronym for General Purpose Computing on Graphics Processing Units) based technology to accelerate both the reconstruction and visualization of ultra-high speed OCT imaging [87, 218, 219].

The signal processing flow chart of the dual-GPUs architecture is illustrated in Fig. OCT GPU, where three major threads are used for the FD OCT system raw data acquisition (Thread 1), the GPU accelerated FD OCT data processing (Thread 2), and the GPU based volume rendering (Thread 3). The three threads synchronize in the pipeline mode, where Thread 1 triggers Thread 2 for every B-scan and Thread 2 triggers Thread 3 for every complete C-scan, as indicated by the dashed arrows. The solid arrows describe the main data stream and the hollow arrows indicate the internal data flow of the GPU. Since the CUDA technology currently does not support direct data transfer between GPU memories, a C-Scan buffer is placed in the host memory for the data relay [219]. Such dual-GPUs architecture separates the computing task of the signal processing and the visualization into different GPUs, which has the following advantages: *1)* Assigning different computing tasks to different GPUs makes the entire system more stable and consistent. For the real-time 4D imaging mode, the volume rendering is only conducted when a complete C-scan is ready, while B-scan frame processing is running continuously. Therefore, if the signal processing and the visu-

Figure 0.7 Signal processing flow chart of Ultra-high speed OCT based on dual-GPUs architecture. Dashed arrows, thread triggering; Solid arrows, main data stream; Hollow arrows, internal data flow of the GPU. Here the graphics memory refers to global memory.

alization are performed on the same GPU, competition for GPU resource will happen when the volume rendering starts while the B-scan processing is still going on, which could result in instability for both tasks; 2) It will be more convenient to enhance the system performance from the software engineering perspective. For example, the A-scan processing could be further accelerated and the point spread function (PSF) could be refined by improving algorithm with GPU-1, while more complex 3D image processing task such as segmentation or target tracking can be added to GPU-2.

Fig.0.8 provides an overview of a stereo-microscope with iOCT as well as a pair of digital cameras allowing simultaneous capturing of both the pair of stereo-images as well as intraoperative OCT-images. The iOCT and digital cameras share a large part of the optical path. This is convenient as zoom adjustments will be equally reflected in the stereo-camera as on the iOCT scanner. While such and similar layouts offer powerful measurement tools for capturing the retina, the quality still depends on the state and alignment of all intervening media. Advanced instruments described next (§3) by-pass these problems by directly measuring inside the patient’s eye.

Figure 0.8 Layout of a stereo-microscope with iOCT and digital cameras. Frontal view upon a commercial stereo-microscope with mounts for digital cameras and iOCT.

3. Advanced Instrumentation

Over the last decades advances in instrumentation have significantly altered retinal surgery practice. The development of *pars-plana vitrectomy* in the 1970's by Machernner formed a key milestone [122]. Kasner had discovered in 1962 that the vitreous humour, the clear gel-like structure that occupies the intraocular space between the lens and retina (Fig. 0.1), could be removed a.o. providing unrestricted access to the retina [26, 102]. Machernner developed *pars-plana vitrectomy* a minimal invasive technique to remove the vitreous humour. In this technique a so-called *vitrectome* is introduced at a distance of 3–4 mm from the limbus, the place where the cornea and sclera meet. This region, the so-called *pars plana*, is free from major vascular structures. The retina typically starts at 6–8 mm posterior to the limbus. There is thus little risk for retinal detachment when making an incision in the pars plana to create access to the intraocular space [26]. A vitrectome, a suction cutter, is then used to remove the vitreous humour, which can be replaced by a balanced salt solution. Often a 3-port approach is adopted where, aside for the vitrectome, a second incision is made to connect a supply line to provide at constant pressure the salt solution. A third incision is used to pass a light guide to provide local illumination. Vitrectomy clears the path for other instruments to operate on the retina. Modern retinal instruments include retinal picks, forceps, diamond-dusted membrane scrapers, soft tip aspiration cannulas, cauterization tools, coagulating laser fibers, chandeliers (illuminating fibers), and injection needles. There is a trend to miniaturize these instruments, with a particular focus on the diameter. In retinal surgery the instrument diameter is expressed in the

System.

Table 0.4 Lookup table - instrument dimensions from Birmingham Wire Gauge

Gauge	20	21	22	23	24	25	26	27	28	29
mm	0.889	0.813	0.711	0.635	0.559	0.508	0.457	0.406	0.356	0.330

Birmingham Wire Gauge (BWG) system, often simply termed *Gauge* and abbreviated as G; Table 0.4 shows the corresponding dimensions in millimeter. When the diameter drops to 25 G (0.5 mm) the required incisions become self-sealing so that there is no need to suture the incisions, and the risk for inflammation is reduced. However, the 25 G instruments are more compliant and may bend (e.g., when trying to reposition the eye). Some retinal surgeons therefore prefer larger and stiffer 23 G (0.65 mm) instruments. Next to the more ‘traditional’ instruments various sophisticated instruments, featuring integrated sensing capability (§3.1, §3.2) as listed in Table 0.5 or enhanced dexterity (§3.4) have been reported. In contrast to methods relying on external measurement (§2) or actuation (§5) these instruments directly measure and act in the intraocular space, bypassing the complex optical path formed by the plurality of lenses and intervening media, and avoiding the effect of the scleral interface. Therefore they potentially allow a more precise acquisition, understanding and control over the interaction with the retinal tissue.

Table 0.5 Overview of sensor-integrated vitreo-retinal instruments

Measurand	Technology [Reference]
retinal interaction force	strain gauge [20, 21, 73]; Fabry-Perot [58, 119]; FBG [60, 66, 81, 82, 95, 110, 146, 176, 184]
scleral interaction force	FBG [79, 80, 86]
proximity, depth	OCT [6, 74, 99, 146, 176]
puncture	impedance [170]
oxygen	photoluminescence [54]

3.1. Force Sensing

Fragile structures at the retina may get damaged when undergoing excessive forces. As these excessive forces often lie below human perceptual thresholds [73], this is not an imaginary problem. Gonenc *et al.* describe iatrogenic retinal breaks, vitreous hemorrhage as well as subretinal hemorrhages following peeling procedures [66]. When too large forces are applied on a cannulated vessel it may tear or get pierced. This may lead to serious bleedings, or unintentional injection of a thrombolytic agent in subretinal layers which would cause severe trauma. Over the years researchers presented several sensors for measuring the force applied on retinal structures Fig.0.9 shows a

Figure 0.9 Force sensing for retinal surgery. Evolution of force sensing over the last years and integration of force sensing technology in instruments for retinal surgery.

time-line with some developments in this regard.

3.1.1. Retinal Interaction Forces

Gupta and Berkelman employed strain gauges glued upon or integrated in the instrument handle [20, 21, 73]. These early works provided a first insight in the governing interaction forces with the retina. Gupta showed that for 75% of the time interaction forces stayed with 7.5mN below human perceptual thresholds. Berkelman developed a 3-Degree-of-Freedom (DoF) sensor based on a double cross flexure beam design [20, 21]. Aside from submillinewton precision, Berkelman’s sensor behaves isotropic in 3 DoFs. With a 12.5mm outer diameter this sensor can only be integrated in the instrument handle. The sensor therefore does not only pick up the interaction forces at the retina, but also forces that develop at the incision in the sclera. Since the latter are typically an order of magnitude larger [96] it is difficult to estimate the interaction forces at the retina. Therefore researchers searched for embedding sensors in the shaft of the surgical instruments to measure directly in the intraocular space.

The first intraocular force sensor by Sun *et al.* employed FBG (Fiber Bragg Grating) optical fibers [183]. Fiber optical sensors are attractive as they can be made very small, immune to electrical noise, and are sterilizable [119]. Sun started with a single 160 μm FBG strain sensor. The sensor was glued in a square channel manufac-

tured along the longitudinal axis of a 0.5 mm diameter titanium wire, mimicking 25G ophthalmic instruments [183]. The sensitive part of the optical fiber, i.e. a 1cm long portion where the Bragg grating resides, was positioned nearby the distal instrument tip such that interaction forces at the sclera would not be picked up. A measurement resolution of 0.25mN was reported. During experiments on fertilized chicken eggs forces between 8 – 12mN were found when peeling the ISM. Forces in the range of 1 – 3mN were measured during vessel cannulation experiments. In a follow-up work a 2-DoF version was developed by routing three fibers along the longitudinal axis of a 0.5mm diameter instrument [95]. This sensor measures transverse force components perpendicular to the instrument axis. Through differential measurement the effect of temperature could be canceled out. Experiments were conducted on respectively a raw egg membrane, a chorioallantoic membrane (CAM), which is an extraembryonic membrane of a fertilized chicken egg [115], and on a rabbit [184]. Peeling forces varied between 0.2 – 1.5 mN, 1.3 – 4.1 mN, for the rabbit an average minimum force to delaminate the hyaloid was 6.7 mN. Damage appeared in the CAM model when forces exceeded 5.1 mN, whereas retinal tears occurred in the rabbit from forces beyond 6.4 mN. Gonenc *et al.* used a similar design on a hook with the Micron, a handheld robotic instrument, to measure forces while peeling different membrane models [66]. He *et al.* introduced a 3-DoF force sensing pick which is also sensitive to axial forces. A super-elastic Nitinol flexure is foreseen to make the tool sensitive in axial direction [82]. RMS error was below 1 mN in all directions [82].

Several works introduced micro-forceps with integrated force sensing [67, 81, 82]. Gonenc *et al.* foresee a method to ensure that grasping does not affect the force reading [67]. Kuru *et al.* developed a modular setup allowing exchange of the forceps within a non-disposable force-sensing tube [110]. Gijbels *et al.* developed a stainless steel cannulation needle with 2-DoF FBG sensing and a 80 μ m needle tip [60, 64]. The sensor resolution was 0.2mN. Repeatable force patterns were measured when cannulating synthetic vessels in a PDMS retina [63] and cadaver pig eyes [64]. Whereas the force sensor is only sensitive to transverse forces typical cannulation needles, including those from Gijbels [60] and Gonenc *et al.* [68], have a distal tip that is bent under an angle close to 45% to ease cannulation [55]. This angulation renders the sensor also sensitive to puncture forces which are hypothesized to mainly occur in the direction of the needle tip. Gonenc *et al.* mounted a force-sensing microneedle on the Micron handheld robotic system and cannulated vessels on a CAM surface. They reported cannulation forces rising from on average 8.8 mN up to 9.33 mN for increasing speed of respectively 0, 3 to 0, 5mm/s [68]. In Gijbels’ work cannulation forces ranged between 0.6mN and 17.5mN, but in 80% of the cases they were below 7mN [64].

Whereas the majority of works involve sensors based on FBGs, a number of instruments have been presented that employed the Fabri-Pérot Interferometry (FPI) measurement principle [58, 119]. With FPI light exiting an optical fiber scatters back

between reflective surfaces at both sides of a flexure body. Depending on the load the flexure deforms affecting the back-scattered light. FPI is in general more affordable than FBG, but manufacturing precise flexures is challenging. A further challenge exists in making sure the instrument operates robust despite the flexure. Liu *et al.* used a common-path Fourier domain optical coherence tomography to interrogate the change of cavity length of an FP cavity. By multiplexing three signals they constructed a 3-dimensional force sensor with diameter below 1mm on a retinal pick [119]. Fifanski *et al.* built also a retinal pick with force sensing based on FPI. This instrument has only 1 DoF but is with 0.6mm Outer Diameter (O.D.) close to current practice [58].

3.1.2. Scleral Interaction Forces

An underappreciated limitation of current robotic systems is the lost perception of forces at the scleral interface where the tool enters the eye. During freehand manipulation, surgeons use this forces to orient the eye or to pivot about the incision such as to limit stress and/or keep the eye steady. In robot-assisted retinal surgery the stiffness of the robotic system attenuates the users perception of the scleral forces [80]. This may induce undesired large scleral forces with the potential for eye injury. For example, Bourla *et al.* reported excessive forces applied to the sclera due to misalignment of the RCM of the Da Vinci [24]. He *et al.* developed a multi-function force sensing ophthalmic tools [79, 80] that simultaneously measures and analyzes tool-tissue forces at the tool tip and at the tool-sclera contact point. A robot control framework based on variable admittance uses this sensory information to reduce the interaction force. He *et al.* reported large scleral forces exceeding 50mN and tool deflection complicating positioning accuracy if no care was paid to the scleral interaction forces, where forces dropped to 3.4mN otherwise [80]. Following the same control-framework, a force-sensitive light-guide was developed by Horise *et al.* that can accommodate to the patient's eye motion. Such smart light-guide could support bimanual surgery as the microsurgeon can use his/her second hand to manipulate other instruments instead of the light-guide [86].

3.1.3. Force Gradients

Instead of measuring the absolute interaction force for some applications such as detection of puncture or contact state it is more robust to look at relative changes rather than absolute forces. For example Gijbels *et al.* and Gonenc *et al.* look at the force transient to detect the puncture of a retinal vein [63, 64, 68]. A threshold of $-3mN/s$ was found to be able to detect punctures with 98% success rate [64]. In 12% of the cases a false-positive detection was made e.g. when upon retraction the threshold was hit. Double-punctures, i.e. where the vein is pierced through, were also successfully detected as such would lead to two rapidly succeeding force transients.

3.2. Optical Coherence Tomography

Force or force gradients can help improve understanding of the current state, but offer little help to anticipate upon a pending contact or state transition, neither do they provide a lot of insight in what is present below the surface. SD OCT systems (§2.3) achieve $< 5\mu\text{m}$ axial resolution in tissue [189] and have imaging windows larger than 2 – 3mm. As such they are considered very useful to enhance depth perception in retinal applications. Several researchers have developed surgical instruments with integrated optical fibers to make this imaging technology available at the instrument tip. The fibers may be directly connected to an OCT-engine or when using iOCT systems they may be routed via an optical switch to the OCT-engine, whereby the switch allows re-routing of the OCT signal to the fiber and alternatively to the intra-operative scanner [105]. The single fiber is typically introduced in a slot along the longitudinal direction of the surgical instrument and inserted alongside the instrument into the eye. The single fiber can then be used to generate an axial OCT-scan or A-scan (Fig.0.6) that provides information of the tissue and tissue layers in front of the OCT-beam that radiates within a narrow cone from the OCT fibre. By making lateral scanning motions, the axial OCT-scan can be used to create *B-scans* or *C-scans*.

Han *et al.* integrated a fiber optic probe with Fourier Domain Common Path OCT (FD CP OCT) [120] into a modified 25G hypodermic needle shaped as a retinal pick. They showed how the multiple layers of the neurosensory retina can be visualised through an A-scan and further reconstructed B-and C-scans from a rat cornea [74]. The fiber of Liu *et al.*[120] was also used by Yang *et al.* to generate B- and C-scans out of A-scans with the Micron, a handheld micromanipulator [208]. Balicki *et al.* embedded a fiber in a 0.5mm retinal pick for peeling the ERM [6]. The instrument was designed so that also the tool tip itself was visible in the A-scan. Through some basic filtering both the tool tip and the target surface could be extracted. In this layout registration is highly simplified as the distance to the target is simply the distance between the observed tip and the observed anatomical structure, hence omitting the need to calibrate the absolute tip location. Song *et al.* integrated an OCT-fiber in a motorized microforceps. It assesses the relative motion of the forceps relative to the target. The fiber is glued along the outside, fixed to one ‘finger’ of the forceps, such as to avoid artifacts when tissue is grasped [178]. Kang [99] developed a CP OCT-guided microinjector based on a similar piezo-actuation stage to conduct subretinal injections at specific depths.

Given the multi-layer structure of the retina, simple peak detection algorithms, may mistakenly provide the distance to a layer that differs from the retinal surface. More sophisticated algorithms were developed to take the multi-layer structure into account a.o. by Cheon *et al.* who proposed a shifted cross-correlation method in combination with a Kalman filter [41] or Borghesan *et al.* who compared algorithms based on an

Unscented Kalman Filter to an algorithm based on the Particle Filter [23]. Recently, within EurEyeCase, a EU-funded project on robotic retinal surgery [196], the first human experiments (5 subjects) with robot-assisted fiber-based OCT were conducted. A needle with integrated OCT-fiber was moved towards the retina. The feasibility of installing a virtual bound at a safe distance from the retina was confirmed [35]. In the same project cannulation needles featuring combined force and OCT-sensing were developed [176] and tested in-vivo on pig eyes [146]. In one of the configurations that were explored four grooves were made along the longitudinal axis of the instrument. In two grooves a pair of FBG optical fibers were inserted and glued. In one of the remaining grooves an OCT-fiber was glued. This latter was used to estimate the distance from the tip of the cannulation needle to the surface. The fiber OCT-fiber was retracted with respect to the cannulation needle such that even during cannulation the OCT-fiber tip was at a certain distance from the surface allowing estimation of the depth of the cannulation relative to the retinal surface.

3.3. Impedance Sensing

Several works have looked at electrical impedance sensing to measure different variables. In an early work Saito *et al.* used electrical conductivity for venipuncture detection in a rabbit [168]. A similar approach was followed by Schoevaerdt *et al.* [170] for eye-surgery. The goal was to estimate the contact with a retinal vessel and a shift in impedance when puncturing a retinal vessel. Similar to force sensing a change in impedance was expected to occur when the sensor passed from a pure vitreous-like environment, towards a contact with a vessel wall and subsequently the contact with blood in a cannulated vessel. Experiments on ex-vivo pig eyes showed a detection rate of 80%. The feasibility of detecting double punctures was also confirmed by Schoevaerdt. A side-product of the impedance sensing was found in the possibility to detect air bubbles in the supply line through which the medicine is to be injected [171]. Given the small size and fragile nature of the targeted vessels the presence of air in a supply line forms an important problem. The air may pass through the tiny lumen of a cannulation needle and end up in the targeted vessel. Due to the lower pressure in the vessel (compared to the high pressure to push the drugs through the needle tip), the air could rapidly expand inside as it could potentially damage the targeted vessel itself.

3.4. Dexterous Instruments

Where conventional retinal tools are generally straight, several instruments featuring distal dexterity have been designed up to now [19, 32, 84, 89, 94, 116, 155, 197, 199]. These instruments enter the intraocular space in a straight fashion but can then be reconfigured, taking on a curved or a bent shape inside the eye. Thanks to the distal dexterity a larger part of the targeted region can be reached with reduced risk for

Figure 0.10 Dexterous vitreoretinal instruments. overview of novel instruments featuring distal actuation capability.

colliding with the lens. Anatomic targets may also be reached under different angles. This in its turn can help reduce the force that is applied at the sclera.

Ikuta *et al.* developed a micro active forceps 1.4 mm in diameter, with built-in fiberscope [94]. In Ikuta’s design a sophisticated interface allows bending the 5mm distal segment over a range of 45 degrees while still allowing normal operation of the gripper. Wei *et al.* introduced a 550 μm pre-shaped super-elastic NiTi tube. Restrained by a cannula 0.91 mm in diameter, the tube enters the eye in a straight fashion. However the part that protrudes out of the cannula takes on its pre-shaped form once again. By regulating the relative displacement between NiTi tube and cannula the bending angle is adjusted [197]. Hubschman *et al.* developed a micro-hand of which each finger is 4 mm long and 800 μm wide and consists of 6 pneumatically actuated phalanxes that can bend 180 degrees and each lift up to 5 mN force [89]. A stent deployment unit (SDU) was a new development from Wei *et al.* [199]. The SDU consists out of three concentric NiTi elements: a preshaped outer tube with preshaped radius of 5 mm that bends at a specified angle when extended outside a stainless steel support tube; a stent pushing element to deliver the stent and a guidewire to puncture and create access to the vessel. With outer tool diameter of 550 μm the instrument is compatible with modern dimensions. The 70 μm guidewire was used to successfully cannulate vessels in agar and in a CAM model. The authors recommended smaller stents than the 200 μm that was employed to be able to target smaller structures.

Another example of an instrument with distal DoF is found in the IRIS system. The IRIS has an outer diameter of 0.9mm but features 2 distal DoFs, each with $\pm 90^\circ$ bending angle for only a 3 mm long section [84, 155]. Cao *et al.* developed an endo-illuminator with 2 bending DoFs. Shape Memory Alloy (SMA) is used as driving method. Despite good miniaturization potential the reported endo-illuminator was only a 10x scaled version of the targeted 25G design. More recently Lin *et al.* introduced a miniature forceps mounted on a concentric tube compatible with 23G vitreo-retinal surgery [116]. The gripper was actuated by a NiTi pull wire which is said not to interfere with the shape of the concentric tubes when actuated. The development of flexible instruments for retinal surgery is advocated a.o. by Bergeles *et al.* who computed a reduction of retinal forces in the order of 30% when moving flexible instruments through a model of the vitreous humour compared to steering rigid instruments through such environment [19]. The importance of this work is to be seen in light of the growing interest towards vitrectomy-less interventions [180].

4. Augmented Reality

Figure 0.11 OCT Volume. High quality volume rendering of an intraoperative OCT cube from a patient with macular foramen (left); and with surgical instrument (right).

During an intervention surgeons tend to immerse themselves mentally into the intraocular space, favoring visual above non-visual sensory channels. Under the assumption that visual feedback would minimally distract the natural flow of the operation, augmented reality has been explored extensively to convey additional contextual information. However, augmentation of information has several specific challenges: firstly, the processing and rendering of the data has to be performed efficiently to provide timely feedback to the user. This is especially important for situations where the additional data directly provides surgical guidance, as in these cases any delay introduced by the visualization and augmentation would create lag and negatively affect the surgical performance. Assuming the needle movement is 1 mm/s, each 10 ms of delay

in one control loop will bring $10\ \mu\text{m}$ position error. Secondly, identification of the required information to be augmented in each surgical steps as well as registration of multimodal images are not straightforward. As a third challenge, the visualization and rendering methods for 3D volumetric data are highly demanding when it comes to computational performance, especially when a high visual fidelity is to be achieved.

Advanced rendering methods which apply realistic illumination simulation produce high-quality results, such as the OCT volume in Fig.0.11 rendered with Monte-Carlo volume raycasting. These have been shown to improve perception of important structures; however, it takes currently several seconds to generate these images and thus is not directly applicable to real-time imaging. Therefore, optimizing approaches for fast rendering and high-quality augmentation is an important research task.

This section explains mosaicing, subsurface imaging, depth visualization and overlaying of pre- and intra-operatively acquired data. Furthermore, a novel approach using auditory feedback as a form of augmentation will be discussed.

4.1. Mosaicing

Figure 0.12 **Mosaicing**. Mosaicing result obtained from a set of nine images [31].

The acquisition of high-resolution retinal images with a large field of view (FOV) is challenging due to technological, physiological and economic reasons. The majority of imaging devices being used in retinal applications are slit lamp biomicroscopes; OCT machines and ophthalmic microscopes visualize only a small portion of the retina, complicating the task of localizing and identifying surgical targets, increasing treatment duration and patient discomfort. To optimize ophthalmic procedures, image processing and advanced visualization methods can assist in creating intra-operative retina maps for view expansion. An example of such mosaicing methods, described in [157], is a combination of direct and feature-based methods, suitable for the textured nature of the human retina. The researchers in this work described three major enhancements to the original formulation. The first is a visual tracking method using local illumination compensation to cope with the challenging visualization conditions. The second is an efficient pixel selection scheme for increased computational efficiency. The third is an entropy-based mosaic update method to dynamically improve the retina map during exploration. To evaluate the performance of the proposed method, they conducted several experiments on human subjects with a computer-assisted slit-lamp prototype. They also demonstrated the practical value of the system for photo documentation, diagnosis and intraoperative navigation.

4.2. Subsurface Imaging

Recent ophthalmic imaging modalities such as microscope integrated OCT enable intraoperative visualization of microstructural anatomies in sub-tissue domains. Therefore, conventional microscopic images are subject to modification in order to integrate visualization of these new modalities. Augmentation of iOCT data on en-face images to the surgeon comes with challenges including instrument localization and OCT-optical image registration (see Fig.0.13). Studies describe robust segmentation methods to obtain the needle point cloud within the OCT volume and use retinal vessel structures for online registration of OCT and optical images of the retina [221]. Due to the infrared light source of the OCT and using the geometrical features of the surgical instruments, segmentation results are robust to illumination variation and speck reflection.

4.3. Depth Visualization

In the conventional vitreo-retinal surgeries one of the important weaknesses is the lack of intraocular depth. Currently, surgeons rely on their prior experience to approximate the depth from the shadow of their instrument on the retina. Recent studies show that modern intraoperative imaging modalities such as iOCT are able to provide accurate depth information. Therefore, augmented reality can play an important role here to intuitively visualize the depth information.

Figure 0.13 Screenshot of injection guidance application. Left: Augmented view of the surgical scene, showing the camera view with the overlaid OCT scanning locations as well as the projected intersection point with the RPE layer. Current and last B-Scan are marked with white and blue bars for illustrative purposes; Right: schematic view of the 3D relationships between B-Scans (blue), current needle estimate (green), and intersection point with the target surface (red). These relationships cannot easily be inferred from a simple 2D microscope image.

4.4. Vessel Enhancement

The morphology of blood vessels is an important indicator for most of the retinal diseases. The accuracy of blood vessel segmentation in retinal fundus images affects the quality of retinal image analysis and consequently the quality of diagnosis. Contrast enhancement is one of the crucial steps in any of retinal blood vessel segmentation approaches. The reliability of the segmentation depends on the consistency of the contrast over the image. Bandara and Giragama [10] presented an assessment of the suitability of a recently invented spatially adaptive contrast enhancement technique for enhancing retinal fundus images for blood vessel segmentation. The enhancement technique was integrated with a variant of the Tyler Coye algorithm, which has been improved with a Hough line-transformation-based vessel reconstruction method. The proposed approach was evaluated on two public datasets, STARE [1, 85] and DRIVE [181]. The assessment was done by comparing the segmentation performance with five widely used contrast enhancement techniques based on wavelet transforms, contrast limited histogram equalization, local normalization, linear un-sharp masking and contourlet transforms. The results revealed that the assessed enhancement technique is well suited for the application and also outperforms all compared techniques.

In addition to retinal fundus images, OCT and OCTA are other imaging modalities to offer retinal vessel visualization. As discussed before, OCT is a non-invasive,

high-resolution medical imaging modality that can resolve morphological features, including blood vessel structures, in biological tissue as small as individual cells, at imaging depths on the order of 1 mm below the tissue surface. An extension of OCT, named OCT angiography (OCTA), is able to image non-invasively the vasculature of biological tissue by removing the imaging data corresponding to static tissue and emphasizing the regions that exhibit tissue motion. OCTA has demonstrated great potential for characterization of vascular-related ocular diseases such as glaucoma, age-related macular degeneration, and diabetic retinopathy. Quantitative analysis of OCT and OCTA images, such as segmentation and thickness measurement of tissue layers, pattern analysis to identify regions of tissue where the morphology has been affected by a pathology from regions of healthy tissue, segmentation and sizing of blood and lymph vasculature, etc., has a significant clinical value as it can assist physicians with the diagnosis and treatment of various diseases. However, morphological features of interest in OCT and OCTA are masked or compromised by speckle noise, motion artefacts and shadow artefacts generated by superficial blood vessels over deeper tissue layers due to scattering and absorption of light by the blood cells. Tan *et al.* [188] introduced a novel image processing algorithm based on a modified Bayesian residual transform (MBRT). Tan’s algorithm was developed for the enhancement of morphological and vascular features in OCT and OCTA images.

4.5. Tool Tracking

In order to integrate robotic manipulators in retinal microsurgery and to augment the clinicians perceptual ability, a critical component is the capacity to accurately and reliably estimate the location of an instrument when in the field of view. As microscopes have video recording capabilities, a number of methods have thus focused on real-time visual tracking of instruments from image data.

A major challenge to do so from an algorithm point of view is that the instrument appearance is difficult to model over time. Initially, methods relied on knowing the instrument geometry to track the instrument [152, 187]. Alternatively, visual surveying has been the basis of a number of methods in order to overcome the need to know the instrument structure before hand [28, 156]. Unfortunately, such methods have difficulties in dealing with prolonged tracking time and require failure checking systems. More recent methods have leveraged machine learning methods to provide fast and robust solutions to the instrument tracking problem. This has ranged from using Boosting methods [185, 186] to Random Forests [158], as well as a variety of methods that update learned models dynamically to improve robustness [159]. Unsurprisingly however, recent use of deep learning methods have been shown to work extremely well in terms of 2D instrument pose localization, speed and robustness [109, 111].

Perhaps even more promising is the use of high-resolution OCT information at

the 3D location of the instrument tip. Given new integrated iOCT capabilities, some preliminary results for tracking instruments with iOCT image data have been shown possible (see Fig.0.14) and are promising [160, 200]. Such combined multi-modal instrument tracking approaches may be the key for precise intraocular tracking of surgical instruments. Without a doubt, this will have important relevance in robotic assisted retinal micro surgery, as the intraoperative OCT has an axial resolution of $510\mu\text{m}$, which allows for precise depth information to be estimated and appears far better than pure stereo-based estimation [156].

Figure 0.14 *Modality-specific instrument tracking approaches.* These tracking approaches are developed and combined to a robust and real-time multi-modal instrument tracking approach. The tracked instrument is indicated by the yellow arrow.

4.6. Auditory Augmentation

In data augmentation and perception, the visual modality is currently dominant. However, conveying all the available information in an operational environment through the same modality may risk overstimulation and a high cognitive load which could lead to inattentive blindness (see Fig.0.15). In modern surgical rooms there are many visual displays. Sometimes their number is even higher than the number of surgeons and physicians in the room. Following all these monitors during a surgical procedure can be very difficult.

Augmenting the cognitive field with additional perceptual modalities such as audio can sometimes offer a solution to this problem. Audio as a modality plays a substantial role in our perception and provides us with focused or complementary information in an intuitive fashion. Auditory display, and specifically sonification, aims at exploiting the potential of the human auditory system to expand and improve perception. This modality has been less exploited in augmented reality and robotic applications so far. Sonification is the transformation of perceptualized data into non-speech audio. The

Figure 0.15 *One frame taken from Lumera 700 with integrated iOCT-Rescan 700 (Carl Zeiss AG).* - Ophthalmic operation in posterior segment on a patient with sub-retinal haemorrhage; surgeon is precisely injecting rtPA in sub-retinal domain. Surgeons see this side-by-side view of the optical image and OCT image intraoperatively. Yellow circles show the areas that need surgeons attention. (Left) One frame taken from Lumera 700 with integrated iOCT-Rescan 700 (Carl Zeiss AG) - Ophthalmic operation in anterior segment; surgeon is performing Descemet Membrane Endothelial Keratoplasty (DMEK) operation. Surgeons see this side-by-side view of the optical image and OCT image intra-operatively. Yellow circles show areas that need surgeons attention. (Right).

temporal and spatial advantages of the audio modality suggest sonification as an alternative or complement to visualization systems. The siren alarm of ambulances and parking guidance systems are the most known examples of sonification, which provide us respectively with the urgency and distance function in an intuitive way. The omnidirectional nature of sound relieves the surgeon from steadily looking at the monitors and switching between monitors and patient.

The works in [129, 165] suggest solutions for using sonification for surgical data augmentation and precise navigation. Sonification methods proposed for robotic vitreo-retinal surgery give the surgeon aural feedback about the status of the operation. These studies investigate different digital audio effects on a music track to indicate the current anatomical area where the moving surgical tool is. Data regarding the corresponding area can be acquired from several sources, including optical coherence tomography.

5. State-of-the-art Robotic Systems

As described in §1, in retinal microsurgery the requirements for fine motor control are high, exceeding the fundamental physiological capability of many individuals. In order to enhance the capabilities of surgeons, a variety of robotic concepts have been explored that span the spectrum between conventional surgery and full robotic autonomy. As we move along that spectrum, major approaches include completely handheld systems that are completely ungrounded and maintain much of the essence of conventional surgery, cooperative-control systems in which the surgeon’s hand and the robot are both in direct contact with the surgical instrument, and master-slave teleoperation systems in which the “master” human-input device is distinct from the “slave” surgical

Table 0.6 Overview of systems for robotic retinal surgery

Group	Config.	Actuation	RCM	Refs.	Comments
Automatical Center of Lille	TO	AT	CT	[15, 71]	RCM w/ distal insertion
Beihang Univ.	TO	AT	LI	[78]	RCM w/ proximal insertion
Carnegie Mellon Univ.	HH	PZ	—	[123, 212]	handheld 6-DoF parallel
Columbia/Vanderbilt	TO	AT	—	[197, 198]	parallel + distal continuum
ETH Zurich	TO	MA	—	[16, 108]	untethered microrobot
Imperial College London	TO	AT	—	[116]	continuum
Johns Hopkins Univ.	CC	AT	LI	[83, 194]	RCM w/ distal insertion
Johns Hopkins Univ.	TO	AT	LI	[84, 179]	RCM + distal continuum
Johns Hopkins Univ.	HH	PZ	—	[99, 178]	handheld 1-DoF prismatic
King’s/Moorfields Robot	CC	AT	LI	[121]	RCM w/ distal insertion
KU Leuven	CC	IT	LI	[29, 65]	RCM w/ proximal insertion
McGill Univ.	TO	IT	—	[90, 91]	parallel macro-micro
NASA-JPL/MicroDexterity	TO	IT	—	[39, 169]	serial, cable-driven
Northwestern Univ.	TO	AT	—	[69, 97]	parallel
TU Eindhoven/Preceyes	TO	AT	LI	[132, 133]	RCM w/ distal insertion
TU Munich	TO	PZ	—	[142, 143]	hybrid parallel-serial
UCLA	TO	AT	CT	[155, 205]	RCM w/ distal insertion
Univ. of Tokyo	TO	AT	CT	[92, 193]	RCM w/ distal insertion
Univ. of Tokyo	TO	AT	—	[139, 192]	parallel + distal rotation
Univ. of Utah	TO	PZ	—	[140]	serial
Univ. of Western Australia	TO	AT	CT	[215]	RCM w/ distal PZ insertion

List of abbreviations used: AT = admittance-type electric motor, CC = cooperatively controlled, CT = circular track, HH = handheld, IT = impedance-type electric motor, LI = linkage-based, MA = magnetic, PZ = piezoelectric actuators, TO = teleoperation.

robotic manipulator. As technology continues to advance, all robotic systems have the potential to be partially automated, and all but handheld devices have the potential to be fully automated, although automation has not typically been the primary motivation for the development of robotic retinal-surgery platforms. Table 0.6 provides an overview of 21 distinct robotic retinal-surgery platforms that have been described to date. An overview of the majority of these systems is assembled in Fig. 0.16. The fundamental working principles behind these systems, and associated parameters, are explained below.

5.1. Common Mechatronic Concepts

In this section, we begin with an introduction to some of the actuation and mechanism design concepts that are common across multiple robotic retinal-surgery platforms.

Figure 0.16 Representative examples of robotic retinal-surgery platforms from the categories of (a) Cooperatively controlled robots, (b) Teleoperated systems, (c) Handheld systems and (d) Microrobotic systems.

5.1.1. Electric-motor Actuation: Impedance-type vs. Admittance-type

The electric motor is by far the most commonly used actuator in the design of surgical robots. However, even within electric-motor-based systems, varying the inertial and frictional properties—typically through the use of transmissions, such as gearing—can lead to drastically different robot dynamics, to the point of completely changing the input-output causality of the system.

At one extreme, *impedance-type robots* use direct-drive motors with no (or little) gearing or capstan-like cable transmissions, which results in joints with low friction and low inertia that are easily backdrivable when powered off. The pose of impedance-type robots must be actively controlled, and gravity compensation is typically employed to prevent the robot from sagging under its own weight. It can be assumed that all external loads will disturb the robot from its desired pose to some degree, although the feedback control mitigates these disturbances. From one perspective, impedance-type robots have inherent safety characteristics, in that their maximum force capability is fairly low, which limits the risk of harm to humans in direct contact. In the event of a power-loss system failure, these systems can be removed rapidly from the scene because they can be simply backdriven by hand, however, they may not “fail safe” in the sense that they may not remain in their last commanded position in the event of a power-loss system failure.

At the other extreme, *admittance-type robots* have a significant amount of gearing, reflected inertia, and nonlinear friction, making the joints nonbackdrivable to a substantial degree, even when powered off. An admittance-type robot will hold its pose whenever it is not actively commanded to move, and requires a control system to move (whereas an impedance-type robot requires a control system to hold its pose). An admittance-type robot exhibits high precision, and it can be assumed that, when interacting with soft tissue, environmental disturbances have a negligible effect on the pose of the robot. From one perspective, admittance-type robots are “fail safe” in the sense that, in the event of a power-loss system failure, they remain in their last commanded position. A quick-release mechanism may need to be added if one wants to remove the instrument relatively quickly from the patient’s eye in an emergency situation. Admittance-type robots may have a very high maximum force capability, which represents an inherent safety risk when in direct contact with humans.

Impedance-type and admittance-type robots can be viewed as two ends of a continuous spectrum, without definitive boundaries. For the purposes of this chapter, if a robot is difficult or impossible to move by a human when it is powered off then it will be considered an admittance-type robot; otherwise, it will be considered an impedance-type robot.

5.1.2. Piezoelectric Actuation

Piezoelectric actuators exhibit a strain (i.e. they stretch) when a voltage is applied. These actuators are capable of extremely precise motions, typically measured in nanometers. In addition, motions can be commanded at high bandwidth. However, standard piezoelectric actuators are typically not capable of large motions.

Piezoelectric stick-slip actuators utilize a piezoelectric element that stretches when a voltage is applied (for example, by $1\ \mu\text{m}$), with a distal element that is moved by the piezoelectric element through static friction. When the piezoelectric element is rapidly retracted, the inertia of the distal element causes slipping relative to the piezoelectric element, resulting in a net displacement of the distal element. The result is an actuator that is similar in behavior to a stepper motor with extremely small steps, but with a stochastic step size. By taking multiple successive steps, large net motions are possible. Piezoelectric stick-slip actuators behave much like admittance-type actuators during normal operation, in that they are very precise and they maintain their position when not commanded to move. However, they can be easily backdriven when powered off by overcoming the static friction.

Other *piezoelectric motors* use various forms of “inchworm” strategies in which multiple piezoelectric actuators are successively stretched and relaxed in a sequence that results in a net stepping behavior. From a high-level control perspective, these piezoelectric motors behave much like piezoelectric stick-slip actuators. However, piezoelectric motors are able to generate larger forces and resist higher loads before being backdriven. Another example of piezoelectric motor is the ultrasonic motor (SQL-RV-1.8 SQUIGGLE motor, New Scale Technologies, N.Y., USA) used in the Micron handheld robotic instrument [210], which uses a ring of piezoelectric elements to rotate a threaded rod, thus producing linear actuation with a range of motion that is limited only by the length of the threaded rod.

5.1.3. Remote-center-of-motion (RCM) Mechanisms

When the surgical instrument passes through a scleral trocar, it must be constrained to move with four DoFs in order to respect the constraint of the trocar; this includes 3-DoF rotation about the center of the trocar, and 1-DoF translation parallel to the shaft of the instrument (Fig. 0.17). Some retinal robots implement this kinematic constraint in software. Other robots use a dedicated *remote-center-of motion (RCM) mechanism* to mechanically implement the kinematic constraint. RCM mechanisms provide an additional layer of safety, in that no system failure could cause the robot to violate the kinematic constraint of the trocar and potentially harm the sclera.

There are two basic RCM designs that have dominated the design of retinal robots. The most common is a linkage mechanism (e.g., double parallelogram) preceded proximally by a rotary joint, with axes that intersect at a point (which is the location of the RCM). The second most common is a circular track preceded proximally by a rotary

Figure 0.17 Instrument motion DoF divided into [Left] 4 DoF that do not alter the location of an optimal pivot point located central in the pars plana; [Right] 2 DoF that alter the orientation of the eye in its orbit by displacing the pars plana.

joint, with axes that intersect at a point. Both of these base mechanisms are typically succeeded distally by a rotary joint to rotate the instrument about its shaft and a prismatic joint to translate the instrument along its shaft (i.e., to insert/withdraw the instrument), which completes the 4-DoF mechanism. However, recent innovations in linkage-based RCM mechanisms have eliminated the distal prismatic joint, simplifying the portion of the robot that is closest to the eye and microscope. The instrument-translation DoF is in this case enabled by a more complex proximal mechanism [65, 78].

In order to rotate the eye in its orbit to image the complete retina (Fig. 0.17), RCM mechanisms must be preceded proximally by additional DoF to move the location of the RCM point. This is typically accomplished by a simple 3-DoF Cartesian stage, which need not have the precision of the RCM, since its only function is eye rotation, and it is not directly involved in the control of the instrument with respect to the retina. It must be noted that the inherent safety motivating the use of an RCM mechanism is somewhat reduced by the addition of this proximal positioning system, as its motion can easily violate the trocar constraint and should be conducted with sufficient care.

5.2. Handheld Systems

The first class of retinal robotic systems that we consider are handheld devices in which mechatronic components intervene between the surgeon’s hand and the tip of

the instrument. Of all the systems that we consider, handheld systems are the closest to existing clinical practice and workflow, with the surgeon retaining a great deal of direct control over the instrument, including the ability to rapidly remove the instrument from the eye. Because handheld devices are mechanically ungrounded, they are able to affect, but not fully control, the steady-state pose of the instrument. In this regard, handheld systems are only robotic systems in the broadest sense, and might be better described as “mechatronic” systems. They are best suited to compensating for small motions, particularly over time scales that are too fast for the surgeon to react. Unlike teleoperated systems, which can enhance accuracy by filtering error out of commands sent to the manipulator, handheld systems can reduce error only by means of active compensation.

The handheld-system concept that has received the most attention is the Micron system from Carnegie Mellon University [11, 123, 161, 212, 213]. The Micron uses a 6-DoF Stewart-platform (a.k.a., hexapod) parallel mechanism driven by piezoelectric motors. The motion of the handle is tracked (e.g., optically [126] or electromagnetically [124, 125]) using an external system. The control system tries to extract the intentional motion of the operator and to cancel out all unintentional motion, including the tremor of the operator’s hand.

Researchers at Johns Hopkins University have created a class of “SMART” (sensorized micromanipulation aided robotic-surgery tools) instruments that incorporate 1-DoF motion control into a handheld instrument. SMART instruments incorporate a common-path OCT fiber into the instrument to measure the distance between the instrument and the retina, and use a piezoelectric motor to move the instrument’s end-effector prismatically. This active DoF, which is directed normal to the retina during operation, is automatically controlled using real-time feedback from the OCT in an attempt to maintain a fixed distance between the instrument’s end-effector and the retina, in spite of surgeon tremor. To date, the group has developed a microforceps [178] and a microinjector [99] based on this concept.

5.3. Cooperative-control Systems

The defining feature of cooperative control, sometimes referred to as “hands-on” cooperative control or “co-manipulation”, is that both the surgeon’s hand and the robot cooperatively hold the surgical instrument. The resulting instrument motion is defined by input commands from both the surgeon and the robot. To a lesser extent the external environment will also affect instrument motion (mainly through bending of the thin instrument shaft). Cooperative-control systems retain much of the manipulation experience of a traditional surgery for the surgeon. Cooperative-control systems can also be used in a teleoperation configuration with only minor modifications, but the reverse is not true in general.

The earliest example of a cooperative-control system for robotic retinal surgery is the Steady-Hand Eye Robot (SHER) [81, 83, 194] developed at the Johns Hopkins University. The SHER comprises a 3-DoF Cartesian robot, followed distally by a linkage-based RCM mechanism, followed distally by a passive rotation joint for rotation about the instrument shaft. Because the robot does not include a dedicated prismatic actuator for instrument insertion/withdrawal, in general the RCM point at the trocar is implemented virtually, involving all DoF of the robot. However, the RCM mechanism was designed so that the mechanical RCM will correspond to the trocar (and virtual RCM) when the instrument’s end-effector is interacting with the macula; in that location, very little movement of the Cartesian robot is required. The SHER is an admittance-type robot. A force sensor integrated into the instrument handle measures the force applied by the user. This force is used as an input to control the velocity of the robot (i.e., admittance control), which in the simplest case is a linear relationship that creates the effect of virtual damping. The small forces conveyed by human hand tremor can be attenuated through filtering, leading to the “steady hand” designation.

A similar admittance-type paradigm is currently being pursued at King’s College London and Moorfields Eye Hospital [121]. The system comprises a 7-DoF positioning robot (a 6-DoF Stewart platform followed distally by a rotational actuator), followed distally by a linkage-based RCM mechanism, followed distally by a prismatic actuator for instrument insertion/withdrawal.

The system from KU Leuven [29, 65, 64] is the only cooperative-control platform that follows the impedance paradigm: the robot is impedance-type, and the controller generates a retarding force that is proportional to the velocity of the device (i.e., impedance control), creating the effect of virtual damping. This system does not require embedding a force sensor in the operator’s handle. Using the impedance controller, it is also possible to mitigate unintentional and risky motions through the application of force to the surgeon-manipulated instrument. The KU Leuven system comprises a 3-DoF Cartesian robot, followed distally by a linkage-based RCM mechanism, followed distally by a passive joint for rotation about the instrument shaft. The RCM mechanism is also responsible for instrument insertion/withdrawal.

5.4. Teleoperated Systems

Teleoperated systems comprise two distinct robotic subsystems connected via a communication channel: a “slave” manipulator that mechanically manipulates the surgical instrument, and a “master” human-input device that is directly manipulated by the surgeon. The master device typically takes the form of a haptic interface, but other types of input devices (e.g., joysticks, 3D mice) have been used as well. Because there is not a direct physical connection between the master and slave, teleoperation systems provide more opportunities to substantially change the surgical experience of the sur-

geon, including position and force scaling, as well as other ergonomic improvements such as moving the surgeon away from the operating microscope to a surgeon console. In the following, we focus on the design of the slave manipulator, since master input devices are easily exchanged.

The first teleoperated retinal-surgery robot was the Stereotaxical Microtelemanipulator for Ocular Surgery (SMOS), created by researchers at the Automatical Center of Lille [15, 71]. The SMOS slave comprised a 3-DoF Cartesian robot, followed distally by a circular-track RCM mechanism, followed distally by a prismatic actuator for instrument insertion/withdrawal, followed distally by a joint for rotation about the instrument shaft. In the years that followed, three different groups developed systems with very similar designs. The first was a group at the University of Western Australia [215]. Although the majority of their robot is driven by electric motors, the distal insertion/withdrawal stage uses a piezoelectric motor. The second was a group at the University of Tokyo [92]; it should be noted that this system was an evolution from an earlier prototype with a different circular-track-RCM design [193]. The third was a group at UCLA, with the Intraocular Robotic Interventional and Surgical System (IRISS) [155, 205]. IRISS had two key innovations over SMOS. The first was a tool-changer design that enabled two different instruments to be changed (automatically) in a given hand. The second innovation was the use of two circular-track RCM mechanisms, separated by the same distance as the left- and right-hand scleral trocars, mounted to a single base positioning unit, which enables a single slave robot to be used for bimanual manipulation (i.e., one slave ‘body’ with two ‘arms’). All of the systems described above are admittance-type robots.

Although linkage-based RCM mechanisms have dominated the designs of platforms based on cooperative control, they have received relatively little attention in the context of teleoperation. The system that is the most mature, and has received the most attention, is the PRECEYES Surgical System developed by a collaboration between TU Eindhoven and AMC Amsterdam and commercialized by Preceyes [132, 133]. The slave is an admittance-type robot, comprising a 3-DoF Cartesian robot, followed distally by the RCM mechanism, followed distally by a prismatic actuator for instrument insertion/withdrawal, followed distally by a joint for rotation about the instrument shaft. The slave is equipped with a quick-release instrument holder such that the instrument can be swiftly removed in case of an emergency. A further noteworthy feature is that the 3-DoF proximal positioning stage is integrated in the patients head-rest such that there is more knee space for the operator who sits closely to the patient and manipulates a bed-mounted master manipulator. Recently a group from Beihang University [78] developed a system that is similar to the PRECEYES system, but they removed the most distal translation stage used for instrument insertion/withdrawal and modified the more-proximal RCM mechanism to provide that DoF, similar to the system from KU Leuven (see §5.3).

Four groups have developed solutions based on parallel robots, all of which implement the RCM in software. The first such system was developed at McGill University [90, 91]. It was based upon two 3-DoF Cartesian robots that contacted a flexible element at two distinct locations; controlling the positions of the two 3-DoF robots enabled control of the tip of the flexible element through elastic beam bending. Each of the six actuators was designed as a two-stage, macro-micro actuator, with a high-precision piezoelectric element mounted on an impedance-type linear electric motor. The three ‘parallel’ systems that followed were all based upon a 6-DoF Stewart platform, and all were of the admittance type, including systems from Northwestern University [69, 97], the University of Tokyo [139, 192], and Columbia University [197]. The system from the University of Tokyo is similar to the Northwestern system, but also included an additional distal rotation DoF for rotation of the instrument about its shaft axis. The system from Columbia, the Intro-Ocular Dexterity Robot (IODR), is similar to the Northwestern system, but with a major innovation: it includes a distal 2-DoF continuum device to add dexterity inside of the eye (i.e., distal to the RCM implemented at the scleral trocar).

In the years that followed, other snake-like continuum devices have been developed that enable the instrument’s end-effector to approach and manipulate the retina from arbitrary orientation. The system from Johns Hopkins University [84, 179] is quite similar to the continuum device of the IODR, but is deployed from the SHER platform (see §5.3). The system from Imperial College London uses nested superelastic tubes, to be deployed from a unit located on the microscope. [116].

Two systems have been developed based on piezoelectric stick-slip actuators. Both systems implement the RCM in software, and both systems exhibit compact designs motivated by the goal of mounting the slave manipulator on the patient’s head. The first was the system from TU Munich [142, 143], called iRAM!S (Robot-assisted Microscopic Manipulation for Vitreo-retinal Ophthalmologic Surgery). iRAM!S uses ‘hybrid parallel-serial’ kinematics comprising a serial chain of simple parallel mechanisms, leading to a compact design reminiscent of RCM mechanisms. The second is the system from the University of Utah [140], which uses a conventional 6-DoF serial-chain kinematic structure (3-DoF Cartesian robot followed distally by a 3-DoF spherical wrist) with the goal of eliminating uncontrolled and unsensed DoF in the kinematic chain.

Finally, one system that stands out as being quite distinct from any other concept discussed above is the system developed as a collaboration between NASA-JPL and MicroDexterity [39, 169]. In that system, the slave manipulator is a cable-driven impedance-type robot with serial-chain kinematics.

5.5. Untethered “Microrobots”

A more exotic robotic approach has been pursued at ETH Zurich, where researchers have been developing untethered magnetic devices that can navigate from the sclera to the retina, driven wirelessly by applied magnetic fields, to deliver potent drugs [16, 17, 18, 53]. Although the tiny untethered devices are referred to as “microrobots” for lack of a better term, the robotic intelligence in the system lies entirely in the external magnetic control system. The applied magnetic fields are generated by the OctoMag system, which comprises eight electromagnets designed to surround the patient’s head without interfering with the operating microscope [108]. From the perspective of control, magnetic actuation shares many properties of other impedance-type robots, assuming the vitreous has been removed and replaced with a liquid solution. With magnetic microrobots, force control at the retina can be accomplished in an open-loop fashion, giving magnetic microrobots an inherent safety when compared to traditional robots. However, the forces that can be generated are also quite small, which complicates or even prohibits performing certain surgical procedures.

An alternate concept has been explored for use with an intact vitreous in which the microrobot takes the form of a miniature screw driven by magnetic torque [127, 145]. The same field-generation and localization systems can be applied with this concept as well, but the nonholonomic motion of screws through soft tissue requires more sophisticated motion planning and control. Recently, researchers pushed the miniaturization envelope even further, and presented the navigation of micropropelling swimmers inside an intact porcine vitreous humour, with their results evaluated with OCT measurements [207]. The fact that the vitreous humour would not need to be removed is an appealing property warranting further investigation.

5.6. Clinical Use Cases

Recently, a number of first-in-human experiments have been reported. The teleoperated PRECEYES system (§5.4) has been used for initiating a membrane-peeling procedure on six patients during a human trial at Oxford university [50]. The robot was used successfully to lift up a flap of the ERM or the ILM away from the macula surface using a bevelled needle or pick. Subsequently, a subretinal injection was conducted successfully in three patients [51]. In the framework of EurEyeCase, PRECEYES has been used in another human trial at the Rotterdam Eye hospital. Here, for the first time a virtual fixture was implemented based on real-time acquired distance measurements from an OCT fiber [35], demonstrating the feasibility of in-vivo use of OCT-integrated instruments.

Contemporaneously, the cooperative-control system from KU Leuven (§5.3) was been used for the first-in-human demonstration of robot-assisted vein cannulation [61]. Four patients with RVO were treated, demonstrating the feasibility of robot-assisted

cannulation and injection of the thrombolytic agent ocriplasmin (Thrombogenics NV) to dissolve clots obstructing retinal veins.

5.7. General Considerations w.r.t. Safety and Usability

Regardless of the surgical robot used, there is still a risk of human error, which may lead to iatrogenic trauma and blindness. For example, excessive tool pivoting around the entry incision may lead to astigmatism, wound leak, or hypotony. Accidental motions of the tool may still puncture the retina or cause bleeding, or even touch the intraocular lens and cause a cataract [52]. All of these risks are indeed present, since the previously described robotic systems do not demonstrate “intelligence”, they merely replicate or scale-down the motions of the commanding surgeon. Thus, robots can improve surgical dexterity but not necessarily surgical performance. Functionalization of the tools with force or pressure sensors, as well as ophthalmic image processing, can improve the perception of the surgeon and enable him/her to link with artificial intelligence algorithms towards further improving the success rate of interventions.

A typical step in retinal surgery is a rotation of the eye in its orbit to visualize different regions of the retina. This is accomplished by applying forces at the scleral trocars with the instrument shafts. When done bimanually, surgeons have force feedback to ensure that their hands are working together to accomplish the rotation, without putting undue stress on the sclera. When using more than one robotic manipulator in retinal surgery, whether in a cooperative or teleoperated paradigm, the control system must ensure that the robots work in a coordinated fashion. This kinematically constrained problem is solved in [198].

Further, all teleoperation systems and especially systems using curved and shape-changing instruments or untethered agents require re-training of the surgical personnel to get accustomed to this remote-manipulation paradigm, which may disrupt surgical workflow. Many of the master interfaces have been designed to make this transition as intuitive as possible, and are based on either recreating the kinematic constraints of handheld and cooperative-control systems (i.e., with the surgeon’s hand on the instrument handle outside of the eye) or on creating kinematics that effectively place the surgeon’s hand at the end-effector of the instrument inside the eye (with the kinematic constraint of the trocar explicitly implemented in the interface). However, recent work suggests that placing the surgeon’s hand at the end-effector of the instrument, but not explicitly presenting the kinematic constraints of the trocar to the user, may lead to improved performance, likely due to the improved ergonomics that it affords [141].

6. Closed-loop Feedback and Guidance

Benefiting from feedback from sensorized instruments (§3), it becomes possible to establish high-bandwidth feedback schemes that update in real time with the changing

anatomy. This section describes different feedback and guidance schemes, including haptic feedback and other force-servoing schemes. Through sensor fusion, it becomes possible to implement multi-rate estimation schemes that mix information and measurements, derived from pre-operative or intra-operative imaging, with local sensor measurements. Feedback and guidance schemes share commonalities across hardware configurations, but in the following the discussion is organized by category, describing feedback schemes tailored for hand-held systems (§6.1), cooperative-control systems, (§6.2) and finally teleoperation systems (§6.3).

6.1. Closed-loop Control for Handheld Systems

The hand-held system Micron from Carnegie Mellon University (§5.2) tracks its own motion using a custom optical tracking system [126], performs filtering to determine the undesired component of motion [123], and deflects its own tip using high-bandwidth actuators in order to counteract the undesired movement of the tip [212]. Control of Micron is based on internal-model control, which provides a frequency-domain design technique that can handle underdamped dynamics, is robust under conditions of model error, and addresses time delay [27]. Due to the active nature of error compensation, performance is limited by time delay [123]. As a result, control implementation with Micron has frequently incorporated feedforward tremor suppression based on Kalman state estimation [13]. Besides the image-guided applications described in (§7), in order to provide tremor compensation when image guidance is not used, the system incorporates a high-shelf filter with negative gain as a tremor-canceling filter, providing what may be thought of as relative motion scaling below 2 Hz, with full suppression above 2 Hz [123]. Previously, notch filtering of the neurogenic component of physiological tremor was implemented [161, 163], but over time experimentation made clear that achievement of significant accuracy enhancement for surgical tasks requires error suppression at frequencies considerably lower than had been foreseen, even to frequencies that overlap with voluntary movement. The controller can also be programmed to limit velocity, which may help to avoid tissue damage [137].

Micron has been used also in combination with force-sensing tools to enhance safety in tissue manipulation. Gonenc *et al.* [66] integrated a 2-DoF force-sensing hook tool with an earlier-generation 3-DoF Micron prototype for superior performance in membrane peeling operations. By mapping the force information into auditory signals in real time, the forces could be kept below a safety threshold throughout the operation. Furthermore, Gonenc *et al.* mounted a force-sensing micro-needle tool on Micron, enabling an assistive feedback mechanism for cannulating retinal veins more easily [68]. The implemented feedback mechanism informs the operator upon vessel puncture and prevents overshoot based on the time derivative of sensed tool

tip forces. In [67], a compact, lightweight, force-sensing micro-forceps module was integrated with Micron and the existing tremor cancellation software was extended to inject micro-vibrations on the tool tip trajectory when necessary to assist membrane delamination. Experiments on bandages and raw chicken eggs have revealed that controlled micro-vibrations provide ease in delaminating membranes. Automatic force-limiting control has also been demonstrated with the 6-DoF Micron system for membrane peeling using a parallel force/position control system [202].

An alternate handheld system developed for retinal surgery and highlighted in (§5.2), is SMART from Johns Hopkins University [178]. SMART is a microsurgical forceps that can actively stabilize tool tip motion along the tool axis by using a fiber-optic OCT to measure the distance between tool tip and target tissue. The OCT signals are sent via feedback control to a piezoelectric motor that provides active tremor compensation during grasping and peeling functions. This closed-loop positioning function can be particularly useful for 1-DOF motion stabilization when a target tissue and an environment is delicate, and undesired collision needs to be avoided.

6.2. Closed-loop Control for Cooperative-control Systems

Cooperative control is a shared control scheme where both the operator and the robot hold the surgical instrument (see §5.3). The force exerted by the operator guides the robot to comply with his/her movements. These robotic systems can be augmented with virtual fixtures [166], and can be fitted with smart instruments possessing various sensing modalities. By way of example, smart instruments with force sensing capability may prove essential for safe interaction between the robot and the patient. The Johns Hopkins University team have developed a family of force-sensing instruments [81, 82, 95, 119] with fiber-optic sensors integrated into the distal portion of the instrument that is typically located inside the eye. Auditory [44] and haptic [7] force-feedback mechanisms have demonstrated the potential value of regulating the tool-to-tissue forces. Initially, the JHU team have employed cooperative-control methods that modulate the robot behavior based on operator input and/or tool tip forces [7, 106]. Later on, they extended these methods to take into consideration the interaction forces between tool shaft and sclera.

6.2.1. Robot Control Algorithms based on Tool-tip Force Information

The earliest application of micro-force sensing in cooperative robot control was proposed by Kumar *et al.* [106]. Balicki *et al.* [7] implemented this control scheme on the SHER as one of the available behaviors for assisting retinal surgery. Force scaling cooperative control maps, or amplifies, the human-imperceptible forces sensed at the tool tip (F_t) to handle interaction forces (F_h) by modulating robot velocity $\dot{x} = \alpha(F_h + \gamma F_t)$. Scaling factors of $\alpha=1$, and $\gamma=500$ were chosen to map the 0–10 mN manipulation forces at the tool tip to input forces of 0–05 N at the handle. Furthermore, a force

limiting behavior was developed to increase maneuverability when low tip forces are present [7]. The method incorporates standard linear cooperative control with an additional velocity constraint that is inversely proportional to the tip force:

$$\dot{x} = \begin{cases} V_{lim}(F_t), & -F_h < V_{lim}(F_t) \text{ and } F_t < 0 \\ V_{lim}(F_t), & -F_h > V_{lim}(F_t) \text{ and } F_t > 0 \\ \propto F_h, & \text{otherwise} \end{cases}$$

where $V_{lim}(F_t)$ is a velocity limiting function described graphically in Fig. 0.18. This force limiting behavior effectively dampens the manipulation velocities.

Figure 0.18 Velocity limiting function. Constraint parameters m and b were chosen empirically. Forces lower than $f_1 = 1$ mN do not limit the velocity. Velocity limit was set at $v_2 = 0.1$ mm/s for forces above $f_2 = 7.5$ mN [7].

6.2.2. Robot Control Algorithms based on Sclera Force Information

An underappreciated limitation of current robotic systems is the suboptimal user perception of the forces present at the point that the tool passes through the sclera. With multi-function force sensing tools (§3.1.2), He *et al.* measures the tool-tissue forces at both the tip and the interface with the sclera. A variable admittance control method was introduced [80] to take advantage of this knowledge. The control law is: $\dot{x}_{ss} = \alpha(A_{sh}F_{sh} + \gamma A_{ss}F_{ss})$ where $\dot{x}_{ss}$ is the desired velocity of where the robot/tool contact the sclerotomy in the sclera, F_{sh} and F_{ss} are the handle input force and sclera contact force resolved in the sclera frame, respectively, γ denotes the constant scalar as the force scaling factor, α denotes the constant scalar as the admittance gain, and A_{sh} and A_{ss} are the diagonal admittance matrices associated with the handle input force and sclera contact force in the sclera frame, respectively. A virtual RCM can be realized by setting $A_{sh} = \text{diag}(0, 0, 1, 1, 1, 1)$ and $A_{ss} = I$. The admittance matrix A_{sh} removes the

Figure 0.19 JHU SHER variable admittance control. Left: robot variable admittance control framework based on sclera force/position input. Right: Admittance variation (linear or non-linear) along the insertion depth

transverse force components that can lead to undesired lateral motion, and preserves the 4-DoF motion that is allowed by the RCM constraints. In addition, the sclera force feedback is to servo the sclera contact force toward zero. This strengthens the virtual RCM with robustness against eye motion attributed to other instrument and/or patient movement. When the surgeon is performing RVC, the tool tip is close (or in contact) to the retina, and an RCM is desired to minimize the eye motion and the target tissue. When the surgeon needs to reposition the eye to adjust view, the tool is kept away from the retina to avoid collision. Therefore, the measured insertion depth of the tool can be used to adjust the robot admittance to provide the appropriate robot behavior. We can define $A_{sh} = \text{diag}(1 - \beta, 1 - \beta, 1, 1, 1, 1)$ and $A_{ss} = \text{diag}(1 + \beta, 1 + \beta, 1, 1, 1, 1)$, where $\beta \in [0, 1]$ could vary linearly along with the tool insertion depth as shown in Fig. 0.19 or non-linearly [49, 77]. When the insertion depth is smaller than the given lower bound l_{lb} , $\beta = 0$ and $A_{sh} = A_{ss} = I$, we have the force-scaling control mode that provides the freedom to reposition the eye with scaled sclera force feedback. When the insertion depth is larger than the given upper bound l_{ub} , $\beta = 1$ and it switches to virtual RCM with doubled gain for minimizing the transverse forces at the sclerotomy.

6.3. Closed-loop Control for Teleoperated Systems

Most of the results in closed-loop control of cooperative-control systems can be applied to teleoperated systems with only minor modifications. However, in contrast to handheld systems or cooperative-control systems, teleoperation systems additionally offer the possibility to completely decouple the operator at the master side from the surgical robotic slave. Thanks to this decoupling it becomes possible to tackle physiological tremor in a number of different manners. First, it is possible to inject physical damping in the master robot's controller, effectively removing the high-frequency mo-

tion of the operator’s hand. Second, it is possible to filter the signal that is sent, e.g., as a reference trajectory for the slave robot to follow, in such a manner that all high-frequency components are filtered out. In a scaled teleoperation scenario, a constant scale factor is used to scale the master command to a scaled reference signal for the slave robot. In this third scenario, the amplitude of the physiological tremor would simply be transmitted in a down-scaled fashion to the slave robot. It goes without saying that a combination of above three methods may be implemented as well.

Teleoperation schemes also offer a number of options to implement virtual walls. One may choose to install a virtual wall at the master side that renders resistive forces upon entry into the forbidden zone. The operator will be slowed down. The slave robot that follows the master’s motion will equally slow down as soon as the master enters the wall. Alternative one may choose to decouple the slave’s motion from the master’s motion with an intermediate ‘proxy’, and effectively halt slave motion upon entry in the forbidden zone. For example, in [98] Jingjing *et al.* propose to compute a homo-centric sphere with radius below that of a spherical-shaped eye as boundary between a safe and a dangerous zone. In a scaled teleoperation scenario this decoupling could correspond to ‘zeroing’ the scale factor between master and slave. In principle, decoupling allows installing stiffer virtual walls at the slave side. In such case penetration can be kept minimal and potentially lower than in the case of a cooperatively controlled systems where the penetration will be lower-bounded by the stiffness of the robot and its controller. In practice, the difference in stiffness may not always be significant [62], especially given the fact that operators are trained individuals that naturally operate in a responsible fashion and typically work at low speeds.

Whereas most practical teleoperation schemes are ‘uni-lateral’, which means that all control signals travel down from master to slave with only visual feedback traveling back upwards to the operator, one may equally consider ‘bi-lateral’ control [76, 113]. By reflecting back position errors or forces measured at the slave to the master, the operator could be in principle be made aware more rapidly of the interactions that are taking place at the slave side. Balicki *et al.* have implemented both uni- and bi-lateral controllers [8]. Bi-lateral controllers can be made responsive to any kind of position or force tracking error [76, 113]. For the former it suffices to compute e.g. from the robot encoders of master and slave, the tracking error. For the latter one needs to measure the interaction forces of the eye that one wants to feedback. While quite some force-sensing instruments have been developed in the past (as depicted in Fig.0.9), most of the governing forces stay well below human thresholds [73]. ‘Force scaling’ would thus need to be applied if one wants to render the forces to a perceivable level. While bi-lateral controllers tend to enhance the operator’s awareness offering a more transparent way of operation in reality this may lead to stability issues [76, 113]. Balicki further proposes *cooperative teleoperation* behaviour. In this hybrid control scheme a robot designed for cooperative control can be jointly controlled by mixing

inputs from an operator handling the robot and from a second operator who provides inputs at a master console [9]. While this approach may combine the benefits from both worlds it does require attendance of two experts who would need training to accustom to this new way of operation.

Note that while ample works in the literature describe contributions to set up visual, auditory or haptic feedback, so far hardly any work analysed the usability and the benefit of one feedback type versus another. This was also a finding of Griffin *et al.* who conducted a systematic review of the role of haptic feedback in robotic retinal surgery to conclude that even in a broader sense proper studies on human factors and ergonomics in robotic retinal surgery are missing[70].

7. Image-guided Robotic Surgery

7.1. Image-guidance Based on Video

The earliest work in this area was that of Dewan *et al.* [47], who described active constraints based on stereo vision to limit the motion of the JHU Steady-Hand Robot to follow a surface or a contour using admittance control. This work was done without a microscope, performing stereo disparity-based 3D reconstruction to constrain the robot for open-sky manipulation. Disparity-based techniques were used by Richa *et al.* [156] to warn the surgeon of proximity to the retina; this was tested quantitatively in a water-filled eye phantom and qualitatively in rabbit eyes *in vivo*, although the proximity threshold was set to 2 mm, which is very large for retinal procedures.

Becker *et al.* used a similar disparity-based stereo technique to develop active constraints for Micron, demonstrating accuracy enhancement in station-keeping, contour-following, a repeated move-and-hold task, and membrane peeling [11]. Implementation of active constraints with handheld systems such as Micron is fundamentally different from setting up admittance-based virtual fixtures with a grounded robot arm, however. Because Micron is not mechanically grounded, it cannot apply force to resist the motion of the human operator. Therefore, active constraints must be implemented as position-based virtual fixtures [11], in which a corrective displacement of the instrument tip is automatically applied in order to constrain the tip motion to the fixture. In such an approach, the null position of the tip manipulator of the handheld system is taken as the user input to the system, and the reference position is adjusted in order to implement the fixture. Just as an admittance-type robot enables implementation of “hard” (unyielding) fixtures by setting the admittance to zero in a given direction, or “soft” (yielding) fixtures by setting the admittance to a reduced but nonzero value, likewise with position-based virtual fixtures, a hard fixture can be implemented by prohibiting all motion in a given direction, whereas a soft fixture can be implemented by providing (down)scaled motion in a given direction within the vicinity of a given location, subject to the range of motion of the manipulator.

Becker *et al.* [14] also used this approach to develop a virtual fixture for vessel cannulation, scaling motion by a factor of 0.5 perpendicular to the target vessel while allowing unscaled motion parallel to the target vessel. This work was demonstrated *ex vivo* in an open-sky porcine retina.

In a similar porcine retina model *ex vivo*, Becker *et al.* [12] implemented a hard fixture for semiautomated scanning for patterned laser retinal photocoagulation. This work performed visual servoing using the aiming beam of the treatment laser. To accommodate the limited range of motion of the 3-DoF Micron prototype at that time [123], the operator provided the gross motion from point to point. Whenever a yet-untreated target was detected within reach of the manipulator, the control system servoed the tip to the target, fired the laser, and returned the tip to its null position. Yang *et al.* [210] subsequently updated this work to demonstrate fully automated scanning, using a newer Micron prototype with much greater range of motion [212]. This updated work featured a hybrid visual servoing scheme, in which motion in the retinal plane was controlled via visual servoing using the microscope cameras, while motion perpendicular to the retina was handheld by closed-loop control using the optical tracker that accompanies Micron [126]—essentially, the visual compliance approach of Castaño and Hutchinson [34]. Direct comparison with the semiautomated approach showed that accuracy was similar at rates below 1 target per second, but that at higher rates the performance of the semiautomated approach dropped off due to the difficulty of the human operator in moving accurately between targets [211]. Also following a hybrid servoing approach similar to [210, 211], Yu *et al.* [216] presented a technique for hybrid visual servoing using microscope cameras for guidance in the plane of the retina and a separate miniature B-mode OCT probe for guidance along the axis of the instrument.

Open-sky implementation allows good performance with stereo disparity-based reconstruction of the retinal surface. However, when it comes to operating in the intact eyeball, this approach is highly problematic due to the complex and nonlinear optics of the eye [213]. Recently, Probst *et al.* [154] presented a semi-dense deep matching approach that involves convolutional neural networks for tool landmark detection and 3D anatomical reconstruction; however, to date the work has been demonstrated only open-sky.

Methods that specifically model the ocular optics have been developed [18, 16], but these have still yielded error on the order of hundreds of microns. To address this problem for Micron, Yang *et al.* [213] exploited the manipulation capability of the instrument in order to implement a structured-light approach. Before starting an intervention, this approach involves generating one or two circular scans with the laser aiming beam that are detected by the microscope cameras as ellipses. The size, aspect ratio, and orientation of the ellipses allow the retinal surface to be reconstructed. This approach to reconstruction was used by Yang *et al.* [213] to demonstrate automated

patterned laser photocoagulation with a handheld instrument in intact porcine eyes *ex vivo*. This approach can be generalized by building aiming beams into other instruments besides those for laser treatment; Mukherjee *et al.* [138] took such an approach in preliminary experiments toward vessel cannulation. Such aiming beams also have potential to be used for proximity sensing to the surface, and intraoperative updates of retinal surface reconstruction [167].

The eyeball moves during surgery, caused sometimes by the patient, and sometimes caused by the surgeon either intentionally in order to change the view of the retina or unintentionally as a result of intraocular manipulation. In order to keep anatomical active constraints registered to the patient, it is important to accurately track the motion of the retina. There are many algorithms for segmentation of fundus images, but such algorithms are generally designed for offline use, and do not provide robust tracking in the presence of illumination changes or avoidance of mistaking intraocular instruments for vessels. To address this need, Braun *et al.* [25] developed a retinal tracking algorithm for intraoperative use with active constraints that uses an exploratory algorithm for rapid vessel tracing [30], with an occupancy grid for mapping and iterative closest point (ICP) for localization in the presence of instrument occlusions and varying illumination. More recently, this work was augmented by incorporating loop-closure detection [136].

Figure 0.20 Examples of research in image-guided robotic retinal surgery. Systems are shown during preclinical testing. Left: intraoperative tracking of needles for iOCT-based servoing during retinal and subretinal injections [221]. Right: hybrid visual servoing for patterned laser photocoagulation using the Micron hand-held robotic system, performed during vitrectomy surgery in a porcine eye *ex vivo* [213].

7.2. Image-guidance Based on OCT

OCT represents an alternative imaging means that provides far higher resolution than microscope cameras, albeit at a much higher cost. Systems for iOCT are available commercially from numerous microscope manufacturers [33]. Efforts have begun to exploit this new means of imaging. Zhou *et al.* [222] described two methods to segment an intraocular surgical needle: one using morphological features of the needle as detected in the OCT images, and the other using a fully convolutional network. The methods were demonstrated in porcine eyes *ex vivo*. However, these methods require volumetric data sets, and do not work in real time. To address this shortcoming, Weiss *et al.* [200] presented a technique that uses five parallel iOCT B-scans to track the needle, detecting the elliptical section of the needle in each scan. The same group has also presented a similar technique, using a larger number of B-scans (128 along x , and the same number along y), to perform marker-free robot hand-eye calibration, which they have demonstrated in intact porcine eyes *ex vivo* [220]. Tracking of the needle after it enters tissue remains an open research problem [200], which the group has begun to address by registering the needle to a computer-aided design (CAD) model before the tip enters the retina, and then predicting the tip position and orientation after subretinal entry using the known input commands [221].

OCT information can be acquired not only through the pupil, but also through intraocular surgical instruments. The laboratory of J. U. Kang at Johns Hopkins University has developed common-path swept-source OCT (CP-SSOCT) with a fiber-optic probe that is fabricated to fit within a 26 G needle [99]. The system provides an OCT A-scan with resolution of $1.8\text{ }\mu\text{m}$ over a range of $3686.4\text{ }\mu\text{m}$ from the tip of the probe. Automated scanning of an OCT probe to acquire 3D retinal imagery using this technology was demonstrated using Micron [208, 209]. Yang *et al.* obtained stabilized OCT images of A-mode and M-mode scans in air and live rabbit eyes.

The Kang group has also demonstrated a wide range of capabilities using SMART handheld instruments combining their OCT technology with 1-DoF axial actuation (see §5.2). The technology can perform functions such as servoing to a selected stand-off distance from the surface [41], or actively compensating hand tremor along the axial dimension of the instrument [177]. They have also combined the technology with a motorized microforceps for epiretinal membranectomy, and have demonstrated improved accuracy and reduced task completion time in an artificial task involving picking up $125\text{ }\mu\text{m}$ optical fibers from a soft polymer surface [40]. They have used the depth servoing capability to perform subretinal injections with enhanced accuracy in a porcine retina *ex vivo* in an open-sky experiment [99]. Compared to freehand injection, where depth varied over a range of $200\text{ }\mu\text{m}$, the rms error of OCT-guided injection in gelatin and *ex vivo* bovine eyes stayed below $7.48\text{ }\mu\text{m}$ and $10.95\text{ }\mu\text{m}$, respectively [99]. The group has also developed the capability to calculate lateral displacement

from the value of cross-correlation coefficient based on the speckle model, and used this to demonstrate real-time estimation of scanning speed in freehand retinal scanning in order to reduce distortion due to motion artifacts [118].

8. Conclusion and Future Work

Robotic microsurgery is still in its infancy but has already begun to change the perspective of retinal surgery. Robotic retinal surgery has been successfully carried out in a limited number of patients, using a few surgical systems such as the PRECEYES Surgical System [46] or the system by KU Leuven [61]. Performing new treatments such as retinal vein cannulation becomes now technically feasible due to improved stability and tremor cancellation functionality. New developments such as among others augmented reality, haptic guidance, micrometer-scale distance sensing will further impact the efficiency and reliability of these interventions. Recent contributions have lead to the arrival of instruments featuring superior dexterity. These futuristic devices could initiate clinical efforts to design radically new surgical techniques. Innovations in material science, drug development, retinal chips, and gene and cell therapy are expected to create a whole new set of engineering challenges. Together with advances in imaging, robotics has become among the most promising trends in advancing the field of retinal microsurgery. As a result an increasing number of academic institutions embark on research projects to investigate ever more powerful systems. This section identifies the main challenges ahead, striving to outline the direction of the next decade of research in retinal surgery.

8.1. Practical Challenges Ahead

There are several challenges in the road ahead. Existing robotic systems are still very expensive and should be first adopted and accepted by both surgeons and patients to become sufficiently useful in the Operating Room (OR). At that point the OR culture will need to change. Dedicated training programs would need to be developed and robotic surgery should be included in the surgical curriculum. Further, robotic systems need to be developed so that currently impossible interventions are achieved, as, for example, retinal vein cannulation or subretinal delivery of novel therapeutics. Technical feasibility alone is not sufficient, as the safety and effectiveness of supplied substances and drugs must be validated as well. An important challenge for robot developers is hence to establish of a solid collaboration with the pharmaceutical industry. The adoption of robotic systems in commonplace procedures such as epiretinal membrane peeling, which despite its dexterity requirement is routinely and successfully performed in the western world, does not support the high cost of introducing a robot into the OR. The added value is too restricted for these scenarios. Therefore, we anticipate that the way forward for retinal surgical robotics will depend on a combination

of the following three key characteristics: **1)** system optimization including enhancing the usability, reduction of cost and miniaturization in order to reduce the space occupation in the OR; **2)** the capability to delivery targeted drugs and substances ultra-minimally invasively, opening the path to new treatment methods and **3)** automation to enable the parallel execution of a plurality of surgical procedures operated by a surgery supervisor.

8.2. System Optimization

In the case of developing robotic technology for microsurgery, more effort is needed in studying human factors to design more effective human-robot interfaces that are intuitive enough to perform complicated maneuvers inside the eye. Little attention has been paid so far to coordinate the control of multiple instruments at once. In manual interventions surgeons regularly reposition the eye to optimize the view angle after obtaining a good view they conduct then ultra-precise manipulations. While this concerns very different gestures surgeons are used to quickly switch between them. Virtual fixtures that coordinate and constrain the relative motion between instruments, such as proposed by Balicki [9], could be further explored to this end. Increased surgical time and cost remain serious concerns for robotic surgery. Several strategies can be followed to limit these concerns, such as for example making sure that robotic surgeons possess equal control over what happens with and within the eye. Another essential feature is the possibility to quickly exchange tools such as developed in prior work by Nambi *et al.* [140]. Further optimization of space would be needed as well. Especially in the case where the surgeon remains close to the patient space occupancy of the robotic device is crucial as it should not affect the surgeon’s already poor ergonomic conditions negatively. Multi-disciplinary teams should work together to understand how to build optimal compact and low-cost systems. Clinicians have traditionally worked together with biomedical engineers to design systems for specific applications, but have not been successful in translating these systems to the clinic. Most commonly, academic research occurs in isolation of the constraints that a real OR poses. For example, academic teams have designed robots that are challenging to integrate into existing clinical procedures, therefore limiting their adoption. It is time to pay greater attention to devise streamlined robot design approaches that consider more profoundly the constraints of the OR, staff position, and assistant/surgeon position, together with ergonomics, microscope constraints, and the challenges of the actual application at hand. The robotic retinal surgery community can therefore leverage the extensive work conducted by Padoy *et al.* on OR reconstruction and tracking [147, 191].

8.3. Novel Therapy Delivery Methods

Depth perception and control is difficult in retinal surgery in general but is especially problematic for subretinal injections where in the absence of visual feedback a precision in the order of $25\ \mu\text{m}$ is needed to ensure critical layers such as the retinal pigment epithelium are not damaged irreparably [221]. The development of OCT has opened up new perspectives in this context, offering the capacity to image disease on the micrometer level and at early disease states. This spurs the development of novel tools and delivery systems that allow interventions in early stages before major complications arise. As new drugs, new prosthesis, cell and gene therapy are being developed, we expect a growth in the development of new miniature delivery instruments and micro-injectors that, e.g. under iOCT guidance, deliver these therapeutic substances with extreme precision, targeting specific retinal layers [99, 221]. In this context microrobotics have made their appearance. Being the smallest representative of surgical and interventional devices, they offer tremendous opportunities to push miniaturization to the extreme. Ultimately they could enable interaction with few and even individual cells. Microrobots are one of the newest research areas in surgical robotics. Retinal surgery has been one of the major drivers for this technology. Microrobots have been proposed for intraocular drug delivery and retinal vein cannulation, and their mobility has been evaluated in animal models *in vivo*. The evaluated microrobots are propelled by electromagnetic fields (see §5.5). Electromagnetic-based actuation is preferred in small-scale actuation due to the favorable scaling of electromagnetic forces and torques with respect to device volume. Even though the minuscule size of the steerable magnetic devices makes the application of forces challenging currently, it can be expected that as the engineering capacity at the microscale levels matures, microdevices will become valuable tools of future retinal surgical operating rooms, primarily as means to precisely deliver novel therapeutics, and subsequently as mechanisms to enable ever more precise interventions.

8.4. Towards Autonomous Interventions

We expect a progressive adoption of automated features, similar to other fields in robotic surgery [103], which could ultimately lead to full autonomous execution of parts of the surgery. A long-term effort would enable a surgeon to supervise a set of robots that perform routine procedure steps autonomously and only call on his/her expertise during critical patient-specific steps. The surgeon would then guide the robot through the more complex tasks. Reaching this goal will require the analysis of data generated from a large number of interventions. Significant research on the topic, primarily on understanding the surgical phases of cataract surgery, has been conducted by Janin *et al.* [112], among others. Coupled with realistic retinal simulators, such as those developed by Cotin *et al.* [43], we expect that robots will be able to undertake certain

aspects of surgery, such as port placement and vitrectomy, in the near future. Visual servoing frameworks such as the ones developed by Riviere *et al.* [213] would enable automated cauterization of leaky vessels in diabetic retinopathy, therefore speeding up potentially lengthy interventions. Finally, the upcoming field of surgical data science (Maier-Hein *et al.* [128]) is expected to play an increasingly important role in robotic retinal surgery.

ACKNOWLEDGMENTS

Emmanuel B. Vander Poorten’s contribution to this publication was supported by EurEyeCase via the EU Framework Programme for Research and Innovation-Horizon 2020 (No. 645331) and STABLEYES an internal KU Leuven C3-fund [3E160419].

Cameron Riviere’s contribution to this publication was supported by the U. S. National Institutes of Health (grant no. R01EB000526).

Jake J. Abbott’s contribution to this publication was supported by the National Eye Institute of the National Institutes of Health under award Number R21EY027528.

Christos Bergeles’ contribution to this publication was supported by an ERC Starting Grant [714562]. Further, this report is independent research funded by the National Institute for Health Research (Invention for Innovation, i4i; II-LB-0716-20002).

M. Ali Nasseris contribution to this publication was supported by Ophthalmology Department of Klinikum rechts der Isar, TUM; State of Bavaria; and Carl Zeiss Meditec AG.

Jin U Kang’s contribution to this publication was supported by an NIH grant R01 EY021540 and Coulter Translational Fund.

Koorosh Faridpooya’s contribution was supported by the Foundation Rotterdam Eye Hospital and grants from the Foundation Coolsingel and by EurEyeCase via the EU Framework Programme for Research and Innovation-Horizon 2020 (No. 645331).

Iulian Iordachita’s contribution to this publication was supported by an U.S. National Institutes of Health, grant number 1R01EB023943.

The views expressed in this publication are those of the authors and not necessarily those of the National Institutes of Health, the NHS, the National Institute for Health Research, or the Department of Health.

1. A., H. and Goldbaum, M. (2013). Stare - structured analysis of the retina.
2. Allf, B. and De Juan, E. (1987). In vivo cannulation of retinal vessels. *Graefe's Archive for Clinical and Experimental Ophthalmology*, 225(3):221–225.
3. Almony, A., Nudleman, E., Shah, G. K., Blinder, K. J., Elliott, D. B., Mittra, R. A., and Tewari, A. (2012). Techniques, rationale, and outcomes of internal limiting membrane peeling. *Retina*, 32(5):877–891.
4. Ang, W. T., Riviere, C. N., and Khosla, P. K. (2000). An active hand-held instrument for enhanced microsurgical accuracy. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pages 878–886. Springer.
5. Appiah, A. P., Hirose, T., and Kado, M. (1988). A review of 324 cases of idiopathic premacular gliosis. *American Journal of Ophthalmology*, 106(5):533–535.
6. Balicki, M., Han, J.-H., Iordachita, I., Gehlbach, P., Handa, J., Taylor, R., and Kang, J. (2009). Single fiber optical coherence tomography microsurgical instruments for computer and robot-assisted retinal surgery. *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pages 108–115.
7. Balicki, M., Uneri, A., Iordachita, I., Handa, J., Gehlbach, P., and Taylor, R. (2010). Micro-force sensing in robot assisted membrane peeling for vitreoretinal surgery. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pages 303–310. Springer.
8. Balicki, M., Xia, T., Jung, M. Y., Deguet, A., Vagvolgyi, B., Kazanzides, P., and Taylor, R. (2011). Prototyping a hybrid cooperative and tele-robotic surgical system for retinal microsurgery. *MIDAS journal*, 2011:815.
9. Balicki, M. A. (2014). *Augmentation of human skill in microsurgery*. PhD thesis, The Johns Hopkins University.
10. Bandara, A. M. R. R. and Giragama, P. W. G. R. M. P. B. (2017). A retinal image enhancement technique for blood vessel segmentation algorithm. In *IEEE International Conference on Industrial and Information Systems*, pages 1–5.
11. Becker, B. C., MacLachlan, R. A., Lobes, L. A., Hager, G. D., and Riviere, C. N. (2013). Vision-based control of a handheld surgical micromanipulator with virtual fixtures. *IEEE Transactions on Robotics*, 29(3):674–683.
12. Becker, B. C., MacLachlan, R. A., Lobes Jr, L. A., and Riviere, C. N. (2010a). Semiautomated intraocular laser surgery using handheld instruments. *Lasers in Surgery and Medicine*, 42(3):264–273.
13. Becker, B. C., MacLachlan, R. A., and Riviere, C. N. (2011). State estimation and feedforward tremor suppression for a handheld micromanipulator with a Kalman filter. In *IEEE/RSJ International Conference on Intelligent Robots and Systems*, pages 5160–5165.
14. Becker, B. C., Voros, S., Lobes Jr, L. A., Handa, J. T., Hager, G. D., and Riviere, C. N. (2010b). Retinal vessel cannulation with an image-guided handheld robot. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 5420–5423.
15. Ben Gayed, M., Guerrouad, A., Diaz, C., Lepers, B., and Vidal, P. (1987). An advanced control micromanipulator for surgical applications. *Systems Science*, 13(1–2):123–134.
16. Bergeles, C., Kratochvil, B. E., and Nelson, B. J. (2012). Visually servoing magnetic intraocular microdevices. *IEEE Transactions on Robotics*, 28(4):798–809.
17. Bergeles, C., Kummer, M. P., Kratochvil, B. E., Framme, C., and Nelson, B. J. (2011). Steerable intravitreal inserts for drug delivery: in vitro and ex vivo mobility experiments. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pages 33–40. Springer.
18. Bergeles, C., Shamaei, K., Abbott, J. J., and Nelson, B. J. (2010). Single-camera focus-based localizing of intraocular devices. *IEEE Transactions on Biomedical Engineering*, 57(8):2064–2074.
19. Bergeles, C., Sugathapala, M., and Yang, G.-Z. (2018). Retinal surgery with flexible robots: Biomechanical advantages. In *International Symposium on Medical Robotics*, pages 1–6.
20. Berkelman, P. J., Whitcomb, L. L., Taylor, R. H., and Jensen, P. (2000). A miniature instrument tip force sensor for robot/human cooperative microsurgical manipulation with enhanced force feedback. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pages 897–906. Springer.
21. Berkelman, P. J., Whitcomb, L. L., Taylor, R. H., and Jensen, P. (2003). A miniature microsurgical

- instrument tip force sensor for enhanced force feedback during robot-assisted manipulation. *IEEE Transactions on Robotics and Automation*, 19(5):917–921.
22. Boppart, S. A., Brezinski, M. E., and Fujimoto, J. G. (2001). Surgical guidance and intervention.
23. Borghesan, G., Ourak, M., Lankenau, E., Hüttmann, G., Schulz-Hildebrandt, H., Willekens, K., Stalmans, P., Reynaerts, D., and Vander Poorten, E. (2018). Single scan OCT-based retina detection for robot-assisted retinal vein cannulation. *Journal of Medical Robotics Research*, 3(02):1840005.
24. Bourla, D. H., Hubschman, J. P., Culjat, M., Tsirbas, A., Gupta, A., and Schwartz, S. D. (2008). Feasibility study of intraocular robotic surgery with the da vinci surgical system. *Retina*, 28(1):154–158.
25. Braun, D., Yang, S., Martel, J. N., Riviere, C. N., and Becker, B. C. (2018). EyeSLAM: Real-time simultaneous localization and mapping of retinal vessels during intraocular microsurgery. *International Journal of Medical Robotics and Computer Assisted Surgery*, 14(1):e1848.
26. Brod, R. D. (2009). Surgery for diseases of the vitreous and retina. *The Journal of Lancaster General Hospital*, 4(1):4–9.
27. Brosilow, C. and Joseph, B. (2002). *Techniques of model-based control*. Prentice-Hall.
28. Burschka, D., Corso, J. J., Dewan, M., Lau, W., Li, M., Lin, H., Marayong, P., Ramey, N., Hager, G. D., Hoffman, B., Larkin, D., and Hasser, C. (2005). Navigating inner space: 3-d assistance for minimally invasive surgery. *Robotics and Autonomous Systems*, 52:5–26.
29. Caers, P., Gijbels, A., De Volder, M., Gorissen, B., Stalmans, P., Reynaerts, D., and Vander Poorten, E. B. (2011). Precision experiments on a comanipulated robotic system for use in retinal surgery. In *Proceedings of the SCATh Joint Workshop on New Technologies for Computer/Robot Assisted Surgery*, pages 1–7.
30. Can, A., Shen, H., Turner, J. N., Tanenbaum, H. L., and Roysam, B. (1999). Rapid automated tracing and feature extraction from retinal fundus images using direct exploratory algorithms. *IEEE Transactions on Information Technology in Biomedicine*, 3(2):125–138.
31. Can, A., Stewart, C. V., Roysam, B., and Tanenbaum, H. L. (2000). A feature-based technique for joint, linear estimation of high-order image-to-mosaic transformations: application to mosaicing the curved human retina. In *Proceedings of the IEEE Conference on Computer Vision and Pattern Recognition*, volume 2, pages 585–591.
32. Cao, K., Pinon, R., Schachar, I., Jayasundera, T., and Awatar, S. (2014). Automatic instrument tracking endo-illuminator for intra-ocular surgeries. *Journal of Medical Devices*, 8(3):030932.
33. Carrasco-Zevallos, O. M., Viehland, C., Keller, B., Draelos, M., Kuo, A. N., Toth, C. A., and Izatt, J. A. (2017). Review of intraoperative optical coherence tomography: technology and applications. *Biomedical Optics Express*, 8(3):1607–1637.
34. Castaño, A. and Hutchinson, S. (1994). Visual compliance: Task-directed visual servo control. *IEEE Transactions on Robotics and Automation*, 10(3):334–342.
35. Cereda, M. G., K., F., J.C., V., van Romunde, S., Naus, G., Beelen, M., Douven, Y., D. d. M., and EurEyeCase (2018). First in-human clinical evaluation of a robot-controlled instrument with a real-time distance sensor in the vitreous cavity.
36. Charles, S. (1996). Dexterity enhancement for surgery. *Computer Integrated Surgery: Technology and Clinical Applications*, pages 467–471.
37. Charles, S. (2003). Techniques and tools for dissection of epiretinal membranes. *Graefe’s Archive for Clinical and Experimental Ophthalmology*, 241(5):347–352.
38. Charles, S. and Calzada, J. (2010). *Vitreous Microsurgery*. Lippincott Williams & Wilkins.
39. Charles, S., Das, H., Ohm, T., Boswell, C., Rodriguez, G., Steele, R., and Istrate, D. (1997). Dexterity-enhanced telerobotic microsurgery. In *IEEE International Conference on Advanced Robotics*, pages 5–10.
40. Cheon, G. W., Gonenc, B., Taylor, R. H., Gehlbach, P. L., and Kang, J. U. (2017). Motorized microforceps with active motion guidance based on common-path ssoct for epiretinal membranectomy. *IEEE/ASME Transactions on Mechatronics*, 22(6):2440–2448.
41. Cheon, G. W., Huang, Y., Cha, J., Gehlbach, P. L., and Kang, J. U. (2015). Accurate real-time depth control for CP-SSOCT distal sensor based handheld microsurgery tools. *Biomedical Optics Express*, 6(5):1942–1953.

42. Choma, M. A., Sarunic, M. V., Yang, C., and Izatt, J. A. (2003). Sensitivity advantage of swept source and fourier domain optical coherence tomography. *Optics Express*, 11(18):2183–2189.
43. Cotin, S., Keppi, J. J., Allard, J., Bessard, R., Courtecuisse, H., and Gaucher, D. (2015). Project RESET. RETinal Surgery systEm for Training. *Acta Ophthalmologica*, 93(S255):ABS15–0477.
44. Cutler, N., Balicki, M., Finkelstein, M., Wang, J., Gehlbach, P., McGready, J., Iordachita, I., Taylor, R., and Handa, J. T. (2013). Auditory force feedback substitution improves surgical precision during simulated ophthalmic surgery. *Investigative Ophthalmology & Visual Science*, 54(2):1316–1324.
45. de Smet, M. D., Meenink, T. C. M., Janssens, T., Vanheukelom, V., Naus, G. J. L., Beelen, M. J., Meers, C., Jonckx, B., and Stassen, J.-M. (2016). Robotic assisted cannulation of occluded retinal veins. *PloS One*, 11(9):e0162037.
46. de Smet, M. D., Naus, G. J. L., Faridpooya, K., and Mura, M. (2018). Robotic-assisted surgery in ophthalmology. *Current Opinion in Ophthalmology*, 29(3):248–253.
47. Dewan, M., Marayong, P., Okamura, A. M., and Hager, G. D. (2004). Vision-based assistance for ophthalmic micro-surgery. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pages 49–57. Springer.
48. Drexler, W. and Fujimoto, J. G. (2008). *Optical coherence tomography: technology and applications*. Springer Science & Business Media.
49. Ebrahimi, A., He, C., Roizenblatt, M., Patel, N., Sefati, S., Gehlbach, P., and Iordachita, I. (2018). Real-time sclera force feedback for enabling safe robot-assisted vitreoretinal surgery. In *2018 40th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, pages 3650–3655. IEEE.
50. Edwards, T. L., Xue, K., Meenink, H. C. M., Beelen, M. J., Naus, G. J. L., Simunovic, M. P., Latasiewicz, M., Farmery, A. D., de Smet, M. D., and MacLaren, R. E. (2018a). First-in-human study of the safety and viability of intraocular robotic surgery. *Nature Biomedical Engineering*, 2:649–656.
51. Edwards, T. L., Xue, K., Meenink, T., Beelen, M., Naus, G., Simunovic, M. P., Latasiewicz, M., Farmery, A., de Smet, M. D., and MacLaren, R. E. (2018b). A first-in-man trial assessing robotic surgery inside the human eye to perform a subretinal injection. *Investigative Ophthalmology & Visual Science*, 59:5936.
52. Elhousseini, Z., Lee, E., and Williamson, T. H. (2016). Incidence of lens touch during pars plana vitrectomy and outcomes from subsequent cataract surgery. *Retina*, 36(4):825–829.
53. Ergeneman, O., Bergeles, C., Kummer, M. P., Abbott, J. J., and Nelson, B. J. (2011a). Wireless intraocular microrobots: opportunities and challenges. In Rosen, J., Hannaford, B., and Satava, R. M., editors, *Surgical Robotics: Systems Applications and Visions*, pages 271–311. Springer.
54. Ergeneman, O., Dogangil, G., Kummer, M. P., Abbott, J. J., Nazeeruddin, M. K., and Nelson, B. J. (2008). A magnetically controlled wireless optical oxygen sensor for intraocular measurements. *IEEE Sensors Journal*, 8(1):29–37.
55. Ergeneman, O., Pokki, J., Počepcová, V., Hall, H., Abbott, J. J., and Nelson, B. J. (2011b). Characterization of puncture forces for retinal vein cannulation. *Journal of Medical Devices*, 5(4):044504.
56. Feltgen, N., Junker, B., Agostini, H., and Hansen, L. L. (2007). Retinal endovascular lysis in ischemic central retinal vein occlusion: one-year results of a pilot study. *Ophthalmology*, 114(4):716–723.
57. Fercher, A. F., Hitzengerger, C. K., Kamp, G., and El-Zaiat, S. Y. (1995). Measurement of intraocular distances by backscattering spectral interferometry. *Optics Communications*, 117(1-2):43–48.
58. Fifanski, D., Rivera, J., Clogenson, M., Baur, M., Bertholds, A., Llosas, P., Wolfensberger, T., and Henein, S. (2017). VivoForce instrument for retinal microsurgery. In *Surgetica*, pages 155–157.
59. Fleming, I., Balicki, M., Koo, J., Iordachita, I., Mitchell, B., Handa, J., Hager, G., and Taylor, R. (2008). Cooperative robot assistant for retinal microsurgery. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pages 543–550. Springer.
60. Gijbels, A., Reynaerts, D., Stalmans, P., and Vander Poorten, E. (2014a). Design and manufacturing of a 2-dof force sensing needle for retinal surgery. In *4th Joint Workshop on Computer/Robot Assisted Surgery*, pages 71–74.
61. Gijbels, A., Smits, J., Schoevaerds, L., Willekens, K., Vander Poorten, E. B., Stalmans, P., and Reynaerts, D. (2018). In-human robot-assisted retinal vein cannulation, a world first. *Annals of Biomedical Engineering*, pages 1–10.

62. Gijbels, A., Vander Poorten, E. B., Gorissen, B., Devreker, A., Stalmans, P., and Reynaerts, D. (2014b). Experimental validation of a robotic comanipulation and telemanipulation system for retinal surgery. In *5th IEEE RAS/EMBS International Conference on Biomedical Robotics and Biomechatronics*, pages 144–150. IEEE.
63. Gijbels, A., Vander Poorten, E. B., Stalmans, P., and Reynaerts, D. (2015). Development and experimental validation of a force sensing needle for robotically assisted retinal vein cannulations. In *IEEE International Conference on Robotics and Automation*, pages 2270–2276.
64. Gijbels, A., Willekens, K., Esteveny, L., Stalmans, P., Reynaerts, D., and Vander Poorten, E. B. (2016). Towards a clinically applicable robotic assistance system for retinal vein cannulation. In *IEEE International Conference on Biomedical Robotics and Biomechatronics*, pages 284–291.
65. Gijbels, A., Wouters, N., Stalmans, P., Van Brussel, H., Reynaerts, D., and Vander Poorten, E. (2013). Design and realisation of a novel robotic manipulator for retinal surgery. In *IEEE/RSJ International Conference on Intelligent Robots and Systems*, pages 3598–3603.
66. Gonenc, B., Balicki, M. A., Handa, J., Gehlbach, P., Riviere, C. N., Taylor, R. H., and Iordachita, I. (2012). Preliminary evaluation of a micro-force sensing handheld robot for vitreoretinal surgery. In *IEEE/RSJ International Conference on Intelligent Robots and Systems*, pages 4125–4130.
67. Gonenc, B., Gehlbach, P., Handa, J., Taylor, R. H., and Iordachita, I. (2014a). Motorized force-sensing micro-forceps with tremor cancelling and controlled micro-vibrations for easier membrane peeling. In *IEEE RAS/EMBS International Conference on Biomedical Robotics and Biomechatronics*, pages 244–251.
68. Gonenc, B., Taylor, R. H., Iordachita, I., Gehlbach, P., and Handa, J. (2014b). Force-sensing microneedle for assisted retinal vein cannulation. In *IEEE Sensors Conference*, pages 698–701.
69. Grace, K. W., Colgate, J. E., Glucksberg, M. R., and Chun, J. H. (1993). A six degree of freedom micromanipulator for ophthalmic surgery. *IEEE International Conference on Robotics and Automation*, 1:630–635.
70. Griffin, J. A., Zhu, W., and Nam, C. S. (2017). The role of haptic feedback in robotic-assisted retinal microsurgery systems: A systematic review. *IEEE Transactions on Haptics*, 10(1):94–105.
71. Guerrouad, A. and Vidal, P. (1989). SMOS: stereotaxical microtelemanipulator for ocular surgery. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 879–880.
72. Gupta, A., Gonenc, B., Balicki, M., Olds, K., Handa, J., Gehlbach, P., Taylor, R. H., and Iordachita, I. (2014). Human eye phantom for developing computer and robot-assisted epiretinal membrane peeling. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 6864–6867.
73. Gupta, P. K., Jensen, P. S., and de Juan, E. (1999). Surgical forces and tactile perception during retinal microsurgery. In *International Conference on Medical Image Computing and Computer-Assisted Intervention*, pages 1218–1225. Springer.
74. Han, J.-H., Balicki, M., Zhang, K., Liu, X., Handa, J., Taylor, R., and Kang, J. U. (2009). Common-path fourier-domain optical coherence tomography with a fiber optic probe integrated into a surgical needle. In *Conference on Lasers and Electro-Optics*, page CMCC2. Optical Society of America.
75. Harwell, R. C. and Ferguson, R. L. (1983). Physiologic tremor and microsurgery. *Microsurgery*, 4(3):187–192.
76. Hashtrudi-Zaad, K. and Salcudean, S. E. (2001). Analysis of control architectures for teleoperation systems with impedance/admittance master and slave manipulators. *The International Journal of Robotics Research*, 20(6):419–445.
77. He, C., Ebrahimi, A., Roizenblatt, M., Patel, N., Yang, Y., Gehlbach, P. L., and Iordachita, I. (2018a). User behavior evaluation in robot-assisted retinal surgery. In *2018 27th IEEE International Symposium on Robot and Human Interactive Communication (RO-MAN)*, pages 174–179. IEEE.
78. He, C.-Y., Huang, L., Yang, Y., Liang, Q.-F., and Li, Y.-K. (2018b). Research and realization of a master-slave robotic system for retinal vascular bypass surgery. *Chinese Journal of Mechanical Engineering*, 31(1):78.
79. He, X., Balicki, M., Gehlbach, P., Handa, J., Taylor, R., and Iordachita, I. (2013). A novel dual force sensing instrument with cooperative robotic assistant for vitreoretinal surgery. In *IEEE International*

- Conference on Robotics and Automation*, pages 213–218.
80. He, X., Balicki, M., Gehlbach, P., Handa, J., Taylor, R., and Iordachita, I. (2014a). A multi-function force sensing instrument for variable admittance robot control in retinal microsurgery. In *IEEE International Conference on Robotics and Automation*, pages 1411–1418.
 81. He, X., Balicki, M. A., Kang, J. U., Gehlbach, P. L., Handa, J. T., Taylor, R. H., and Iordachita, I. I. (2012a). Force sensing micro-forceps with integrated fiber bragg grating for vitreoretinal surgery. In *Optical Fibers and Sensors for Medical Diagnostics and Treatment Applications XII*, volume 8218, page 82180W. International Society for Optics and Photonics.
 82. He, X., Handa, J., Gehlbach, P., Taylor, R., and Iordachita, I. (2014b). A submillimetric 3-DOF force sensing instrument with integrated fiber bragg grating for retinal microsurgery. *IEEE Transactions on Biomedical Engineering*, 61(2):522–534.
 83. He, X., Roppenecker, D., Gierlach, D., Balicki, M., Olds, K., Gehlbach, P., Handa, J., Taylor, R., and Iordachita, I. (2012b). Toward clinically applicable steady-hand eye robot for vitreoretinal surgery. In *ASME International Mechanical Engineering Congress and Exposition*, pages 145–153. American Society of Mechanical Engineers.
 84. He, X., Van Geirt, V., Gehlbach, P., Taylor, R., and Iordachita, I. (2015). IRIS: Integrated robotic intraocular snake. In *IEEE International Conference on Robotics and Automation*, pages 1764–1769.
 85. Hoover, A., Kouznetsova, V., and Goldbaum, M. (2000). Locating blood vessels in retinal images by piecewise threshold probing of a matched filter response. *IEEE Transactions on Medical imaging*, 19(3):203–210.
 86. Horise, Y., He, X., Gehlbach, P., Taylor, R., and Iordachita, I. (2015). Fbg-based sensorized light pipe for robotic intraocular illumination facilitates bimanual retinal microsurgery. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 13–16.
 87. Huang, Y., Liu, X., and Kang, J. U. (2012). Real-time 3D and 4D Fourier domain Doppler optical coherence tomography based on dual graphics processing units. *Biomedical optics express*, 3(9):2162–2174.
 88. Hubschman, J., Son, J., Allen, B., Schwartz, S., and Bourges, J. (2011). Evaluation of the motion of surgical instruments during intraocular surgery. *Eye*, 25(7):947.
 89. Hubschman, J. P., Bourges, J. L., Choi, W., Mozayan, A., Tsirbas, A., Kim, C. J., and Schwartz, S. D. (2010). the microhand: a new concept of micro-forceps for ocular robotic surgery. *Eye*, 24(2):364.
 90. Hunter, I. W., Doukoglou, T. D., Lafontaine, S. R., Charette, P. G., Jones, L. A., Sagar, M. A., Mallinson, G. D., and Hunter, P. J. (1993). A teleoperated microsurgical robot and associated virtual environment for eye surgery. *Presence: Teleoperators & Virtual Environments*, 2(4):265–280.
 91. Hunter, I. W., Lafontaine, S., Nielsen, P. M. F., Hunter, P. J., and Hollerbach, J. M. (1990). Manipulation and dynamic mechanical testing of microscopic objects using a tele-micro-robot system. In *IEEE International Conference on Robotics and Automation*, pages 3–9.
 92. Ida, Y., Sugita, N., Ueta, T., Tamaki, Y., Tanimoto, K., and Mitsuishi, M. (2012). Microsurgical robotic system for vitreoretinal surgery. *International Journal of Computer Assisted Radiology and Surgery*, 7(1):27–34.
 93. Ikuta, K., Kato, T., and Nagata, S. (1996). Micro active forceps with optical fiber scope for intraocular microsurgery. In *Micro Electro Mechanical Systems, 1996, MEMS’96, Proceedings. An Investigation of Micro Structures, Sensors, Actuators, Machines and Systems. IEEE, The Ninth Annual International Workshop on*, pages 456–461. IEEE.
 94. Ikuta, K., Kato, T., and Nagata, S. (1998). Optimum designed micro active forceps with built-in fiberscope for retinal microsurgery. *Medical Image Computing and Computer Assisted Intervention*, LNCS 1496:411–420.
 95. Iordachita, I., Sun, Z., Balicki, M., Kang, J. U., Phee, S. J., Handa, J., Gehlbach, P., and Taylor, R. (2009). A sub-millimetric, 0.25 mN resolution fully integrated fiber-optic force-sensing tool for retinal microsurgery. *International Journal of Computer Assisted Radiology and Surgery*, 4(4):383–390.
 96. Jagtap, A. D. and Riviere, C. N. (2004). Applied force during vitreoretinal microsurgery with hand-held instruments. *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, 4:2771–2774.

97. Jensen, P. S., Grace, K. W., Attariwala, R., Colgate, J. E., and Glucksberg, M. R. (1997). Toward robot-assisted vascular microsurgery in the retina. *Graefe's Archive for Clinical and Experimental Ophthalmology*, 235(11):696–701.
98. Jingjing, X., Long, H., Lijun, S., and Yang, Y. (2014). Design and research of a robotic aided system for retinal vascular bypass surgery. *Journal of Medical Devices*, 8(4):044501.
99. Kang, J. and Cheon, G. (2018). Demonstration of subretinal injection using common-path swept source OCT guided microinjector. *Applied Sciences*, 8(8):1287.
100. Kang, J. U., Han, J.-H., Liu, X., Zhang, K., Song, C. G., and Gehlbach, P. (2010). Endoscopic functional fourier domain common-path optical coherence tomography for microsurgery. *IEEE Journal of Selected Topics in Quantum Electronics*, 16(4):781–792.
101. Kang, J. U., Huang, Y., Zhang, K., Ibrahim, Z., Cha, J., Andrew Lee, W., Brandacher, G., and Gehlbach, P. (2012). Real-time three-dimensional Fourier-domain optical coherence tomography video image guided microsurgies. *Journal of Biomedical Optics*, 17(8):081403–1.
102. Kasner, D. (1969). Vitrectomy: a new approach to management of vitreous. *Highlights Ophthalmol*, 11:304.
103. Kassahun, Y., Yu, B., Tibebu, A. T., Stoyanov, D., Giannarou, S., Metzen, J. H., and Vander Poorten, E. (2016). Surgical robotics beyond enhanced dexterity instrumentation: a survey of machine learning techniques and their role in intelligent and autonomous surgical actions. *International journal of computer assisted radiology and surgery*, 11(4):553–568.
104. Koch, P. (2017). Advanced vitreoretinal surgery. *Acta Ophthalmologica*, 95(S259).
105. Krug, M. and Lankenau, E. (2017). Wo2017167850 (a1) - oct system.
106. Kumar, R., Berkelman, P., Gupta, P., Barnes, A., Jensen, P. S., Whitcomb, L. L., and Taylor, R. H. (2000). Preliminary experiments in cooperative human/robot force control for robot assisted microsurgical manipulation. In *IEEE International Conference on Robotics and Automation*, volume 1, pages 610–617.
107. Kummer, M. P., Abbott, J. J., Dinser, S., and Nelson, B. J. (2007). Artificial vitreous humor for in vitro experiments. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 6406–6409.
108. Kummer, M. P., Abbott, J. J., Kratochvil, B. E., Borer, R., Sengul, A., and Nelson, B. J. (2010). OctoMag: An electromagnetic system for 5-dof wireless micromanipulation. *IEEE Transactions on Robotics*, 26(6):1006–1017.
109. Kurmann, T., Neila, P. M., Du, X., Fua, P., Stoyanov, D., Wolf, S., and Sznitman, R. (2017). Simultaneous Recognition and Pose Estimation of Instruments in Minimally Invasive Surgery. In *Medical Image Computing and Computer-Assisted Intervention*.
110. Kuru, I., Gonenc, B., Balicki, M., Handa, J., Gehlbach, P., Taylor, R. H., and Iordachita, I. (2012). Force sensing micro-forceps for robot assisted retinal surgery. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 1401–1404.
111. Laina, I., Rieke, N., Rupprecht, C., Vizcaíno, J. P., Eslami, A., Tombari, F., and Navab, N. (2017). Concurrent Segmentation and Localization for Tracking of Surgical Instruments. In *Medical Image Computing and Computer-Assisted Intervention*, pages 664–672.
112. Lalys, F., Riffaud, L., Bouget, D., and Jannin, P. (2012). A framework for the recognition of high-level surgical tasks from video images for cataract surgeries. *IEEE Transactions on Biomedical Engineering*, 59(4):966–976.
113. Lawrence, D. A. (1992). Designing teleoperator architectures for transparency. In *Robotics and Automation, 1992. Proceedings., 1992 IEEE International Conference on*, pages 1406–1411. IEEE.
114. Leitgeb, R., Hitzenberger, C., and Fercher, A. F. (2003). Performance of Fourier domain vs. time domain optical coherence tomography. *Optics Express*, 11(8):889–894.
115. Leng, T., Miller, J. M., Bilbao, K. V., Palanker, D. V., Huie, P., and Blumenkranz, M. S. (2004). The chick chorioallantoic membrane as a model tissue for surgical retinal research and simulation. *Retina*, 24(3):427–434.
116. Lin, F.-Y., Bergeles, C., and Yang, G.-Z. (2015). Biometry-based concentric tubes robot for vitreoretinal surgery. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 5280–5284.

117. Liu, G., Zhang, J., Yu, L., Xie, T., and Chen, Z. (2009). Real-time polarization-sensitive optical coherence tomography data processing with parallel computing. *Applied Optics*, 48(32):6365–6370.
118. Liu, X., Huang, Y., and Kang, J. U. (2012a). Distortion-free freehand-scanning OCT implemented with real-time scanning speed variance correction. *Optics Express*, 20(15):16567–16583.
119. Liu, X., Iordachita, I., He, X., Taylor, R., and Kang, J. (2012b). Miniature fiber-optic force sensor based on low-coherence Fabry-Perot interferometry for vitreoretinal microsurgery. *Biomedical Optics Express*, 3(5):1062–1076.
120. Liu, X., Li, X., Kim, D.-H., Ilev, I., and Kang, J. U. (2008). Fiber-optic Fourier-domain common-path OCT. *Chinese Optics Letters*, 6(12):899–901.
121. Mablekos-Alexiou, A., Ourselin, S., Da Cruz, L., and Bergeles, C. (2018). Requirements based design and end-to-end dynamic modeling of a robotic tool for vitreoretinal surgery. In *IEEE International Conference on Robotics and Automation*, pages 135–141.
122. Machemer, R. (1995). The development of pars plana vitrectomy: a personal account. *Graefes Archive for Clinical and Experimental Ophthalmology*, 233(8):453–468.
123. MacLachlan, R. A., Becker, B. C., Cuevas Tabarés, J., Podnar, G. W., Lobes Jr, L. A., and Riviere, C. N. (2012). Micron: an actively stabilized handheld tool for microsurgery. *IEEE Transactions on Robotics*, 28(1):195–212.
124. MacLachlan, R. A., Hollis, R. L., Jaramaz, B., Martel, J. N., Urish, K. L., and Riviere, C. N. (2017). Multirate Kalman filter rejects impulse noise in frequency-domain-multiplexed tracker measurements. In *IEEE Sensors Conference*, pages 591–593.
125. MacLachlan, R. A., Parody, N., Mukherjee, S., Hollis, R. L., and Riviere, C. N. (2016). Electro-magnetic tracker for active handheld robotic systems. In *IEEE Sensors Conference*, pages 52–54.
126. MacLachlan, R. A. and Riviere, C. N. (2009). High-speed microscale optical tracking using digital frequency-domain multiplexing. *IEEE Transactions on Instrumentation and Measurement*, 58(6):1991–2001.
127. Mahoney, A. W., Nelson, N. D., Parsons, E. M., and Abbott, J. J. (2012). Non-ideal behaviors of magnetically driven screws in soft tissue. In *IEEE/RSJ International Conference on Robotics and Intelligent Systems*, pages 3559–3564.
128. Maier-Hein, L., Vedula, S. S., Speidel, S., Navab, N., Kikinis, R., Park, A., Eisenmann, M., Feussner, H., Forestier, G., Giannarou, S., et al. (2017). Surgical data science for next-generation interventions. *Nature Biomedical Engineering*, 1(9):691–696.
129. Matinfar, S., Nasser, M. A., Eck, U., Kowalsky, M., Roodaki, H., Navab, N., Lohmann, C. P., Maier, M., and Navab, N. (2018). Surgical soundtracks: automatic acoustic augmentation of surgical procedures. *International Journal of Computer Assisted Radiology and Surgery*, 13(9):1345–1355.
130. McCannel, C. A., Olson, E. J., Donaldson, M. J., Bakri, S. J., Pulido, J. S., and Donna, M. (2012). Snoring is associated with unexpected patient head movement during monitored anesthesia care vitreoretinal surgery. *Retina*, 32(7):1324–1327.
131. McIntosh, R. L., Rogers, S. L., Lim, L., Cheung, N., Wang, J. J., Mitchell, P., Kowalski, J. W., Nguyen, H. P., and Wong, T. Y. (2010). Natural history of central retinal vein occlusion: an evidence-based systematic review. *Ophthalmology*, 117(6):1113–1123.
132. Meenink, H. C. M., Hendrix, R., Naus, G. J. L., Beelen, M. J., Nijmeijer, H., Steinbuch, M., van Oosterhout, E. J. G. M., and de Smet, M. D. (2012). Robot-assisted vitreoretinal surgery. In Gomes, P., editor, *Medical Robotics: Minimally Invasive Surgery*, pages 185–209. Woodhead Publishing.
133. Meenink, T., Naus, G., de Smet, M., Beelen, M., and Steinbuch, M. (2013). Robot assistance for micrometer precision in vitreoretinal surgery. *Investigative Ophthalmology & Visual Science*, 54(15):5808.
134. Mehta, S. and Hubbard III, G. B. (2013). Avoiding neck strain in vitreoretinal surgery: an ergonomic approach to indirect ophthalmoscopy and laser photocoagulation. *Retina*, 33(2):439–441.
135. Mitchell, P., Smith, W., Chey, T., Wang, J. J., and Chang, A. (1997). Prevalence and associations of epiretinal membranes: the Blue Mountains Eye Study, Australia. *Ophthalmology*, 104(6):1033–1040.
136. Mukherjee, S., Kaess, M., Martel, J. N., and Riviere, C. N. (2019). EyeSAM: Graph-based localization and mapping of retinal vasculature during intraocular microsurgery. *International Journal of Computer Assisted Radiology and Surgery*, page submitted.
137. Mukherjee, S., MacLachlan, R., and Riviere, C. N. (2016). Velocity-limiting control of an active

- handheld micromanipulator. *Journal of Medical Devices*, 10(3):030944–1–030944–2.
138. Mukherjee, S., Yang, S., MacLachlan, R. A., Lobes, L. A., Martel, J. N., and Riviere, C. N. (2017). Toward monocular camera-guided retinal vein cannulation with an actively stabilized handheld robot. In *IEEE International Conference on Robotics and Automation*, pages 2951–2956.
139. Nakano, T., Sugita, N., Ueta, T., Tamaki, Y., and Mitsuishi, M. (2009). A parallel robot to assist vitreoretinal surgery. *International Journal of Computer Assisted Radiology and Surgery*, 4(6):517.
140. Nambi, M., Bernstein, P. S., and Abbott, J. J. (2016). A compact telemanipulated retinal-surgery system that uses commercially available instruments with a quick-change adapter. *Journal of Medical Robotics Research*, 1(2):1630001.
141. Nambi, M., Bernstein, P. S., and Abbott, J. J. (2017). Effect of haptic-interface virtual kinematics on the performance and preference of novice users in telemanipulated retinal surgery. *IEEE Robotics and Automation Letters*, 2(1):64–71.
142. Nasser, M. A., Eder, M., Eberts, D., Nair, S., Maier, M., Zapp, D., Lohmann, C. P., and Knoll, A. (2013a). Kinematics and dynamics analysis of a hybrid parallel-serial micromanipulator designed for biomedical applications. In *IEEE/ASME International Conference on Advanced Intelligent Mechatronics*, pages 293–299.
143. Nasser, M. A., Eder, M., Nair, S., Dean, E., Maier, M., Zapp, D., Lohmann, C. P., and Knoll, A. (2013b). The introduction of a new robot for assistance in ophthalmic surgery. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 5682–5685.
144. Nassif, N., Cense, B., Park, B. H., Yun, S. H., Chen, T. C., Bouma, B. E., Tearney, G. J., and de Boer, J. F. (2004). In vivo human retinal imaging by ultrahigh-speed spectral domain optical coherence tomography. *Optics Letters*, 29(5):480–482.
145. Nelson, N. D., Delacenserie, J., and Abbott, J. J. (2013). An empirical study of the role of magnetic, geometric, and tissue properties on the turning radius of magnetically driven screws. In *IEEE International Conference on Robotics and Automation*, pages 5352–5357.
146. Ourak, M., Smits, J., Esteveny, L., Borghesan, G., Gijbels, A., Schoevaerdt, L., Douven, Y., Scholtes, J., Lankenau, E., Eixmann, T., et al. (2018). Combined oct distance and fbg force sensing cannulation needle for retinal vein cannulation: in vivo animal validation. *International Journal of Computer Assisted Radiology and Surgery*, pages 1–9.
147. Padoy, N., Blum, T., Feussner, H., Berger, M.-O., and Navab, N. (2008). On-line recognition of surgical activity for monitoring in the operating room. In *AAAI*, pages 1718–1724.
148. Patkin, M. (1977). Ergonomics applied to the practice of microsurgery. *Australian and New Zealand Journal of Surgery*, 47(3):320–329.
149. Peng, Y., Tang, L., and Zhou, Y. (2017). Subretinal injection: A review on the novel route of therapeutic delivery for vitreoretinal diseases. *Ophthalmic Research*, 58(4):217–226.
150. Peral-Gutierrez, F., Liao, A. L., and Riviere, C. N. (2004). Static and dynamic accuracy of vitreoretinal surgeons. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, volume 1, pages 2734–2737.
151. Peters, T. and Cleary, K. (2008). *Image-guided interventions: technology and applications*. Springer Science & Business Media.
152. Pezzementi, Z., Voros, S., and Hager, G. D. (2009). Articulated object tracking by rendering consistent appearance parts. In *IEEE International Conference on Robotics and Automation, 2009.*, pages 3940–3947.
153. Probst, J., Hillmann, D., Lankenau, E. M., Winter, C., Oelckers, S., Koch, P., and Hüttmann, G. (2010). Optical coherence tomography with online visualization of more than seven rendered volumes per second. *Journal of Biomedical Optics*, 15(2):026014.
154. Probst, T., Maninis, K.-K., Chhatkuli, A., Ourak, M., Vander Poorten, E., and Van Gool, L. (2018). Automatic tool landmark detection for stereo vision in robot-assisted retinal surgery. *IEEE Robotics and Automation Letters*, 3(1):612–619.
155. Rahimy, E., Wilson, J., Tsao, T., Schwartz, S., and Hubschman, J. (2013). Robot-assisted intraocular surgery: development of the IRISS and feasibility studies in an animal model. *Eye*, 27(8):972.
156. Richa, R., Balicki, M., Sznitman, R., Meisner, E., Taylor, R., and Hager, G. (2012). Vision-based

- proximity detection in retinal surgery. *IEEE Transactions on Biomedical Engineering*, 59(8):2291–2301.
157. Richa, R., Linhares, R., Comunello, E., Von Wangenheim, A., Schnitzler, J.-Y., Wassmer, B., Guillemot, C., Thuret, G., Gain, P., Hager, G., et al. (2014). Fundus image mosaicking for information augmentation in computer-assisted slit-lamp imaging. *IEEE Transactions on Medical Imaging*, 33(6):1304–1312.
158. Rieke, N., Tan, D. J., Alsheakhali, M., Tombari, F., di San Filippo, C. A., Belagiannis, V., Eslami, A., and Navab, N. (2015). Surgical tool tracking and pose estimation in retinal microsurgery. In *MICCAI 2015*, pages 266–273.
159. Rieke, N., Tan, D. J., Tombari, F., Vizcaíno, J. P., di San Filippo, C. A., Eslami, A., and Navab, N. (2016). Real-time online adaption for robust instrument tracking and pose estimation. In *MICCAI 2016*, pages 422–430.
160. Rieke, N., Tombari, F., and Navab, N. (2018). Chapter 4 - computer vision and machine learning for surgical instrument tracking: Focus: Random forest-based microsurgical tool tracking. In Leo, M. and Farinella, G. M., editors, *Computer Vision for Assistive Healthcare*, Computer Vision and Pattern Recognition, pages 105 – 126. Academic Press.
161. Riviere, C. N., Ang, W. T., and Khosla, P. K. (2003). Toward active tremor canceling in handheld microsurgical instruments. *IEEE Transactions on Robotics and Automation*, 19(5):793–800.
162. Riviere, C. N. and Jensen, P. S. (2000). A study of instrument motion in retinal microsurgery. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, volume 1, pages 59–60.
163. Riviere, C. N., Rader, R. S., and Thakor, N. V. (1998). Adaptive canceling of physiological tremor for improved precision in microsurgery. *IEEE Transactions on Biomedical Engineering*, 45(7):839–846.
164. Rogers, S., McIntosh, R. L., Cheung, N., Lim, L., Wang, J. J., Mitchell, P., Kowalski, J. W., Nguyen, H., Wong, T. Y., Consortium, I. E. D., et al. (2010). The prevalence of retinal vein occlusion: pooled data from population studies from the United States, Europe, Asia, and Australia. *Ophthalmology*, 117(2):313–319.
165. Roodaki, H., Navab, N., Eslami, A., Stapleton, C., and Navab, N. (2017). Sonifeye: Sonification of visual information using physical modeling sound synthesis. *IEEE Transactions on Visualization and Computer Graphics*, 23(11):2366–2371.
166. Rosenberg, L. B. (1993). Virtual fixtures: Perceptual tools for telerobotic manipulation. In *IEEE Virtual Reality Annual International Symposium*, pages 76–82.
167. Routray, A., MacLachlan, R., Martel, J., and Riviere, C. (2019). Incremental intraoperative update of retinal reconstruction using laser aiming beam. In *International Symposium on Medical Robotics*, page submitted.
168. Saito, H., Mitsubayashi, K., and Togawa, T. (2006). Detection of needle puncture to blood vessel by using electric conductivity of blood for automatic blood sampling. *Sensors and Actuators A: Physical*, 125(2):446–450.
169. Schenker, P. S., Das, H., and Ohm, T. R. (1995). A new robot for high dexterity microsurgery. In *Computer Vision, Virtual Reality and Robotics in Medicine*, pages 115–122. Springer.
170. Schoevaerdts, L., Esteveny, L., Borghesan, G., Ourak, M., Gijbels, A., Smits, J., Reynaerts, D., and Vander Poorten, E. (2018a). Innovative bio-impedance sensor towards puncture detection in eye surgery for retinal vein occlusion treatment. In *IEEE International Conference on Robotics and Automation*, pages 1–6.
171. Schoevaerdts, L., Esteveny, L., Borghesan, G., Ourak, M., Reynaerts, D., and Vander Poorten, E. (2018b). Automatic air bubble detection based on bio-impedance for safe drug delivery in retinal veins. In *Proceedings of the Hamlyn Symposium on Medical Robotics*, pages 7–8.
172. Sharma, U., Fried, N. M., and Kang, J. U. (2005). All-fiber common-path optical coherence tomography: sensitivity optimization and system analysis. *IEEE Journal of Selected Topics in Quantum Electronics*, 11(4):799–805.
173. Singh, K., Dion, C., Costantino, S., Wajszilber, M., Lesk, M. R., and Ozaki, T. (2009). In vivo measurement of the retinal movements using fourier domain low coherence interferometry. In *Conference on Lasers and Electro-Optics*, page CMR4. Optical Society of America.

174. Singh, K. D., Logan, N. S., and Gilmartin, B. (2006). Three-dimensional modeling of the human eye based on magnetic resonance imaging. *Investigative Ophthalmology & Visual Science*, 47(6):2272–2279.
175. Singh, S. P. N. and Riviere, C. N. (2002). Physiological tremor amplitude during retinal microsurgery. In *Proceedings of the IEEE Northeast Bioengineering Conference*, pages 171–172.
176. Smits, J., Ourak, M., Gijbels, A., Esteveny, L., Borghesan, G., Schoevaerds, L., Willekens, K., Stalmans, P., Lankenau, E., Schulz-Hildebrandt, H., et al. (2018). Development and experimental validation of a combined fbg force and oct distance sensing needle for robot-assisted retinal vein cannulation. In *IEEE International Conference on Robotics and Automation*, pages 129–134.
177. Song, C., Gehlbach, P. L., and Kang, J. U. (2012). Active tremor cancellation by a smart handheld vitreoretinal microsurgical tool using swept source optical coherence tomography. *Optics Express*, 20(21):23414–23421.
178. Song, C., Park, D. Y., Gehlbach, P. L., Park, S. J., and Kang, J. U. (2013). Fiber-optic OCT sensor guided smart micro-forceps for microsurgery. *Biomedical Optics Express*, 4(7):1045–1050.
179. Song, J., Gonenc, B., Gua, J., and Iordachita, I. (2017). Intraocular snake integrated with the steady-hand eye robot for assisted retinal microsurgery. In *IEEE International Conference on Robotics and Automation*, pages 6724–6729.
180. Spaide, R. F. (2002). Macular hole repair with minimal vitrectomy. *Retina*, 22(2):183–186.
181. Staal, J., Abramoff, M. D., Niemeijer, M., Viergever, M. A., and Van Ginneken, B. (2004). Ridge-based vessel segmentation in color images of the retina. *IEEE transactions on medical imaging*, 23(4):501–509.
182. Stout, J. T. and Francis, P. J. (2011). Surgical approaches to gene and stem cell therapy for retinal disease. *Human Gene Therapy*, 22(5):531–535.
183. Sun, Z., Balicki, M., Kang, J., Handa, J., Taylor, R., and Iordachita, I. (2009). Development and preliminary data of novel integrated optical micro-force sensing tools for retinal microsurgery. In *IEEE International Conference on Robotics and Automation*, pages 1897–1902.
184. Sunshine, S., Balicki, M., X., H., Olds, K., Kang, J., Gehlbach, P., Taylor, R., and Iordachita, I. (2013). A force-sensing microsurgical instrument that detects forces below human tactile sensation. *Retina*, 33:200–206.
185. Sznitman, R., Ali, K., Richa, R., Taylor, R. H., Hager, G. D., and Fual, P. (2012). Data-driven visual tracking in retinal microsurgery. In *MICCAI 2012*, pages 568–75.
186. Sznitman, R., Becker, C., and Fuá, P. G. (2014a). Fast Part-Based Classification for Instrument Detection in Minimally Invasive Surgery. In *MICCAI 2014*.
187. Sznitman, R., Richa, R., Taylor, R. H., Jedynak, B., and Hager, G. D. (2014b). Unified detection and tracking of instruments during retinal microsurgery. *IEEE Transactions on Pattern Analysis and Machine Intelligence*, 35(5):1263–1273.
188. Tan, B., Wong, A., and Bizheva, K. (2018). Enhancement of morphological and vascular features in oct images using a modified bayesian residual transform. *Biomedical Optics Express*, 9(5):2394–2406.
189. Tao, Y. K., Ehlers, J. P., Toth, C. A., and Izatt, J. A. (2010). Intraoperative spectral domain optical coherence tomography for vitreoretinal surgery. *Optics Letters*, 35(20):3315–3317.
190. The Editors of Encyclopaedia Britannica (2018). Retina.
191. Twinanda, A. P., Alkan, E. O., Gangi, A., de Mathelin, M., and Padoy, N. (2015). Data-driven spatio-temporal rgb-d feature encoding for action recognition in operating rooms. *International Journal of Computer Assisted Radiology and Surgery*, 10(6):737–747.
192. Ueta, T., Nakano, T., Ida, Y., Sugita, N., Mitsuishi, M., and Tamaki, Y. (2011). Comparison of robot-assisted and manual retinal vessel microcannulation in an animal model. *British Journal of Ophthalmology*, 95(5):731–734.
193. Ueta, T., Yamaguchi, Y., Shirakawa, Y., Nakano, T., Ideta, R., Noda, Y., Morita, A., Mochizuki, R., Sugita, N., Mitsuishi, M., et al. (2009). Robot-assisted vitreoretinal surgery: Development of a prototype and feasibility studies in an animal model. *Ophthalmology*, 116(8):1538–1543.
194. Üneri, A., Balicki, M. A., Handa, J., Gehlbach, P., Taylor, R. H., and Iordachita, I. (2010). New steady-hand eye robot with micro-force sensing for vitreoretinal surgery. In *IEEE RAS/EMBS International Conference on Biomedical Robotics and Biomechatronics*, pages 814–819.

195. van Overdam, K. A., Kilic, E., Verdijk, R. M., and Manning, S. (2018). Intra-ocular diathermy forceps. *Acta Ophthalmologica*, 96(4):420–422.
196. Vander Poorten, E., Esteveny, L., Gijbels, A., Rosa, B., Schoevaerds, L., Willekens, K., Stalmans, P., Reynaerts, D., Naus, G., Beelen, M., et al. (2015). Use case for European robotics in ophthalmologic micro-surgery. In *Proceedings of the 5th Joint Workshop on New Technologies for Computer/Robot Assisted Surgery*, pages 10–12.
197. Wei, W., Goldman, R., Simaan, N., Fine, H., and Chang, S. (2007). Design and theoretical evaluation of micro-surgical manipulators for orbital manipulation and intraocular dexterity. In *IEEE International Conference on Robotics and Automation*, pages 3389–3395.
198. Wei, W., Goldman, R. E., Fine, H. F., Chang, S., and Simaan, N. (2009). Performance evaluation for multi-arm manipulation of hollow suspended organs. *IEEE Transactions on Robotics*, 25(1):147–157.
199. Wei, W., Popplewell, C., Chang, S., Fine, H. F., and Simaan, N. (2010). Enabling technology for microvascular stenting in ophthalmic surgery. *Journal of Medical Devices*, 4(1):014503.
200. Weiss, J., Rieke, N., Nasser, M. A., Maier, M., Eslami, A., and Navab, N. (2018). Fast 5DOF needle tracking in iOCT. *International Journal of Computer Assisted Radiology and Surgery*, 13(6):787–796.
201. Wells, T. S., Yang, S., MacLachlan, R. A., Handa, J. T., Gehlbach, P., and Riviere, C. (2013). Comparison of baseline tremor under various microsurgical conditions. In *IEEE International Conference on Systems, Man, and Cybernetics*, pages 1482–1487.
202. Wells, T. S., Yang, S., MacLachlan, R. A., Lobes Jr., L. A., Martel, J. N., and Riviere, C. N. (2016). Hybrid position/force control of an active handheld micromanipulator for membrane peeling. *International Journal of Medical Robotics and Computer Assisted Surgery*, 12(1):85–95.
203. Wilkins, J. R., Puliafito, C. A., Hee, M. R., Duker, J. S., Reichel, E., Coker, J. G., Schuman, J. S., Swanson, E. A., and Fujimoto, J. G. (1996). Characterization of epiretinal membranes using optical coherence tomography. *Ophthalmology*, 103(12):2142–2151.
204. Willekens, K., Gijbels, A., Schoevaerds, L., Esteveny, L., Janssens, T., Jonckx, B., Feyen, J. H., Meers, C., Reynaerts, D., Vander Poorten, E., et al. (2017). Robot-assisted retinal vein cannulation in an in vivo porcine retinal vein occlusion model. *Acta Ophthalmologica*, 95(3):270–275.
205. Wilson, J. T., Gerber, M. J., Prince, S. W., Chen, C.-W., Schwartz, S. D., Hubschman, J.-P., and Tsao, T.-C. (2018). Intraocular robotic interventional surgical system (IRISS): Mechanical design, evaluation, and masterslave manipulation. *International Journal of Medical Robotics and Computer Assisted Surgery*, 4:e1842.
206. Wojtkowski, M., Bajraszewski, T., Targowski, P., and Kowalczyk, A. (2003). Real-time in vivo imaging by high-speed spectral optical coherence tomography. *Optics Letters*, 28(19):1745–1747.
207. Wu, Z., Troll, J., Jeong, H.-H., Wei, Q., Stang, M., Ziemssen, F., Wang, Z., Dong, M., Schnichels, S., Qiu, T., et al. (2018). A swarm of slippery micropneumatics penetrates the vitreous body of the eye. *Science Advances*, 4(11):eaat4388.
208. Yang, S., Balicki, M., MacLachlan, R. A., Liu, X., Kang, J. U., Taylor, R. H., and Riviere, C. N. (2012). Optical coherence tomography scanning with a handheld vitreoretinal micromanipulator. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 948–951.
209. Yang, S., Balicki, M., Wells, T. S., MacLachlan, R. A., Liu, X., Kang, J. U., Handa, J. T., Taylor, R. H., and Riviere, C. N. (2013). Improvement of optical coherence tomography using active handheld micromanipulator in vitreoretinal surgery. In *Proceedings of the Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, pages 5674–5677.
210. Yang, S., Lobes Jr, L. A., Martel, J. N., and Riviere, C. N. (2015a). Handheld-automated microsurgical instrumentation for intraocular laser surgery. *Lasers in Surgery and Medicine*, 47(8):658–668.
211. Yang, S., MacLachlan, R. A., Martel, J. N., Lobes, L. A., and Riviere, C. N. (2016). Comparative evaluation of handheld robot-aided intraocular laser surgery. *IEEE Transactions on Robotics*, 32(1):246–251.
212. Yang, S., MacLachlan, R. A., and Riviere, C. N. (2015b). Manipulator design and operation for a six-degree-of-freedom handheld tremor-canceling microsurgical instrument. *IEEE/ASME Transactions on Mechatronics*, 20(2):761–772.
213. Yang, S., Martel, J. N., Lobes Jr, L. A., and Riviere, C. N. (2018). Techniques for robot-aided

- intraocular surgery using monocular vision. *International Journal of Robotics Research*, 37(8):931–952.
214. Yiu, G., Marra, K. V., Wagley, S., Krishnan, S., Sandhu, H., Kovacs, K., Kuperwaser, M., and Arroyo, J. G. (2013). Surgical outcomes after epiretinal membrane peeling combined with cataract surgery. *British Journal of Ophthalmology*, 97(9):1197–1201.
215. Yu, D.-Y., Cringle, S. J., and Constable, I. J. (1998). Robotic ocular ultramicrosurgery. *Australian and New Zealand Journal of Ophthalmology*, 26(Suppl.):S6–S8.
216. Yu, H., Shen, J.-H., Joos, K. M., and Simaan, N. (2016). Calibration and integration of B-mode optical coherence tomography for assistive control in robotic microsurgery. *IEEE/ASME Transactions on Mechatronics*, 21(6):2613–2623.
217. Yun, S.-H., Tearney, G. J., de Boer, J. F., Iftimia, N., and Bouma, B. E. (2003). High-speed optical frequency-domain imaging. *Optics Express*, 11(22):2953–2963.
218. Zhang, K. and Kang, J. U. (2010). Real-time 4D signal processing and visualization using graphics processing unit on a regular nonlinear-k Fourier-domain OCT system. *Optics Express*, 18(11):11772–11784.
219. Zhang, K. and Kang, J. U. (2011). Real-time intraoperative 4D full-range FD-OCT based on the dual graphics processing units architecture for microsurgery guidance. *Biomedical Optics Express*, 2(4):764–770.
220. Zhou, M., Hamad, M., Weiss, J., Eslami, A., Huang, K., Maier, M., Lohmann, C. P., Navab, N., Knoll, A., and Nasser, M. A. (2018a). Towards robotic eye surgery: Marker-free, online hand-eye calibration using optical coherence tomography images. *IEEE Robotics and Automation Letters*, 3(4):3944–3951.
221. Zhou, M., Huang, K., Eslami, A., Roodaki, H., Zapp, D., Maier, M., Lohmann, C. P., Knoll, A., and Nasser, M. A. (2018b). Precision needle tip localization using optical coherence tomography images for subretinal injection. In *IEEE International Conference on Robotics and Automation*, pages 4033–4040.
222. Zhou, M., Roodaki, H., Eslami, A., Chen, G., Huang, K., Maier, M., Lohmann, C. P., Knoll, A., and Nasser, M. A. (2017). Needle segmentation in volumetric optical coherence tomography images for ophthalmic microsurgery. *Applied Sciences*, 7(8):748.