The Version of Record of this manuscript has been published by Taylor & Francis in the International Journal of Audiology on 20 December 2019, available at

<https://www.tandfonline.com/doi/full/10.1080/14992027.2019.1705405>

Introduction

Direct Acoustic Cochlear Implants (DACI) are a recent category of acoustic hearing implants. They were developed to address the gap in treatment options for patients with severe to profound mixed hearing loss (Häusler, 2008). These patients receive inadequate benefit from hearing aids, bone conduction (BC) implants or middle ear surgery, while in many cases the sensorineural component of their hearing loss is not severe enough to warrant cochlear implantation. Over the past few years the safety and efficacy of DACI have been established by several clinical studies (Busch et al., 2013; Häusler, 2008; Kludt, Büchner, Schwab, Lenarz, & Maier, 2016; Kludt, D'hondt, Lenarz, & Maier, 2017; Lenarz et al., 2014, 2013; Schwab, Kludt, Maier, Lenarz, & Teschner, 2015). The DACI investigated in our study is an active semi-implantable hearing system that is composed of two parts. The external part comprises the digital sound processor and a radio-frequency coil. The implanted part is an implant with actuator, which is mounted in the mastoid cavity with a fixation system (Figure 1). A conventional stapes prosthesis is attached to the implant's actuator. The vibrating actuator drives the perilymph via the stapes prosthesis through a stapedotomy opening. The implant does not attach to the middle ear ossicles and thus bypasses any existing middle ear problems and compensates the sensorineural hearing loss with amplified vibrations. The DACI concept, method and implantation were explained in detail by Häusler et al. (Häusler, 2008), while the DACI system used in our study is described in more detail by Lenarz et al. (Lenarz et al, 2013).

Interestingly, subjects implanted with a DACI have shown significant improvements in their BC thresholds at 750, 1000 and 1500 Hz after implantation (Lenarz et al, 2013), an effect that has not been reported for subjects who had received another type of active middle ear implant (Baumgartner et al, 2010; Bruchhage et al, 2016). It is unlikely that the sensorineural cochlear function of the study subjects improved after implantation. A more plausible explanation is a mechanical origin for the pre-operative BC hearing loss, as indicated by Arnold et al. and Perez et al. (Arnold, Fawzy, Steinhoff, Kiefer, & Arnold, 2011; Perez, de Almeida, Nedzelski, Chen, 2009). They have reported on

improvements of the BC thresholds after stapes surgery. For DACI this could be an explanation, however, another explanation could be a change in the sound transmission pathways after DACI implantation contributing to BC sound perception. When measuring BC thresholds after implantation, the DACI is switched off. Thus, it cannot actively influence the transmission pathways, but, it can influence them in a passive way.

Stenfelt (Stenfelt, 2016) described five components as being most important for BC sound perception in the normal and impaired ear: (1) sound pressure in the ear canal, (2) inertia of the middle ear ossicles, (3) inertia of cochlear fluids, (4) compression of the cochlear space and (5) pressure transmission from the skull interior. He has investigated the relative importance of these five pathways with an acoustic-impedance model of the inner ear and found that the three components middle ear inertia, inner ear fluid inertia, and inner ear compression dominated the excitation of the cochlea with BC stimulation at the mastoid. Of these three, inner ear fluid inertia seemed to be the most important. Ear canal sound pressure and intracranial sound pressure were the least important contributors (Stenfelt, 2016). If one of the three main contributors is unchanged, only small changes in BC sensitivity are expected. In normal hearing subjects, all pathways contribute to BC sound perception. Typical candidates for a DACI are patients with otosclerosis. In these patients, BC sounds generated within the outer and middle ear (pathways 1 and 2) cannot contribute effectively to sound perception because the stapes footplate is fixed. While a fixed stapes footplate and an interruption of the incudo-stapedial joint produced a 30-40 dB decrease in intra-cochlear sound pressures during air conduction stimulation in chinchilla, the same manipulation only had little effect on the intra-cochlear sound pressures during BC stimulation (Chhan, Bowers, McKinnon, & Rosowski, 2016). In human temporal bones a decrease of both intra-cochlear sound pressures by 20 dB after stapes fixation has been reported when stimulating via BC (Stieger et al, 2015). Therefore, it has been assumed that for BC stimulation, the inner ear sources are dominant compared to the sources from the middle and outer ear (Chhan et al., 2016). Clinically, otosclerosis appears as a depression in BC thresholds known as the

‘Carhart notch’ (Carhart, 1949). This notch is most pronounced around 1.7 kHz, the resonance frequency of the ossicular chain (Homma, Du, Shimizu, & Puria, 2009), but is also seen at lower frequencies (Arnold et al, 2011; Perez, de Almeida, Nedzelski, Chen, 2009; Sabbe, Verhaert, Joossen, Lammens, & Debruyne, 2015). During conventional stapes surgery, pathways 1 and 2 are reconstructed and sound can be transmitted via these pathways again. Additionally, changes in the impedance at the oval window can affect the contribution of the inner ear BC sources after stapes surgery. As a result, the Carhart notch often disappears. This is an explanation for the improvement of the BC thresholds after conventional stapes surgery. However, it does not hold in the same way for DACI. During DACI surgery the ossicular chain is interrupted, which means that pathways 1 and 2 are not reconstructed. Rather DACI stimulate the inner ear directly via a stapes prosthesis and bypass the middle ear. A more efficient contribution from the inner ear components could also be expected after DACI implantation due to the changed impedance at the oval window by the stapedotomy. So, can the improved BC thresholds after DACI implantation be explained by the changed impedance at the OW alone? Our hypothesis was that an additional pathway for BC is created via the fixation system, the actuator and the stapes prosthesis (B in Figure 1). Although the ossicular chain is not reconstructed, there is a direct connection from the skull surface to the inner ear via the DACI system. It is important to know the influence of this new pathway on the post-operative BC thresholds as these are used to monitor the cochlear function after DACI implantation for the evaluation of the long-term safety of acoustic hearing implants.

We have investigated this additional pathway and its contribution to BC sound perception by measuring the promontory and round window membrane (RWM) velocities in human cadaver heads before and after DACI implantation. Additionally, we have tested the hypothesis that the fixation of the stapes footplate leads to a lower RWM response. Stenfelt et al. reported that in BC stimulation, the RW fluid volume displacement was very similar between a normal intact specimen and a specimen with a stapes footplate glued to the otic capsule (Stenfelt, Hato, & Goode, 2004a). With BC

stimulation, the important parameter to measure is the relative vibration of the RWM, i.e. the velocity difference between the RWM and the surrounding bone/promontory (Stenfelt, Hato, & Goode, 2004b).

The aim of this study was therefore to investigate the contribution of the additional pathway via the DACI system to BC sound perception by analyzing the relative vibration of the RWM.

Methods

Eight ears of five fresh-frozen cadaveric whole heads with an average age of 84.5 years, intact temporal bones and no history of ear disease were investigated in this trial. Whole heads were used to have realistic measurement conditions and to take into account all BC pathways. The heads were firmly held in a cloth covered plastic ring during all experiments. The middle ear transfer function (METF) was measured for each ear to compare it with the modified acceptance criteria of the ASTM F2504-05 standard (ASTM, 2005) by Rosowski et al. (Rosowski, Chien, Ravicz, 2007). The study was approved by the medical ethics committee of the University Hospitals Leuven.

The surgical approach included a mastoidectomy and a large posterior tympanotomy to gain access to the stapes and to the RWM. The secondary mucosal membrane of the round window was removed carefully. If needed, a minimal amount of bony overhang was removed to ascertain correct visualization of the RWM. For the measurement of the METF the foam tip of an ER-2 insert earphone was inserted in the ear canal together with the tube of an ER-7C probe microphone (both Etymotic Research, Inc.; Elk Grove Village, IL, USA). The earphone was driven by a sine sweep (100 Hz – 10 kHz) of 100 dB SPL, which was generated by a UPV audio analyzer (Rohde & Schwarz, Munich, Germany). The velocity of the stapes posterior crus (PC) was measured with a laser Doppler vibrometry (LDV) system (OFV-5000 Vibrometer controller and OFV-534 Compact Sensor Head, Polytec GmbH, Waldbronn, Germany), which was mounted to a Zeiss OMPI-1FC (Carl Zeiss Microscopy GmbH, Jena, Germany) via a A-HLV MM-30 micro-manipulator (Polytec GmbH). The Vibrometer Controller was set to 10 mm/s/V with a 20 kHz low-pass and a 100 Hz high-pass filter. To

improve reflection of the laser light, a small piece of reflective tape was placed on the PC. The laser measurement angle to the PC was estimated to be 70° (the visual estimation was $\pm 10^\circ$ precise). The LDV output was fed into the UPV analyzer, which extracted 100 rms values of 1/3 octave bands between 100 and 10000 Hz in dB V. The dB V values were transferred back to velocity in mm/s with the range chosen in the LDV settings (10 mm/s/V).

To calculate the normalized METF, the stapes velocity was divided by the sound pressure in the ear canal (ASTM, 2005). The METF measurements were performed as a quality check of the ears used for our measurements.

In the following sections, the procedures for three different measurement conditions are described.

Baseline Condition (normal): Intact Middle Ear

A B81 high output bone transducer (Radioear, New Eagle, PA, USA) was positioned on the mastoid via a standard audiological headband, 45 mm posterior to the posterior wall of the external ear canal, applying a nominal static force of approximately 5.4 N to the head. It was calibrated to be used on the mastoid before the measurements. The B81 was stimulated with continuous pure tones at the audiometric frequencies 0.5, 0.75, 1, 1.5 and 2 kHz, which were generated by an AD229 audiometer (Interacoustics A/S, Middelfart, Denmark). Higher frequencies were not included as the movement of the round window becomes very complex above 3 kHz. The stimulation level was 60 dB HL for all measured frequencies, which required stimulation voltages of 0.9, 0.3, 0.2, 0.1 and 0.2 V at 0.5, 0.75, 1, 1.5 and 2 kHz, respectively. During stimulation, the velocity magnitude was measured via the LDV, with the same settings as described above via reflective tapes at two locations within the middle ear: (1) the promontory and (2) the RWM. The position of the reflective tapes was kept the same for all measurements (in a central position on the RWM). The measurement angles were estimated to be 70° and 45° for the promontory and RWM measurements, respectively.

Condition 2 (fixed): Fixed Stapes Footplate

Once all measurements in condition 1 were completed, the stapes footplate was fixed to simulate otosclerosis, the main indication for the DACI used in this study. The stapes footplate was prepared with a bonding agent (Clearfil SE BOND, Kuraray Co. Ltd., Kurashiki, Japan) and subsequently fixed with a dental compomer material (Dyract Seal, DENTSPLY DeTrey GmbH, Konstanz, Germany) using a curing light (DEMI™, Kerr Corp., Orange, CA, USA) for 10 seconds. Dental compomers are a hybrid of two other dental materials: dental composites and glass ionomer cement. This method was very effective in fixing the stapes footplate, which can be challenging in the humid oval window area. After vibrotactile confirmation that the stapes footplate was firmly fixed, the B81 was used to stimulate again at the audiometric frequencies with the same levels as during the first step and the velocity magnitude was measured at the promontory and the RWM. These measurements were taken in a limited number of ears (n=4 for the promontory measurements and n=6 for the RWM measurements).

Condition 3 (implanted): After DACI Implantation

After the measurements in the first two conditions were completed, the DACI system was implanted. Subsequent to the interruption of the ossicular chain between the incus and the stapes, the stapes superstructure was removed with a Skeeter® otologic drill (Medtronic, Minneapolis, MN, USA). Afterwards, a recess in the bone was drilled for the fixation system (black arrow in Figure 1), which was fixed with four bone screws (Stryker, Kalamazoo, MI, USA) to the surface of the skull. The actuator of the Codacs™ Direct Acoustic Cochlear Implant system (Cochlear™ Ltd., Sydney, Australia) was inserted into the fixation system and clamped in it. Care was taken that the actuator did not touch any bony structure. The implant receiver/stimulator was positioned under the skin. A stapedotomy was performed with the Skeeter drill under fluid to avoid introduction of air into the cochlea. Subsequently a conventional stapes prosthesis, a platinum/PTFE stapes prosthesis (Spiggle & Theiss, Dieburg, Germany) of 0.4 mm diameter, was inserted through the perforation in the stapes

footplate and firmly crimped to the artificial incus of the DACI actuator. The oval window was not sealed after insertion of the prosthesis. The DACI surgical procedure was described in detail by Lenarz et al. (Lenarz et al, 2013). After confirming good connection of the prosthesis to the artificial incus, the B81 was used to stimulate again at the same frequencies and the same levels as during the first and second condition and the velocity was measured at the promontory and RWM. All measurements in all conditions were performed twice to account for test-retest variability. The laser was repositioned for each measurement series.

Analysis

In this repeated-measures study design each ear served as its own control. Given the comparison of three groups and the low number of ears, Friedman's ANOVA was used for statistical analysis. In cases where a significant overall effect was found, the Wilcoxon signed-rank test was used for further analysis and a Bonferroni correction was applied. Only complete datasets were included for statistical analysis, for which the IBM® SPSS® Statistics Software Package (IBM Corporation, Armonk, NY, USA) was used. For Friedman's ANOVA, a significance level of 0.05 was adopted, for the Wilcoxon signed-rank test a significance level of 0.017 was used, compensating for the three comparisons.

Before analyzing the velocities, corrections for the measurement angles were made. This was done by calculating the cosine of the measurement angle. For the RWM the measurement angle was 45° which resulted in a correction of 0.707 or -3.0 dB. For the promontory the measurement angle was 70° which resulted in a correction of 0.342 or -9.3 dB.

Stenfelt et al. (Stenfelt et al., 2004b) argued that a relevant physical measure of the effect of bone conduction stimulation was to describe the vector difference (which depends on differences in magnitude and phase) between the driven RWM and promontory velocities. Due to limitations in our measurement and analysis system, we only report differences in the magnitude of these velocities. This limitation degrades our ability to distinguish small differences in vector velocities. The differences in

the velocity magnitudes at the RWM and the promontory were calculated by subtracting the promontory velocity magnitude from the RWM velocity magnitude. Our estimations of the vector differences in the RWM and promontory velocities are most valid at 500 Hz, where the RWM and promontory are expected to move roughly in phase (Stenfelt et al, 2004b). Any phase difference at higher frequencies complicated the assessment of any vector difference in velocity at the two locations.

As we stimulated with pure tones, stimulus related velocity was quantified as the rms value of the 1/3 octave band at the stimulation frequency, shown as a peak in the frequency response. The promontory and RWM velocities as well as the relative vibration magnitude of the RWM (difference between the RWM and promontory velocity magnitudes) were compared in the three conditions (normal, fixed and implanted).

Results

Figure 2 shows an exemplary frequency spectrum computed from the measured velocity magnitude produced by a 1 kHz continuous pure tone after DACI implantation at the promontory. For comparison the noise floor (velocity measurement without stimulus) is shown as well. The stimulus related velocity magnitude at the stimulation frequency can be seen as a peak in the graph (here: 1 kHz). In the following, only the rms value of the 1/3 octave band at the stimulation frequency will be reported as this value was used for the statistical analysis.

Figure 3 shows box and whisker plots of the velocity magnitude change at the RWM (light grey) and promontory (dark grey) between the post-implantation (implanted) and control/baseline (normal) condition (implanted – normal) after correction for the measurement angle. At the RWM the median velocity magnitude change ranged from 0.01 mm/s at 1.5 kHz to 0.07 mm/s at 500 Hz. At the promontory the median velocity magnitude change ranged from -0.08 mm/s at 500 Hz to 0.03 at 1000 Hz. The RWM velocity magnitude did not significantly change at any frequency except for 2 kHz, χ^2

(2) = 6.33, $p < 0.05$. Wilcoxon tests were used to follow up on this finding. A Bonferroni correction was applied and so all effects are reported at a 0.017 level of significance. It appeared that the RWM velocity magnitude did not significantly change between the normal and fixed condition, $T = 0$, $p = 0.028$, $r = -0.64$, between the fixed and implanted condition, $T = 9$, $p = 0.753$, $r = -0.09$, or between the normal and the implanted condition, $T = 12$, $p = 0.401$, $r = -0.21$. The promontory velocity magnitude did not change significantly at any frequency ($p > 0.05$).

In Figure 4 the change in the relative vibration magnitude of the RWM (difference between the RWM and promontory velocity magnitude) between the post-implantation (implanted) and control/baseline (normal) condition (implanted-normal) is shown per frequency. The median change ranged from 0.015 mm/s at 1.5 and 2 kHz to 0.13 mm/s at 500 Hz. In individual ears (see Figure 5 for an example of ear 5, which had a normal METF), the change in relative vibration magnitude of the RWM after DACI implantation were as much as 1.38 mm/s (at 500 Hz) and 0.2 mm/s (at 2 kHz). The relative vibration magnitude of the RWM only changed significantly at 500 Hz, $\chi^2(2) = 6.5$, $p < 0.05$. Wilcoxon tests were used to follow up on this finding and a Bonferroni correction was applied. It appeared that the relative vibration of the RWM did not significantly change between the normal and fixed condition, $T = 0$, $p = 0.068$, $r = -0.65$, between the fixed and implanted condition, $T = 0$, $p = 0.068$, $r = -0.65$ or between the normal and the implanted condition, $T = 3$, $p = 0.036$, $r = -0.53$. The median relative velocity magnitude increases were in the order of 4 dB (0.13 mm/s) at 0.5 kHz and of 3 dB at 0.75 kHz (0.045 mm/s), 1.5 kHz (0.015 mm/s) and 2 kHz (0.015 mm/s, calculated from the measured velocities reported above) and were thus below the clinically relevant level of 5 dB.

Figure 6 shows the relative vibration magnitudes of the RWM for all three conditions: before implantation (normal), after stapes footplate fixation (fixed) and after implantation of the DACI (implanted). As reported above, there were no significant changes in the relative velocity magnitude of the RWM between any of the three conditions.

General experimental findings

The test-retest variability was 0.7 dB on average over all conditions and measurement locations, being the smallest at the RWM (Mean: 0.5 dB, standard deviation (SD): 0.6 dB) and highest at the PC (Mean: 1.1 dB, SD: 1.4 dB).

The variability of the maximum promontory velocity magnitudes between ears was high. Standard deviations up to 12 dB were calculated. This showed that there was a relatively large variability between ears in the baseline vibration measurement of the promontory before any manipulations on the ossicles or inner ear windows were done. Although the source of variability was reduced as much as possible by e.g. using the same bone vibrator and the same controlled stimulation location on the mastoid, variations related to the biological characteristics of the individual head such as, e.g., cranial attenuation could not be controlled for.

The noise of the measurement system was assessed by measuring the RWM and promontory velocity without any input to the bone vibrator. The signal-to-noise ratio (SNR) was minimally 20 dB SNR and up to 40 dB SNR in the low frequencies.

A change in the stimulation level did not alter the shape of the frequency response of the promontory or RWM. In the first two ears we increased the stimulation level in steps of 10 dB resulting in a linear increase in the maximum RWM and promontory velocity magnitude (data not shown). In the remaining six ears we only measured at one stimulation level per frequency.

The METF of six ears satisfied the modified minimum acceptance criteria of the ASTM F2504-05 standard (ASTM, 2005) by Rosowski et al. (Rosowski, Chien, Ravicz, 2007). For two ears (ear 3 and ear 6), no valid results for the METF could be obtained during the measurements. Both ears showed results close to the median results, while some ears with a normal METF showed very high or low velocity changes at specific frequencies. However, the absolute values of the METF were of minor importance for the present investigation as not the movement of the ossicles but the relative motion

between the promontory and the RWM velocity was investigated; hence data sets from all 8 ears were used.

Discussion

The velocity magnitudes of the RWM and promontory were measured during BC stimulation in 8 ears from 5 human cadaver heads in three conditions. To estimate the effect of DACI implantation on the cochlear stimulation via BC, the manipulation-induced changes in both the RWM and promontory velocity magnitudes (implanted - normal), as well as the changes in the relative vibration magnitudes of the RWM (i.e. the difference between the RWM and promontory velocity magnitudes) were analyzed.

We have measured the RWM velocity magnitude at a single measurement point. Provided that the vibration pattern of the RWM is equal before and after the alteration procedure, the change in RWM motion gives an estimate of the change in stimulation level of the cochlea, caused by the procedure (Stenfelt et al, 2004b). According to Stenfelt et al. there is a similarity in the characteristics of the magnitude and phase data between various positions on the RWM in the same specimen below 1.5 kHz (Stenfelt et al, 2004b). However, above 500 Hz, the phase difference increases with frequency and is highest at the central position of the RWM compared with other position on the RWM (Stenfelt et al, 2004b). As we did not measure the phase, our results are most valid at 500 Hz. We have also included 2 kHz in our analysis as it is an important frequency for DACI. However, the phase difference between the center position at the RWM and the promontory at 2 kHz is approximately 100° (Stenfelt et al., 2004b), which made it difficult to draw conclusions from the measurement results at this frequency.

The relative vibration magnitude of the RWM, i.e. the velocity magnitude difference between the RWM and the promontory can be used to estimate the effect of an alteration of the BC pathways, here, the implantation of a DACI. Comparing the relative vibration magnitude of the RWM before and after

DACI implantation (see Figure 4) no significant changes were observed, possibly because phase information was not included in the analysis.

For air and bone conducted sound, the RWM vibrates nearly as a whole in an in-and-out motion for frequencies below 1.5 kHz, while above this frequency, the RWM moves primarily in two sections that vibrate with approximately 180° difference (Stenfelt et al, 2004b). Above 3 kHz, the membrane motion is complex with a mixture of modal and traveling wave motions (Stenfelt et al., 2004b). We measured at the central position on the RWM and therefore the measurement results at frequencies above 1.5 kHz must be interpreted with care. Additionally, phase differences between the RWM and the promontory above 500 Hz influenced our results.

The relative vibration magnitude of the RWM did not change significantly after stapes fixation at any frequency. This is not surprising giving the fact that in BC stimulation, the RW fluid volume displacement was reported to be very similar between a normal intact specimen and a specimen with a stapes footplate glued to the otic capsule (Stenfelt et al, 2004a). Stenfelt could simulate the reduction in BC sensitivity after stapes immobilization in his model and indicated that it is a result of the removal of the middle ear component together with alteration of the impedance at the oval window leading to less efficient contribution from the inner ear components (Stenfelt, 2016). This could not be seen in our data and might indicate that the relative velocity of the RWM, suggested as a good indicator for the sensory output of the cochlea (Stenfelt et al, 2004b), does not take into account all BC pathways and might not be a good indicator for the cochlear input drive. However, it could also indicate that the phase information is an important parameter. As we did not measure the phase, we cannot draw conclusions on that. Stieger et al. reported that the promontory acceleration for low frequencies and the promontory velocity for high frequencies can be used as an estimation for cochlear stimulation via BC (Stieger et al, 2018). These different opinions indicate that the cochlear input via BC is still not completely understood. More research is needed to relate the intracochlear pressures and their differences to the promontory, oval and RWM velocity and their differences, respectively, when

stimulating via BC. This is specifically interesting for DACI as they stimulate the inner ear directly and circumvent the outer and middle ear. At the same time an additional pathway for BC sound is created.

A model simulating the new pathway created by the DACI system, similar to what Stenfelt did for the conventional pathways (Stenfelt, 2016), might be an alternative way to estimate the influence of a DACI system on the BC thresholds and its relative importance compared to the conventional pathways.

One shortcoming of our study is that we did not include phase measurements. They could have provided additional information on the relative movements between the RWM and the promontory. Vibration measurements of the fixation system and the bone surrounding the fixation system could also have provided valuable additional information.

Another shortcoming is that we only measured on a single, central point on the RWM. Therefore, the results above 1.5 kHz have to be interpreted with care as they might be influenced by measurement errors.

Conclusions

In this study, no significant changes in the relative vibration magnitude of the RWM after DACI implantation were observed. The lack of significant changes in the relative magnitude of motion suggests that any relative motion within the DACI actuator and fixation system does not effectively contribute to the cochlear response produced by BC stimulation. The improvements in BC thresholds seen in patients implanted with a DACI very likely have their origin in the changed impedance at the oval window after DACI surgery leading to a more efficient contribution from the inner ear components to BC sound.

References

- Arnold A., Fawzy T., Steinhoff H.-J., Kiefer J. & Arnold W. 2011. Influence of a totally open oval window on bone conduction in otosclerosis. *Audiology & Neuro-Otology*, 16(1), 23–28. <https://doi.org/10.1159/000308302>
- ASTM. *ASTM F2504-05 Standard Practice for Describing System Output of Implantable Middle Ear Hearing Devices*. 2005.
- Baumgartner W. D., Böheim K., Hagen R., Müller J., Lenarz T., Reiss S., ... Opie, J. M. 2010. The Vibrant Soundbridge for conductive and mixed hearing losses: European multicenter study results. *Advances in Oto-Rhino-Laryngology*, 69, 38–50. <https://doi.org/10.1159/000318521>
- Bruchhage K. L., Leichtle A., Schoenweiler R., Todt I., Baumgartner W. D., Frenzel H. & Wollenberg B. 2016. Systematic review to evaluate the safety, efficacy and economical outcomes of the Vibrant Soundbridge for the treatment of sensorineural hearing loss. *European Archives of Oto-Rhino-Laryngology*, 1–10. <https://doi.org/10.1007/s00405-016-4361-2>
- Busch S., Kruck S., Spickers D., Leuwer R., Hoth S., Praetorius M., ... Lenarz T. 2013. First clinical experiences with a direct acoustic cochlear stimulator in comparison to preoperative fitted conventional hearing aids. *Otology & Neurotology: Official Publication of the American Otological Society, American Neurotology Society [and] European Academy of Otology and Neurotology*, 34(9), 1711–1718. <https://doi.org/10.1097/MAO.0000000000000225>
- Carhart R. 1949. Clinical reliability of bone conduction audiometry. *Laryngoscope*, 1084–1101.
- Chhan D., Bowers P., McKinnon M. L. & Rosowski J. J. 2016. Middle-ear and inner-ear contribution to bone conduction in chinchilla: The development of Carhart's notch. *Hearing Research*, 340, 144–152. <https://doi.org/10.1016/j.heares.2016.02.015>
- Häusler R. 2008. A Novel Implantable Hearing System with Direct Acoustic Cochlear Stimulation.

Audiology and Neurotology, 13, 247–256.

Homma K., Du Y., Shimizu Y. & Puria S. 2009. Ossicular resonance modes of the human middle ear for bone and air conduction. *The Journal of the Acoustical Society of America*, 125(2), 968–979.
<https://doi.org/10.1121/1.3056564>

Kludt E., Büchner A., Schwab B., Lenarz T. & Maier, H. 2016. Indication of Direct Acoustical Cochlea Stimulation in Comparison to Cochlear Implants. *Hearing Research*.
<https://doi.org/10.1016/j.heares.2016.01.016>

Kludt E., D'hondt C., Lenarz T., & Maier H. 2017. Clinical Validation of a Sound Processor Upgrade in Direct Acoustic Cochlear Implant Subjects. *Otology & Neurotology*, 38.
<https://doi.org/10.1097/MAO.0000000000001404>

Lenarz T., Verhaert N., Desloovere C., Desmet J., D'hondt C., González J. C. F., ... Wasowski A. 2014. A comparative study on speech in noise understanding with a direct acoustic cochlear implant in subjects with severe to profound mixed hearing loss. *Audiology and Neurotology*, 19(3), 164–174. <https://doi.org/10.1159/000358004>

Lenarz T., Zwartenkot J. W., Stieger C., Schwab B., Mylanus E. a M., Caversaccio M., ... Mojallal H. 2013. Multicenter study with a direct acoustic cochlear implant. *Otology & Neurotology : Official Publication of the American Otological Society, American Neurotology Society [and] European Academy of Otology and Neurotology*, 34(7), 1215–1225.
<https://doi.org/10.1097/MAO.0b013e318298aa76>

Perez R., de Almeida J., Nedzelski J.M., Chen J. M. 2009. Variations in the Carhart Notch and Overclosure After Laser-Assisted Stapedotomy in Otosclerosis. *Otology & Neurotology*, 30, 1033–1036.

Rosowski J.J., Chien W., Ravicz M. M. S. 2007. Testing a Method for Quantifying the Output of

Implantable Middle Ear Hearing Devices. *Audiology and Neurotology*, 12(4), 265–276.

<https://doi.org/10.1159/000101474>.Testing

Sabbe A., Verhaert N., Joossen I., Lammens A., & Debruyne F. 2015. Otosclerosis: Shift in bone conduction after stapedotomy. *B-Ent*, 11(3), 183–189.

Schwab B., Kludt E., Maier H., Lenarz T., & Teschner M. 2015. Subtotal petrosectomy and Codacs™: new possibilities in ears with chronic infection. *European Archives of Oto-Rhino-Laryngology : Official Journal of the European Federation of Oto-Rhino-Laryngological Societies (EUFOS) : Affiliated with the German Society for Oto-Rhino-Laryngology - Head and Neck Surgery*.
<https://doi.org/10.1007/s00405-015-3688-4>

Stenfelt S. 2016. Model predictions for bone conduction perception in the human. *Hearing Research*, 340, 135–143. <https://doi.org/10.1016/j.heares.2015.10.014>

Stenfelt S., Hato N. & Goode R. L. 2004a. Fluid volume displacement at the oval and round windows with air and bone conduction stimulation. *The Journal of the Acoustic Society of America*, 115(2), 797–812.

Stenfelt S., Hato N. & Goode, R. L. 2004b. Round window membrane motion with air conduction and bone conduction stimulation. *Hearing Research*, 198(1–2), 10–24.
<https://doi.org/10.1016/j.heares.2004.07.008>

Stieger C., Farahmand R. B., Page B. F., Roushan K., Merchant J. P., Abur D., ... Nakajima H. H. 2015. Pressures in the human cochlea during bone conduction. *AIP Conference Proceedings*, 1703. <https://doi.org/10.1063/1.4939359>

Stieger C., Guan,X., Farahmand R. B., Page B. F., Merchant J. P., Abur D., & Nakajima H. H. 2018. Intracochlear Sound Pressure Measurements in Normal Human Temporal Bones During Bone Conduction Stimulation. *JARO - Journal of the Association for Research in Otolaryngology*,

19(5), 523–539. <https://doi.org/10.1007/s10162-018-00684-1>